

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 4/8/2024 4:34:12 PM

JOB DESCRIPTION

RED-HILL [SUBCONTRACT]
625, 8015
RUSH Weekly Red Hill

JOB NUMBER

380-84152-2

Eurofins Eaton Analytical Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Generated
4/8/2024 4:34:12 PM

Authorized for release by
Rachelle Arada, Project Manager
Rachelle.Arada@et.eurofinsus.com
(626)386-1106

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Surrogate Summary	12
QC Sample Results	14
QC Association Summary	19
Lab Chronicle	21
Certification Summary	23
Method Summary	24
Sample Summary	25
Subcontract Data	26
Chain of Custody	60
Receipt Checklists	62



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2

Job ID: 380-84152-2

Eurofins Eaton Analytical Pomona

Job Narrative 380-84152-2

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 2/22/2024 10:15 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 2.0°C and 2.8°C.

Receipt Exceptions

One or more containers for the following samples were received broken or leaking: HALAWA WELLS UNITS 1 & 2 P1 (380-84152-4. One 8015B_GRO_LL vial from site HALAWA WELLS UNITS 1&2 P1 was received broken.

Subcontract Work

Method 625 PAH Physis LL (EAL) + TICs: This method was subcontracted to Physis Environmental Laboratories. The subcontract laboratory certification is different from that of the facility issuing the final report. The subcontract report is appended in its entirety.

Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

GC Semi VOA

Method 8015B_DRO_LL_CS: Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with preparation batch 570-413924. The laboratory control sample (LCS) was performed in duplicate (LCSD) to provide precision data for this batch.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Client Sample ID: MOANALUA WELLS
PWSID Number: HI0000331

Lab Sample ID: 380-84152-1

No Detections.

Client Sample ID: AIEA GULCH WELLS PUMP 2
PWSID Number: HI0000331

Lab Sample ID: 380-84152-2

No Detections.

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2
PWSID Number: HI0000331

Lab Sample ID: 380-84152-3

No Detections.

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1
PWSID Number: HI0000331

Lab Sample ID: 380-84152-4

No Detections.

Client Sample ID: TB MOANALUA WELLS

Lab Sample ID: 380-84152-5

No Detections.

Client Sample ID: TB AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-84152-6

No Detections.

Client Sample ID: TB AIEA WELLS PUMPS 1&2 (260)

Lab Sample ID: 380-84152-7

No Detections.

Client Sample ID: TB HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-84152-8

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Client Sample ID: MOANALUA WELLS

Lab Sample ID: 380-84152-1

Date Collected: 02/20/24 09:42

Matrix: Drinking Water

Date Received: 02/22/24 10:15

PWSID Number: HI0000331

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 19:32	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	85		38 - 134				02/27/24 19:32	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<27		27	ug/L		02/24/24 17:03	02/26/24 22:11	1
Motor Oil Range Organics [C24-C36]	<27		27	ug/L		02/24/24 17:03	02/26/24 22:11	1
C8-C18	<27		27	ug/L		02/24/24 17:03	02/26/24 22:11	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	113		60 - 130			02/24/24 17:03	02/26/24 22:11	1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics I

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Acenaphthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Acenaphthylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Biphenyl	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Chrysene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Dibenzothiophene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		02/27/24 00:00	03/25/24 22:35	1
Fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Fluorene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Naphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Phenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 22:35	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	73		27 - 133				02/27/24 00:00	03/25/24 22:35	1
(d10-Phenanthrene)	82		43 - 129				02/27/24 00:00	03/25/24 22:35	1
(d12-Chrysene)	92		52 - 144				02/27/24 00:00	03/25/24 22:35	1
(d12-Perylene)	90		36 - 161				02/27/24 00:00	03/25/24 22:35	1
(d8-Naphthalene)	58		25 - 125				02/27/24 00:00	03/25/24 22:35	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Client Sample ID: AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-84152-2

Date Collected: 02/20/24 10:54

Matrix: Drinking Water

Date Received: 02/22/24 10:15

PWSID Number: HI0000331

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 19:06	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	85		38 - 134				02/27/24 19:06	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<27		27	ug/L		02/24/24 17:03	02/26/24 22:32	1
Motor Oil Range Organics [C24-C36]	<27		27	ug/L		02/24/24 17:03	02/26/24 22:32	1
C8-C18	<27		27	ug/L		02/24/24 17:03	02/26/24 22:32	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	109		60 - 130			02/24/24 17:03	02/26/24 22:32	1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics I

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Acenaphthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Acenaphthylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Biphenyl	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Chrysene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Dibenzothiophene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		02/27/24 00:00	03/26/24 00:24	1
Fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Fluorene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Naphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Phenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 00:24	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	65		27 - 133				02/27/24 00:00	03/26/24 00:24	1
(d10-Phenanthrene)	73		43 - 129				02/27/24 00:00	03/26/24 00:24	1
(d12-Chrysene)	79		52 - 144				02/27/24 00:00	03/26/24 00:24	1
(d12-Perylene)	90		36 - 161				02/27/24 00:00	03/26/24 00:24	1
(d8-Naphthalene)	53		25 - 125				02/27/24 00:00	03/26/24 00:24	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2

Lab Sample ID: 380-84152-3

Date Collected: 02/20/24 11:31

Matrix: Drinking Water

Date Received: 02/22/24 10:15

PWSID Number: HI0000331

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 18:40	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	82		38 - 134				02/27/24 18:40	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		02/24/24 17:03	02/26/24 22:53	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		02/24/24 17:03	02/26/24 22:53	1
C8-C18	<26		26	ug/L		02/24/24 17:03	02/26/24 22:53	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	112		60 - 130			02/24/24 17:03	02/26/24 22:53	1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics I

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Acenaphthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Acenaphthylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Biphenyl	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Chrysene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Dibenzothiophene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		02/27/24 00:00	03/26/24 02:12	1
Fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Fluorene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Naphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Phenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 02:12	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	73		27 - 133				02/27/24 00:00	03/26/24 02:12	1
(d10-Phenanthrene)	78		43 - 129				02/27/24 00:00	03/26/24 02:12	1
(d12-Chrysene)	108		52 - 144				02/27/24 00:00	03/26/24 02:12	1
(d12-Perylene)	90		36 - 161				02/27/24 00:00	03/26/24 02:12	1
(d8-Naphthalene)	59		25 - 125				02/27/24 00:00	03/26/24 02:12	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-84152-4

Date Collected: 02/20/24 10:26

Matrix: Drinking Water

Date Received: 02/22/24 10:15

PWSID Number: HI0000331

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 18:14	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	86		38 - 134				02/27/24 18:14	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		02/24/24 17:03	03/13/24 15:30	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		02/24/24 17:03	03/13/24 15:30	1
C8-C18	<25		25	ug/L		02/24/24 17:03	03/13/24 15:30	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	114		60 - 130			02/24/24 17:03	03/13/24 15:30	1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics I

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Acenaphthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Acenaphthylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Biphenyl	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Chrysene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Dibenzothiophene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		02/27/24 00:00	03/26/24 04:01	1
Fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Fluorene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Naphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Phenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/26/24 04:01	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	85		27 - 133				02/27/24 00:00	03/26/24 04:01	1
(d10-Phenanthrene)	89		43 - 129				02/27/24 00:00	03/26/24 04:01	1
(d12-Chrysene)	122		52 - 144				02/27/24 00:00	03/26/24 04:01	1
(d12-Perylene)	94		36 - 161				02/27/24 00:00	03/26/24 04:01	1
(d8-Naphthalene)	70		25 - 125				02/27/24 00:00	03/26/24 04:01	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Client Sample ID: TB MOANALUA WELLS

Lab Sample ID: 380-84152-5

Date Collected: 02/20/24 09:42

Matrix: Water

Date Received: 02/22/24 10:15

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 16:29	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	78		38 - 134				02/27/24 16:29	1

Client Sample ID: TB AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-84152-6

Date Collected: 02/20/24 10:54

Matrix: Water

Date Received: 02/22/24 10:15

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 16:55	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	82		38 - 134				02/27/24 16:55	1

Client Sample ID: TB AIEA WELLS PUMPS 1&2 (260)

Lab Sample ID: 380-84152-7

Date Collected: 02/20/24 11:31

Matrix: Water

Date Received: 02/22/24 10:15

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 17:21	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	86		38 - 134				02/27/24 17:21	1

Client Sample ID: TB HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-84152-8

Date Collected: 02/20/24 10:26

Matrix: Water

Date Received: 02/22/24 10:15

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 17:48	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	97		38 - 134				02/27/24 17:48	1

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Drinking Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-84152-1	MOANALUA WELLS	85
380-84152-2	AIEA GULCH WELLS PUMP 2	85
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	82
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	86
380-84152-4 MS	HALAWA WELLS UNITS 1 & 2 P1	92
380-84152-4 MSD	HALAWA WELLS UNITS 1 & 2 P1	94

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-84152-5	TB MOANALUA WELLS	78
380-84152-6	TB AIEA GULCH WELLS PUMP 2	82
380-84152-7	TB AIEA WELLS PUMPS 1&2 (260)	86
380-84152-8	TB HALAWA WELLS UNITS 1 & 2 P1	97
LCS 570-414522/4	Lab Control Sample	103
LCSD 570-414522/5	Lab Control Sample Dup	99
MB 570-414522/6	Method Blank	84
MRL 570-414522/3	Lab Control Sample	83

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Drinking Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-84152-1	MOANALUA WELLS	113
380-84152-2	AIEA GULCH WELLS PUMP 2	109
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	112
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	114

Surrogate Legend

OTCSN = n-Octacosane (Surr)

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
LCS 570-413924/2-A	Lab Control Sample	113
LCSD 570-413924/3-A	Lab Control Sample Dup	106
MB 570-413924/1-A	Method Blank	114
MRL 570-413924/4-A	Lab Control Sample	109

Surrogate Legend

OTCSN = n-Octacosane (Surr)

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Matrix: BlankMatrix

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)				
		Acenaphtl (27-133)	Phenanth (43-129)	CRY (52-144)	NPT (25-125)	PRY (36-161)
116076-B1	Method Blank	83	86	91	68	88
116076-BS1	Lab Control Sample	79	85	90	65	88
116076-BS2	Lab Control Sample Dup	84	88	94	68	91

Surrogate Legend

(d10-Acenaphthene) = (d10-Acenaphthene)

(d10-Phenanthrene) = (d10-Phenanthrene)

CRY = (d12-Chrysene)

NPT = (d8-Naphthalene)

PRY = (d12-Perylene)

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Matrix: Drinking Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)				
		Acenaphtl (27-133)	Phenanth (43-129)	CRY (52-144)	NPT (25-125)	PRY (36-161)
380-84152-1	MOANALUA WELLS	73	82	92	58	90
380-84152-2	AIEA GULCH WELLS PUMP 2	65	73	79	53	90
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	73	78	108	59	90
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	85	89	122	70	94

Surrogate Legend

(d10-Acenaphthene) = (d10-Acenaphthene)

(d10-Phenanthrene) = (d10-Phenanthrene)

CRY = (d12-Chrysene)

NPT = (d8-Naphthalene)

PRY = (d12-Perylene)

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: MB 570-414522/6
Matrix: Water
Analysis Batch: 414522

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/27/24 15:54	1
Surrogate	MB %Recovery	MB Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	84		38 - 134				02/27/24 15:54	1

Lab Sample ID: LCS 570-414522/4
Matrix: Water
Analysis Batch: 414522

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	400	386		ug/L		97	78 - 120
Surrogate	LCS %Recovery	LCS Qualifier	Limits				
4-Bromofluorobenzene (Surr)	103		38 - 134				

Lab Sample ID: LCSD 570-414522/5
Matrix: Water
Analysis Batch: 414522

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	400	378		ug/L		94	78 - 120	2	10
Surrogate	LCSD %Recovery	LCSD Qualifier	Limits						
4-Bromofluorobenzene (Surr)	99		38 - 134						

Lab Sample ID: MRL 570-414522/3
Matrix: Water
Analysis Batch: 414522

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	10.0	11.9		ug/L		119	50 - 150
Surrogate	MRL %Recovery	MRL Qualifier	Limits				
4-Bromofluorobenzene (Surr)	83		38 - 134				

Lab Sample ID: 380-84152-4 MS
Matrix: Drinking Water
Analysis Batch: 414522

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	<10		400	363		ug/L		91	68 - 122
Surrogate	MS %Recovery	MS Qualifier	Limits						
4-Bromofluorobenzene (Surr)	92		38 - 134						

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: 380-84152-4 MSD
Matrix: Drinking Water
Analysis Batch: 414522

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	<10		400	363		ug/L		91	68 - 122	0	18
Surrogate	%Recovery	MSD Qualifier	MSD Limits								
4-Bromofluorobenzene (Surr)	94		38 - 134								

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Lab Sample ID: MB 570-413924/1-A
Matrix: Water
Analysis Batch: 414270

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 413924

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		02/24/24 17:03	02/26/24 20:24	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		02/24/24 17:03	02/26/24 20:24	1
C8-C18	<25		25	ug/L		02/24/24 17:03	02/26/24 20:24	1
Surrogate	%Recovery	MB Qualifier	MB Limits	Prepared	Analyzed	Dil Fac		
n-Octacosane (Surr)	114		60 - 130	02/24/24 17:03	02/26/24 20:24	1		

Lab Sample ID: LCS 570-413924/2-A
Matrix: Water
Analysis Batch: 414270

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 413924

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	1600	1590		ug/L		100	56 - 127
Surrogate	%Recovery	LCS Qualifier	LCS Limits				
n-Octacosane (Surr)	113		60 - 130				

Lab Sample ID: LCSD 570-413924/3-A
Matrix: Water
Analysis Batch: 414270

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 413924

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	1600	1490		ug/L		93	56 - 127	7	23
Surrogate	%Recovery	LCSD Qualifier	LCSD Limits						
n-Octacosane (Surr)	106		60 - 130						

Lab Sample ID: MRL 570-413924/4-A
Matrix: Water
Analysis Batch: 414270

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 413924

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	0.0200	<0.020		mg/L		76	50 - 150

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Lab Sample ID: MRL 570-413924/4-A
 Matrix: Water
 Analysis Batch: 414270

Client Sample ID: Lab Control Sample
 Prep Type: Total/NA
 Prep Batch: 413924

Surrogate	%Recovery	MRL MRL Qualifier	Limits
<i>n</i> -Octacosane (Surr)	109		60 - 130

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Lab Sample ID: 116076-B1
 Matrix: BlankMatrix
 Analysis Batch: O-44126

Client Sample ID: Method Blank
 Prep Type: Total/NA
 Prep Batch: O-44126_P

Analyte	Blank Result	Blank Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Acenaphthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Acenaphthylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Biphenyl	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Chrysene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Dibenzothiophene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		02/27/24 00:00	03/25/24 17:06	1
Fluoranthene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Fluorene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Naphthalene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Perylene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Phenanthrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1
Pyrene	ND		0.005	0.001	µg/L		02/27/24 00:00	03/25/24 17:06	1

Surrogate	%Recovery	Blank Blank Qualifier	Limits	Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	83		27 - 133	02/27/24 00:00	03/25/24 17:06	1
(d10-Phenanthrene)	86		43 - 129	02/27/24 00:00	03/25/24 17:06	1
(d12-Chrysene)	91		52 - 144	02/27/24 00:00	03/25/24 17:06	1
(d12-Perylene)	88		36 - 161	02/27/24 00:00	03/25/24 17:06	1
(d8-Naphthalene)	68		25 - 125	02/27/24 00:00	03/25/24 17:06	1

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

Lab Sample ID: 116076-BS1
Matrix: BlankMatrix
Analysis Batch: O-44126

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: O-44126_P

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.5	0.398		µg/L		80	31 - 128
1-Methylphenanthrene	0.5	0.539		µg/L		108	66 - 127
2,3,5-Trimethylnaphthalene	0.5	0.482		µg/L		96	55 - 122
2,6-Dimethylnaphthalene	0.5	0.433		µg/L		87	48 - 120
2-Methylnaphthalene	0.5	0.404		µg/L		81	47 - 130
Acenaphthene	0.5	0.453		µg/L		91	53 - 131
Acenaphthylene	0.5	0.588		µg/L		118	43 - 140
Anthracene	0.5	0.509		µg/L		102	58 - 135
Benz[a]anthracene	0.5	0.591		µg/L		118	55 - 145
Benzo[a]pyrene	0.5	0.527		µg/L		105	51 - 143
Benzo[b]fluoranthene	0.5	0.558		µg/L		112	46 - 165
Benzo[e]pyrene	0.5	0.457		µg/L		91	42 - 152
Benzo[g,h,i]perylene	0.5	0.521		µg/L		104	63 - 133
Benzo[k]fluoranthene	0.5	0.513		µg/L		103	56 - 145
Biphenyl	0.5	0.408		µg/L		82	56 - 119
Chrysene	0.5	0.528		µg/L		106	56 - 141
Dibenz[a,h]anthracene	0.5	0.728		µg/L		146	55 - 150
Dibenzo[a,l]pyrene	0.5	0.435		µg/L		87	50 - 150
Dibenzothiophene	0.5	0.48		µg/L		96	46 - 126
Disalicylidenepropanediamine	50	47.6		µg/L		95	50 - 150
Fluoranthene	0.5	0.541		µg/L		108	60 - 146
Fluorene	0.5	0.491		µg/L		98	58 - 131
Indeno[1,2,3-cd]pyrene	0.5	0.746		µg/L		149	50 - 151
Naphthalene	0.5	0.366		µg/L		73	41 - 126
Perylene	0.5	0.489		µg/L		98	48 - 141
Phenanthrene	0.5	0.471		µg/L		94	67 - 127
Pyrene	0.5	0.544		µg/L		109	54 - 156

Surrogate	LCS %Recovery	LCS Qualifier	LCS Limits
(d10-Acenaphthene)	79		27 - 133
(d10-Phenanthrene)	85		43 - 129
(d12-Chrysene)	90		52 - 144
(d12-Perylene)	88		36 - 161
(d8-Naphthalene)	65		25 - 125

Lab Sample ID: 116076-BS2
Matrix: BlankMatrix
Analysis Batch: O-44126

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: O-44126_P

Analyte	Spike Added	LCS DUP Result	LCS DUP Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	0.5	0.411		µg/L		82	31 - 128	2	30
1-Methylphenanthrene	0.5	0.535		µg/L		107	66 - 127	1	30
2,3,5-Trimethylnaphthalene	0.5	0.495		µg/L		99	55 - 122	3	30
2,6-Dimethylnaphthalene	0.5	0.454		µg/L		91	48 - 120	4	30
2-Methylnaphthalene	0.5	0.417		µg/L		83	47 - 130	2	30
Acenaphthene	0.5	0.473		µg/L		95	53 - 131	4	30
Acenaphthylene	0.5	0.599		µg/L		120	43 - 140	2	30
Anthracene	0.5	0.53		µg/L		106	58 - 135	4	30

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

Lab Sample ID: 116076-BS2
Matrix: BlankMatrix
Analysis Batch: O-44126

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: O-44126_P

Analyte	Spike Added	LCS DUP Result	LCS DUP Qualifier	Unit	D	%Rec	%Rec		RPD	Limit
							Limits	RPD		
Benz[a]anthracene	0.5	0.568		µg/L		114	55 - 145	3	30	
Benzo[a]pyrene	0.5	0.534		µg/L		107	51 - 143	2	30	
Benzo[b]fluoranthene	0.5	0.538		µg/L		108	46 - 165	4	30	
Benzo[e]pyrene	0.5	0.464		µg/L		93	42 - 152	2	30	
Benzo[g,h,i]perylene	0.5	0.529		µg/L		106	63 - 133	2	30	
Benzo[k]fluoranthene	0.5	0.497		µg/L		99	56 - 145	4	30	
Biphenyl	0.5	0.424		µg/L		85	56 - 119	4	30	
Chrysene	0.5	0.526		µg/L		105	56 - 141	1	30	
Dibenz[a,h]anthracene	0.5	0.73		µg/L		146	55 - 150	0	30	
Dibenzo[a,l]pyrene	0.5	0.379		µg/L		76	50 - 150	13	30	
Dibenzothiophene	0.5	0.493		µg/L		99	46 - 126	3	30	
Disalicylidenepropanediamine	50	49.6		µg/L		99	50 - 150	4	30	
Fluoranthene	0.5	0.537		µg/L		107	60 - 146	1	30	
Fluorene	0.5	0.5		µg/L		100	58 - 131	2	30	
Indeno[1,2,3-cd]pyrene	0.5	0.749		µg/L		150	50 - 151	1	30	
Naphthalene	0.5	0.38		µg/L		76	41 - 126	4	30	
Perylene	0.5	0.494		µg/L		99	48 - 141	1	30	
Phenanthrene	0.5	0.488		µg/L		98	67 - 127	4	30	
Pyrene	0.5	0.54		µg/L		108	54 - 156	1	30	

Surrogate	LCS DUP		Limits
	%Recovery	Qualifier	
(d10-Acenaphthene)	84		27 - 133
(d10-Phenanthrene)	88		43 - 129
(d12-Chrysene)	94		52 - 144
(d12-Perylene)	91		36 - 161
(d8-Naphthalene)	68		25 - 125

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

GC VOA

Analysis Batch: 414522

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-1	MOANALUA WELLS	Total/NA	Drinking Water	8015B GRO LL	
380-84152-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	8015B GRO LL	
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	8015B GRO LL	
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	8015B GRO LL	
380-84152-5	TB MOANALUA WELLS	Total/NA	Water	8015B GRO LL	
380-84152-6	TB AIEA GULCH WELLS PUMP 2	Total/NA	Water	8015B GRO LL	
380-84152-7	TB AIEA WELLS PUMPS 1&2 (260)	Total/NA	Water	8015B GRO LL	
380-84152-8	TB HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Water	8015B GRO LL	
MB 570-414522/6	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-414522/4	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-414522/5	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-414522/3	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-84152-4 MS	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	8015B GRO LL	
380-84152-4 MSD	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	8015B GRO LL	

GC Semi VOA

Prep Batch: 413924

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-1	MOANALUA WELLS	Total/NA	Drinking Water	3510C	
380-84152-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	3510C	
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	3510C	
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	3510C	
MB 570-413924/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-413924/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-413924/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-413924/4-A	Lab Control Sample	Total/NA	Water	3510C	

Analysis Batch: 414270

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-1	MOANALUA WELLS	Total/NA	Drinking Water	8015B	413924
380-84152-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	8015B	413924
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	8015B	413924
MB 570-413924/1-A	Method Blank	Total/NA	Water	8015B	413924
LCS 570-413924/2-A	Lab Control Sample	Total/NA	Water	8015B	413924
LCSD 570-413924/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	413924
MRL 570-413924/4-A	Lab Control Sample	Total/NA	Water	8015B	413924

Analysis Batch: 419748

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	8015B	413924

Subcontract

Analysis Batch: O-44126

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-1	MOANALUA WELLS	Total/NA	Drinking Water	625 PAH Physis LL (EAL) + TICs	O-44126_P
380-84152-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	625 PAH Physis LL (EAL) + TICs	O-44126_P
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	625 PAH Physis LL (EAL) + TICs	O-44126_P

Eurofins Eaton Analytical Pomona

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Subcontract (Continued)

Analysis Batch: O-44126 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	625 PAH Physis LL (EAL) + TICs	O-44126_P
116076-B1	Method Blank	Total/NA	BlankMatrix	625 PAH Physis LL (EAL) + TICs	O-44126_P
116076-BS1	Lab Control Sample	Total/NA	BlankMatrix	625 PAH Physis LL (EAL) + TICs	O-44126_P
116076-BS2	Lab Control Sample Dup	Total/NA	BlankMatrix	625 PAH Physis LL (EAL) + TICs	O-44126_P

Prep Batch: O-44126_P

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-84152-1	MOANALUA WELLS	Total/NA	Drinking Water	EPA_625	O-44126_P
380-84152-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	EPA_625	O-44126_P
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	EPA_625	O-44126_P
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	EPA_625	O-44126_P
116076-B1	Method Blank	Total/NA	BlankMatrix	EPA_625	O-44126_P
116076-BS1	Lab Control Sample	Total/NA	BlankMatrix	EPA_625	O-44126_P
116076-BS2	Lab Control Sample Dup	Total/NA	BlankMatrix	EPA_625	O-44126_P

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
 SDG: 625, 8015

Client Sample ID: MOANALUA WELLS

Lab Sample ID: 380-84152-1

Date Collected: 02/20/24 09:42

Matrix: Drinking Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 19:32
Total/NA	Prep	3510C			413924	JC	EET CAL 4	02/24/24 17:03
Total/NA	Analysis	8015B		1	414270	SP9M	EET CAL 4	02/26/24 22:11
Total/NA	Prep	EPA_625		1	O-44126_P			02/27/24 00:00
Total/NA	Analysis	625 PAH Physis LL (EAL) + TICs		1	O-44126	YC		03/25/24 22:35

Client Sample ID: AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-84152-2

Date Collected: 02/20/24 10:54

Matrix: Drinking Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 19:06
Total/NA	Prep	3510C			413924	JC	EET CAL 4	02/24/24 17:03
Total/NA	Analysis	8015B		1	414270	SP9M	EET CAL 4	02/26/24 22:32
Total/NA	Prep	EPA_625		1	O-44126_P			02/27/24 00:00
Total/NA	Analysis	625 PAH Physis LL (EAL) + TICs		1	O-44126	YC		03/26/24 00:24

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2

Lab Sample ID: 380-84152-3

Date Collected: 02/20/24 11:31

Matrix: Drinking Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 18:40
Total/NA	Prep	3510C			413924	JC	EET CAL 4	02/24/24 17:03
Total/NA	Analysis	8015B		1	414270	SP9M	EET CAL 4	02/26/24 22:53
Total/NA	Prep	EPA_625		1	O-44126_P			02/27/24 00:00
Total/NA	Analysis	625 PAH Physis LL (EAL) + TICs		1	O-44126	YC		03/26/24 02:12

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-84152-4

Date Collected: 02/20/24 10:26

Matrix: Drinking Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 18:14
Total/NA	Prep	3510C			413924	JC	EET CAL 4	02/24/24 17:03
Total/NA	Analysis	8015B		1	419748	SP9M	EET CAL 4	03/13/24 15:30
Total/NA	Prep	EPA_625		1	O-44126_P			02/27/24 00:00
Total/NA	Analysis	625 PAH Physis LL (EAL) + TICs		1	O-44126	YC		03/26/24 04:01

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Client Sample ID: TB MOANALUA WELLS

Lab Sample ID: 380-84152-5

Date Collected: 02/20/24 09:42

Matrix: Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 16:29

Client Sample ID: TB AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-84152-6

Date Collected: 02/20/24 10:54

Matrix: Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 16:55

Client Sample ID: TB AIEA WELLS PUMPS 1&2 (260)

Lab Sample ID: 380-84152-7

Date Collected: 02/20/24 11:31

Matrix: Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 17:21

Client Sample ID: TB HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-84152-8

Date Collected: 02/20/24 10:26

Matrix: Water

Date Received: 02/22/24 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	414522	A9VE	EET CAL 4	02/27/24 17:48

Laboratory References:

= Physis Environmental Laboratories, 1904 Wright Circle, Anaheim, CA 92806
EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Arizona	State	AZ0830	11-16-24
California	Los Angeles County Sanitation Districts	10109	08-01-24
California	State	3082	07-31-24
Kansas	NELAP	E-10420	08-01-24
Nevada	State	CA00111	07-31-24
Oregon	NELAP	4175	02-03-25
USDA	US Federal Programs	P330-22-00059	06-08-26
Washington	State	C916-18	10-11-24

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Method	Method Description	Protocol	Laboratory
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
625	EPA 625 Base/Neutral and Acid Organics i	EPA	
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4

Protocol References:

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

= Physis Environmental Laboratories, 1904 Wright Circle, Anaheim, CA 92806

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL [SUBCONTRACT]

Job ID: 380-84152-2
SDG: 625, 8015

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-84152-1	MOANALUA WELLS	Drinking Water	02/20/24 09:42	02/22/24 10:15	HI0000331
380-84152-2	AIEA GULCH WELLS PUMP 2	Drinking Water	02/20/24 10:54	02/22/24 10:15	HI0000331
380-84152-3	AIEA WELLS PUMPS 1&2 (260) P2	Drinking Water	02/20/24 11:31	02/22/24 10:15	HI0000331
380-84152-4	HALAWA WELLS UNITS 1 & 2 P1	Drinking Water	02/20/24 10:26	02/22/24 10:15	HI0000331
380-84152-5	TB MOANALUA WELLS	Water	02/20/24 09:42	02/22/24 10:15	
380-84152-6	TB AIEA GULCH WELLS PUMP 2	Water	02/20/24 10:54	02/22/24 10:15	
380-84152-7	TB AIEA WELLS PUMPS 1&2 (260)	Water	02/20/24 11:31	02/22/24 10:15	
380-84152-8	TB HALAWA WELLS UNITS 1 & 2 P1	Water	02/20/24 10:26	02/22/24 10:15	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

March 29, 2024

Rachelle Arada
 Eurofins Eaton Analytical
 750 Royal Oaks Drive
 Suite 100
 Monrovia, CA 91016-

Project Name: RED-HILL Project # 38001111 Job # 380-84152-1
 Physis Project ID: 1407003-488

Dear Rachelle,

Enclosed are the analytical results for samples submitted to PHYSIS Environmental Laboratories, Inc. (PHYSIS) on 2/23/2024. A total of 4 samples were received for analysis in accordance with the attached chain of custody (COC). Per the COC, the samples were analyzed for:

Organics
Polynuclear Aromatic Hydrocarbons by EPA 625.1
Disalicylidenepropanediamine by EPA 625.1
Dibenzo [a,l] Pyrene w/ PAHs by EPA 625.1

Analytical results in this report apply only to samples submitted to PHYSIS in accordance with the COC and are intended to be considered in their entirety.

Please feel free to contact me at any time with any questions. PHYSIS appreciates the opportunity to provide you with our analytical and support services.

Regards,
Rachel Hansen
 Rachel Hansen
 714 602-5320
 Extension 203
 rachelhansen@physislabs.com



PROJECT SAMPLE LIST

Eurofins Eaton Analytical

PHYSIS Project ID: 1407003-488

RED-HILL Project # 38001111 Job # 380-84152-1

Total Samples: 4

PHYSIS ID	Sample ID	Description	Date	Time	Matrix	Sample Type
116077	MOANALUA WELLS	380-84152-1	2/20/2024	9:42	Samplewater	Not Specified
116078	AIEA GULCH WELLS PUMP 2	380-84152-2	2/20/2024	10:54	Samplewater	Not Specified
116079	AIEA WELLS PUMPS 1&2 (260) P2	380-84152-3	2/20/2024	11:31	Samplewater	Not Specified
116080	HALAWA WELLS UNITS 1 & 2 P1	380-84152-4	2/20/2024	10:26	Samplewater	Not Specified

ABBREVIATIONS and ACRONYMS

QM	Quality Manual
QA	Quality Assurance
QC	Quality Control
MDL	method detection limit
RL	reporting limit
R1	project sample
R2	project sample replicate
MS1	matrix spike
MS2	matrix spike replicate
B1	procedural blank
B2	procedural blank replicate
BS1	blank spike
BS2	blank spike replicate
LCS1	laboratory control spike
LCS2	laboratory control spike replicate
LCM1	laboratory control material
LCM2	laboratory control material replicate
CRM1	certified reference material
CRM2	certified reference material replicate
RPD	relative percent difference
LMW	low molecular weight
HMW	high molecular weight

QUALITY ASSURANCE SUMMARY

LABORATORY BATCH: Physis' QM defines a laboratory batch as a group of 20 or fewer project samples of similar matrix, processed together under the same conditions and with the same reagents. QC samples are associated with each batch and were used to assess the validity of the sample analyses.

PROCEDURAL BLANK: Laboratory contamination introduced during method use is assessed through the preparation and analysis of procedural blanks is provided at a minimum frequency of one per batch.

ACCURACY: Accuracy of analytical measurements is the degree of closeness based on percent recovery calculations between measured values and the actual or true value and includes a combination of reproducibility error and systematic bias due to sampling and analytical operations. Accuracy of the project data was indicated by analysis of MS, BS, LCS, LCM, CRM, and/or surrogate spikes on a minimum frequency of one per batch. Physis' QM requires that 95% of the target compounds greater than 10 times the MDL be within the specified acceptance limits.

PRECISION: Precision is the agreement among a set of replicate measurements without assumption of knowledge of the true value and is based on RPD calculations between repeated values. Precision of the project data was determined by analysis of replicate MS₁/MS₂, BS₁/BS₂, LCS₁/LCS₂, LCM₁/LCM₂, CRM₁/CRM₂, surrogate spikes and/or replicate project sample analysis (R₁/R₂) on a minimum frequency of one per batch. Physis' QM requires that for 95% of the compounds greater than 10 times the MDL, the percent RPD should be within the specified acceptance range.

BLANK SPIKES: BS is the introduction of a known concentration of analyte into the procedural blank. BS demonstrates performance of the preparation and analytical methods on a clean matrix void of potential matrix related interferences. The BS is performed in laboratory deionized water, making these recoveries a better indicator of the efficiency of the laboratory method per se.

MATRIX SPIKES: MS is the introduction of a known concentration of analyte into a sample. MS samples demonstrate the effect a particular project sample matrix has on the accuracy of a measurement. Individually, MS samples also indicate the bias of analytical measurements due to chemical interferences inherent in the in the specific project sample spiked. Intrinsic target analyte concentration in the specific project sample can also significantly impact MS recovery.

CERTIFIED REFERENCE MATERIALS: CRMs are materials of various matrices for which analytical information has been determined and certified by a recognized authority. These are used to provide a quantitative assessment of the accuracy of an analytical method. CRMs provide evidence that the laboratory preparation and analysis produces results that are comparable to those obtained by an independent organization.

LABORATORY CONTROL MATERIAL: LCM is provided because a suitable natural seawater CRM is not available and can be used to indicate accuracy of the method. Physis' internal LCM is seawater collected at ~800 meters in the Southern California San Pedro Basin and can be used as a reference for background concentrations in clean, natural seawater for comparison to project samples.

LABORATORY CONTROL SPIKES: LCS is the introduction of a known concentration of analyte into Physis' LCM. LCS samples were employed to assess the effect the seawater matrix has on the accuracy of a measurement. LCS also indicate the bias of this method due to chemical interferences inherent in the in the seawater matrix. Intrinsic LCM concentration can also significantly impact LCS recovery.

SURROGATES: A surrogate is a pure analyte unlikely to be found in any project sample, behaves similarly to

the target analyte and most often used with organic analytical procedures. Surrogates are added in known concentration to all samples and are measured to indicate overall efficiency of the method including processing and analyses.

HOLDING TIME: Method recommended holding times are the length of time a project sample can be stored under specific conditions after collection and prior to analysis without significantly affecting the analyte's concentration. Holding times can be extended if preservation techniques are employed to reduce biodegradation, volatilization, oxidation, sorption, precipitation, and other physical and chemical processes.

SAMPLE STORAGE/RETENTION: In order to maintain chemical integrity prior to analysis, all samples submitted to Physis are refrigerated (liquids) or frozen (solids) upon receipt unless otherwise recommended by applicable methods. Solid samples are retained for 1 year from collection while liquid samples are retained until method recommended holding times elapse.

TOTAL/DISSOLVED FRACTION: In some instances, the results for the dissolved fraction may be higher than the total fraction for a particular analyte (e.g. trace metals). This is typically caused by the analytical variation for each result and indicates that the target analyte is primarily in the dissolved phase, within the sample.

PHYSIS QUALIFIER CODES

CODE	DEFINITION
#	see Case Narrative
ND	analyte not detected at or above the MDL
B	analyte was detected in the procedural blank greater than 10 times the MDL
E	analyte concentration exceeds the upper limit of the linear calibration range, reported value is estimated
H	sample received and/or analyzed past the recommended holding time
J	analyte was detected at a concentration below the RL and above the MDL, reported value is estimated
N	insufficient sample, analysis could not be performed
M	analyte was outside the specified accuracy and/or precision acceptance limits due to matrix interference. The associated B/BS were within limits, therefore the sample data was reported without further clarification
SH	analyte concentration in the project sample exceeded the spike concentration, therefore accuracy and/or precision acceptance limits do not apply
SL	analyte results were lower than 10 times the MDL, therefore accuracy and/or precision acceptance limits do not apply
NH	project sample was heterogeneous and sample homogeneity could not be readily achieved using routine laboratory practices, therefore accuracy and/or precision acceptance limits do not apply
Q	analyte was outside the specified QAPP acceptance limits for precision and/or accuracy but within Physis derived acceptance limits, therefore the sample data was reported without further clarification
R	Physis' QM allows for 5% of the target compounds greater than 10 times the MDL to be outside the specified acceptance limits for precision and/or accuracy. This is often due to random error and does not indicate any significant problems with the analysis of these project samples

CASE NARRATIVE

QUALIFIER NOTES

In addition to the use of analyte specific Physis Qualifier Codes where applicable, the following were also noted.

ND

MDL is listed due to report format restrictions; it is not used in reporting. Analytical results reported are ND at the RL.

BIANALYTICALS

REPORT

TERRA AURA
ENVIRONMENTAL LABORATORIES, INC.

Innovative Solutions for Nature

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Base/Neutral Extractable Compounds

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 116077-R1 MOANALUA WELLS 380-84152-1 Matrix: Samplewater											
Disalicylideneprapanediamine	EPA 625.1	µg/L	ND	1	0.05	0.1	Total		O-44126	27-Feb-24	25-Mar-24
Sample ID: 116078-R1 AIEA GULCH WELLS PUMP 2 380-8 Matrix: Samplewater											
Disalicylideneprapanediamine	EPA 625.1	µg/L	ND	1	0.05	0.1	Total		O-44126	27-Feb-24	26-Mar-24
Sample ID: 116079-R1 AIEA WELLS PUMPS 1&2 (260) P2 3 Matrix: Samplewater											
Disalicylideneprapanediamine	EPA 625.1	µg/L	ND	1	0.05	0.1	Total		O-44126	27-Feb-24	26-Mar-24
Sample ID: 116080-R1 HALAWA WELLS UNITS 1 & 2 P1 38 Matrix: Samplewater											
Disalicylideneprapanediamine	EPA 625.1	µg/L	ND	1	0.05	0.1	Total		O-44126	27-Feb-24	26-Mar-24



Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 116077-R1	MOANALUA WELLS 380-84152-1	Matrix: Samplewater					Sampled: 20-Feb-24 9:42			Received: 23-Feb-24	
(d10-Acenaphthene)	EPA 625.1	% Recovery	73	1			Total		O-44126	27-Feb-24	25-Mar-24
(d10-Phenanthrene)	EPA 625.1	% Recovery	82	1			Total		O-44126	27-Feb-24	25-Mar-24
(d12-Chrysene)	EPA 625.1	% Recovery	92	1			Total		O-44126	27-Feb-24	25-Mar-24
(d12-Perylene)	EPA 625.1	% Recovery	90	1			Total		O-44126	27-Feb-24	25-Mar-24
(d8-Naphthalene)	EPA 625.1	% Recovery	58	1			Total		O-44126	27-Feb-24	25-Mar-24
1-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
1-Methylphenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
2,3,5-Trimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
2,6-Dimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
2-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Acenaphthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Acenaphthylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Benz[a]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Benzo[a]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Benzo[b]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Benzo[e]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Benzo[g,h,i]perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Benzo[k]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Biphenyl	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Chrysene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Dibenz[a,h]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Dibenzo[a,l]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Dibenzothiophene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24

Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Fluorene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Indeno[1,2,3-cd]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Naphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Phenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24
Pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	25-Mar-24



Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 116078-R1	AIEA GULCH WELLS PUMP 2 380-8 Matrix: Samplewater						Sampled:	20-Feb-24 10:54	Received:	23-Feb-24	
(d10-Acenaphthene)	EPA 625.1	% Recovery	65	1			Total		O-44126	27-Feb-24	26-Mar-24
(d10-Phenanthrene)	EPA 625.1	% Recovery	73	1			Total		O-44126	27-Feb-24	26-Mar-24
(d12-Chrysene)	EPA 625.1	% Recovery	79	1			Total		O-44126	27-Feb-24	26-Mar-24
(d12-Perylene)	EPA 625.1	% Recovery	90	1			Total		O-44126	27-Feb-24	26-Mar-24
(d8-Naphthalene)	EPA 625.1	% Recovery	53	1			Total		O-44126	27-Feb-24	26-Mar-24
1-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
1-Methylphenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2,3,5-Trimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2,6-Dimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Acenaphthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Acenaphthylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benz[a]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[a]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[b]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[e]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[g,h,i]perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[k]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Biphenyl	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Chrysene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenz[a,h]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenzo[a,l]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenzothiophene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24

Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Fluorene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Indeno[1,2,3-cd]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Naphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Phenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24



Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 116079-R1	AIEA WELLS PUMPS 1&2 (260) P2 3 Matrix: Samplewater						Sampled:	20-Feb-24	11:31	Received:	23-Feb-24
(d10-Acenaphthene)	EPA 625.1	% Recovery	73	1			Total		O-44126	27-Feb-24	26-Mar-24
(d10-Phenanthrene)	EPA 625.1	% Recovery	78	1			Total		O-44126	27-Feb-24	26-Mar-24
(d12-Chrysene)	EPA 625.1	% Recovery	108	1			Total		O-44126	27-Feb-24	26-Mar-24
(d12-Perylene)	EPA 625.1	% Recovery	90	1			Total		O-44126	27-Feb-24	26-Mar-24
(d8-Naphthalene)	EPA 625.1	% Recovery	59	1			Total		O-44126	27-Feb-24	26-Mar-24
1-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
1-Methylphenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2,3,5-Trimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2,6-Dimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Acenaphthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Acenaphthylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benz[a]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[a]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[b]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[e]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[g,h,i]perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[k]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Biphenyl	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Chrysene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenz[a,h]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenzo[a,l]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenzothiophene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24

Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Fluorene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Indeno[1,2,3-cd]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Naphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Phenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24



Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 116080-R1	HALAWA WELLS UNITS 1 & 2 P1 38	Matrix: Samplewater					Sampled:	20-Feb-24 10:26		Received:	23-Feb-24
(d10-Acenaphthene)	EPA 625.1	% Recovery	85	1			Total		O-44126	27-Feb-24	26-Mar-24
(d10-Phenanthrene)	EPA 625.1	% Recovery	89	1			Total		O-44126	27-Feb-24	26-Mar-24
(d12-Chrysene)	EPA 625.1	% Recovery	122	1			Total		O-44126	27-Feb-24	26-Mar-24
(d12-Perylene)	EPA 625.1	% Recovery	94	1			Total		O-44126	27-Feb-24	26-Mar-24
(d8-Naphthalene)	EPA 625.1	% Recovery	70	1			Total		O-44126	27-Feb-24	26-Mar-24
1-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
1-Methylphenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2,3,5-Trimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2,6-Dimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
2-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Acenaphthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Acenaphthylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benz[a]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[a]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[b]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[e]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[g,h,i]perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Benzo[k]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Biphenyl	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Chrysene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenz[a,h]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenzo[a,l]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Dibenzothiophene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24

Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Fluorene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Indeno[1,2,3-cd]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Naphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Phenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24
Pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-44126	27-Feb-24	26-Mar-24



QUALITY CONTROL REPORT

TERRA ENVIRONMENTAL LABORATORIES, INC. AURA

Innovative Solutions for Nature

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

Base/Neutral Extractable Compounds

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODE
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Sample ID: 116076-B1		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:			
		Method: EPA 625.1			Batch ID: O-44126			Prepared: 27-Feb-24		Analyzed: 25-Mar-24			
Disalicylidenepranediamin	Total	ND	1	0.05	0.1	µg/L							
Sample ID: 116076-BS1		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:			
		Method: EPA 625.1			Batch ID: O-44126			Prepared: 27-Feb-24		Analyzed: 25-Mar-24			
Disalicylidenepranediamin	Total	47.6	1	0.05	0.1	µg/L	50	0	95	50 - 150%	PASS		
Sample ID: 116076-BS2		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:			
		Method: EPA 625.1			Batch ID: O-44126			Prepared: 27-Feb-24		Analyzed: 25-Mar-24			
Disalicylidenepranediamin	Total	49.6	1	0.05	0.1	µg/L	50	0	99	50 - 150%	PASS	4	30 PASS

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY	PRECISION	QA CODEc
							LEVEL	RESULT	% LIMITS	% LIMITS	
Sample ID: 116076-B1		QAQC Procedural Blank			Matrix: BlankMatrix		Sampled:		Received:		
	Method: EPA 625.1					Batch ID: O-44126	Prepared: 27-Feb-24	Analyzed: 25-Mar-24			
(d10-Acenaphthene)	Total	83	1			% Recovery	100	83	27 - 133%	PASS	
(d10-Phenanthrene)	Total	86	1			% Recovery	100	86	43 - 129%	PASS	
(d12-Chrysene)	Total	91	1			% Recovery	100	91	52 - 144%	PASS	
(d12-Perylene)	Total	88	1			% Recovery	100	88	36 - 161%	PASS	
(d8-Naphthalene)	Total	68	1			% Recovery	100	68	25 - 125%	PASS	
1-Methylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
1-Methylphenanthrene	Total	ND	1	0.001	0.005	µg/L					
2,3,5-Trimethylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
2,6-Dimethylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
2-Methylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
Acenaphthene	Total	ND	1	0.001	0.005	µg/L					
Acenaphthylene	Total	ND	1	0.001	0.005	µg/L					
Anthracene	Total	ND	1	0.001	0.005	µg/L					
Benz[a]anthracene	Total	ND	1	0.001	0.005	µg/L					
Benzo[a]pyrene	Total	ND	1	0.001	0.005	µg/L					
Benzo[b]fluoranthene	Total	ND	1	0.001	0.005	µg/L					
Benzo[e]pyrene	Total	ND	1	0.001	0.005	µg/L					
Benzo[g,h,i]perylene	Total	ND	1	0.001	0.005	µg/L					
Benzo[k]fluoranthene	Total	ND	1	0.001	0.005	µg/L					
Biphenyl	Total	ND	1	0.001	0.005	µg/L					
Chrysene	Total	ND	1	0.001	0.005	µg/L					
Dibenz[a,h]anthracene	Total	ND	1	0.001	0.005	µg/L					
Dibenzo[a,i]pyrene	Total	ND	1	0.001	0.005	µg/L					

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODE
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Dibenzothiophene	Total	ND	1	0.001	0.005	µg/L							
Fluoranthene	Total	ND	1	0.001	0.005	µg/L							
Fluorene	Total	ND	1	0.001	0.005	µg/L							
Indeno[1,2,3-cd]pyrene	Total	ND	1	0.001	0.005	µg/L							
Naphthalene	Total	ND	1	0.001	0.005	µg/L							
Perylene	Total	ND	1	0.001	0.005	µg/L							
Phenanthrene	Total	ND	1	0.001	0.005	µg/L							
Pyrene	Total	ND	1	0.001	0.005	µg/L							

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION	QA CODEc
							LEVEL	RESULT	%	LIMITS	%	LIMITS
Sample ID: 116076-BS1		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:		
Method: EPA 625.1		Batch ID: O-44126			Prepared: 27-Feb-24		Analyzed: 25-Mar-24					
(d10-Acenaphthene)	Total	79	1			% Recovery	100	0	79	27 - 133%	PASS	
(d10-Phenanthrene)	Total	85	1			% Recovery	100	0	85	43 - 129%	PASS	
(d12-Chrysene)	Total	90	1			% Recovery	100	0	90	52 - 144%	PASS	
(d12-Perylene)	Total	88	1			% Recovery	100	0	88	36 - 161%	PASS	
(d8-Naphthalene)	Total	65	1			% Recovery	100	0	65	25 - 125%	PASS	
1-Methylnaphthalene	Total	0.398	1	0.001	0.005	µg/L	0.5	0	80	31 - 128%	PASS	
1-Methylphenanthrene	Total	0.539	1	0.001	0.005	µg/L	0.5	0	108	66 - 127%	PASS	
2,3,5-Trimethylnaphthalene	Total	0.482	1	0.001	0.005	µg/L	0.5	0	96	55 - 122%	PASS	
2,6-Dimethylnaphthalene	Total	0.433	1	0.001	0.005	µg/L	0.5	0	87	48 - 120%	PASS	
2-Methylnaphthalene	Total	0.404	1	0.001	0.005	µg/L	0.5	0	81	47 - 130%	PASS	
Acenaphthene	Total	0.453	1	0.001	0.005	µg/L	0.5	0	91	53 - 131%	PASS	
Acenaphthylene	Total	0.588	1	0.001	0.005	µg/L	0.5	0	118	43 - 140%	PASS	
Anthracene	Total	0.509	1	0.001	0.005	µg/L	0.5	0	102	58 - 135%	PASS	
Benz[a]anthracene	Total	0.591	1	0.001	0.005	µg/L	0.5	0	118	55 - 145%	PASS	
Benzo[a]pyrene	Total	0.527	1	0.001	0.005	µg/L	0.5	0	105	51 - 143%	PASS	
Benzo[b]fluoranthene	Total	0.558	1	0.001	0.005	µg/L	0.5	0	112	46 - 165%	PASS	
Benzo[e]pyrene	Total	0.457	1	0.001	0.005	µg/L	0.5	0	91	42 - 152%	PASS	
Benzo[g,h,i]perylene	Total	0.521	1	0.001	0.005	µg/L	0.5	0	104	63 - 133%	PASS	
Benzo[k]fluoranthene	Total	0.513	1	0.001	0.005	µg/L	0.5	0	103	56 - 145%	PASS	
Biphenyl	Total	0.408	1	0.001	0.005	µg/L	0.5	0	82	56 - 119%	PASS	
Chrysene	Total	0.528	1	0.001	0.005	µg/L	0.5	0	106	56 - 141%	PASS	
Dibenz[a,h]anthracene	Total	0.728	1	0.001	0.005	µg/L	0.5	0	146	55 - 150%	PASS	
Dibenzo[a,l]pyrene	Total	0.435	1	0.001	0.005	µg/L	0.5	0	87	50 - 150%	PASS	

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODE _c
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Dibenzothiophene	Total	0.48	1	0.001	0.005	µg/L	0.5	0	96	46 - 126%	PASS		
Fluoranthene	Total	0.541	1	0.001	0.005	µg/L	0.5	0	108	60 - 146%	PASS		
Fluorene	Total	0.491	1	0.001	0.005	µg/L	0.5	0	98	58 - 131%	PASS		
Indeno[1,2,3-cd]pyrene	Total	0.746	1	0.001	0.005	µg/L	0.5	0	149	50 - 151%	PASS		
Naphthalene	Total	0.366	1	0.001	0.005	µg/L	0.5	0	73	41 - 126%	PASS		
Perylene	Total	0.489	1	0.001	0.005	µg/L	0.5	0	98	48 - 141%	PASS		
Phenanthrene	Total	0.471	1	0.001	0.005	µg/L	0.5	0	94	67 - 127%	PASS		
Pyrene	Total	0.544	1	0.001	0.005	µg/L	0.5	0	109	54 - 156%	PASS		

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc	
							LEVEL	RESULT	%	LIMITS	%	LIMITS		
Sample ID: 116076-BS2		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:			Received:			
		Method: EPA 625.1			Batch ID: O-44126			Prepared: 27-Feb-24			Analyzed: 25-Mar-24			
(d10-Acenaphthene)	Total	84	1			% Recovery	100	0	84	27 - 133%	PASS	6	30	PASS
(d10-Phenanthrene)	Total	88	1			% Recovery	100	0	88	43 - 129%	PASS	3	30	PASS
(d12-Chrysene)	Total	94	1			% Recovery	100	0	94	52 - 144%	PASS	4	30	PASS
(d12-Perylene)	Total	91	1			% Recovery	100	0	91	36 - 161%	PASS	3	30	PASS
(d8-Naphthalene)	Total	68	1			% Recovery	100	0	68	25 - 125%	PASS	5	30	PASS
1-Methylnaphthalene	Total	0.411	1	0.001	0.005	µg/L	0.5	0	82	31 - 128%	PASS	2	30	PASS
1-Methylphenanthrene	Total	0.535	1	0.001	0.005	µg/L	0.5	0	107	66 - 127%	PASS	1	30	PASS
2,3,5-Trimethylnaphthalene	Total	0.495	1	0.001	0.005	µg/L	0.5	0	99	55 - 122%	PASS	3	30	PASS
2,6-Dimethylnaphthalene	Total	0.454	1	0.001	0.005	µg/L	0.5	0	91	48 - 120%	PASS	4	30	PASS
2-Methylnaphthalene	Total	0.417	1	0.001	0.005	µg/L	0.5	0	83	47 - 130%	PASS	2	30	PASS
Acenaphthene	Total	0.473	1	0.001	0.005	µg/L	0.5	0	95	53 - 131%	PASS	4	30	PASS
Acenaphthylene	Total	0.599	1	0.001	0.005	µg/L	0.5	0	120	43 - 140%	PASS	2	30	PASS
Anthracene	Total	0.53	1	0.001	0.005	µg/L	0.5	0	106	58 - 135%	PASS	4	30	PASS
Benz[a]anthracene	Total	0.568	1	0.001	0.005	µg/L	0.5	0	114	55 - 145%	PASS	3	30	PASS
Benzo[a]pyrene	Total	0.534	1	0.001	0.005	µg/L	0.5	0	107	51 - 143%	PASS	2	30	PASS
Benzo[b]fluoranthene	Total	0.538	1	0.001	0.005	µg/L	0.5	0	108	46 - 165%	PASS	4	30	PASS
Benzo[e]pyrene	Total	0.464	1	0.001	0.005	µg/L	0.5	0	93	42 - 152%	PASS	2	30	PASS
Benzo[g,h,i]perylene	Total	0.529	1	0.001	0.005	µg/L	0.5	0	106	63 - 133%	PASS	2	30	PASS
Benzo[k]fluoranthene	Total	0.497	1	0.001	0.005	µg/L	0.5	0	99	56 - 145%	PASS	4	30	PASS
Biphenyl	Total	0.424	1	0.001	0.005	µg/L	0.5	0	85	56 - 119%	PASS	4	30	PASS
Chrysene	Total	0.526	1	0.001	0.005	µg/L	0.5	0	105	56 - 141%	PASS	1	30	PASS
Dibenz[a,h]anthracene	Total	0.73	1	0.001	0.005	µg/L	0.5	0	146	55 - 150%	PASS	0	30	PASS
Dibenzo[a,l]pyrene	Total	0.379	1	0.001	0.005	µg/L	0.5	0	76	50 - 150%	PASS	13	30	PASS

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODE _c	
							LEVEL	RESULT	%	LIMITS	%	LIMITS		
Dibenzothiophene	Total	0.493	1	0.001	0.005	µg/L	0.5	0	99	46 - 126%	PASS	3	30	PASS
Fluoranthene	Total	0.537	1	0.001	0.005	µg/L	0.5	0	107	60 - 146%	PASS	1	30	PASS
Fluorene	Total	0.5	1	0.001	0.005	µg/L	0.5	0	100	58 - 131%	PASS	2	30	PASS
Indeno[1,2,3-cd]pyrene	Total	0.749	1	0.001	0.005	µg/L	0.5	0	150	50 - 151%	PASS	1	30	PASS
Naphthalene	Total	0.38	1	0.001	0.005	µg/L	0.5	0	76	41 - 126%	PASS	4	30	PASS
Perylene	Total	0.494	1	0.001	0.005	µg/L	0.5	0	99	48 - 141%	PASS	1	30	PASS
Phenanthrene	Total	0.488	1	0.001	0.005	µg/L	0.5	0	98	67 - 127%	PASS	4	30	PASS
Pyrene	Total	0.54	1	0.001	0.005	µg/L	0.5	0	108	54 - 156%	PASS	1	30	PASS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

PHYSIS

TENTATIVELY IDENTIFIED COMPOUNDS

ENVIRONMENTAL LABORATORIES, INC.

Innovative Solutions for Nature

Sample ID: 116080

Retention Time	Area (% of total)	Concentration (ng/L)	Library/ID	Cas Number	Match Quality (%)
35.5498	5.4266	1111	Anthracene-D10-	1719-06-8	98
10.4905	1.2071	247	Hydroperoxide, 1-ethylbutyl	24254-56-6	92
11.2100	0.5380	110	Cyclopropane, 1,1,2,3-tetramethyl-	74752-93-5	90

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Sample ID: 116077

Retention Time	Area (% of total)	Concentration (ng/L)	Library/ID	Cas Number	Match Quality (%)
35.5542	4.1223	1111	Anthracene-D10-	1719-06-8	98
10.4886	0.7473	201	Hydroperoxide, 1-ethylbutyl	24254-56-6	91
63.6326	0.5743	155	Heneicosane	629-94-7	96
25.3651	0.4267	115	Benzoic acid, 4-ethoxy-, ethyl ester	23676-09-7	99

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Sample ID: Lab Blank B1_44126

Retention Time	Area (% of total)	Concentration (ng/L)	Library/ID	Cas Number	Match Quality (%)
35.5530	4.1765	1111	Anthracene-D10-	1719-06-8	98
10.4899	0.6768	180	Hydroperoxide, 1-ethylbutyl	24254-56-6	90
25.3635	0.4394	117	Benzoic acid, 4-ethoxy-, ethyl ester	23676-09-7	98
11.2524	0.4313	115	3,3-Diethoxy-1-propyne	10160-87-9	83

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Sample ID: 116079

Retention Time	Area (% of total)	Concentration (ng/L)	Library/ID	Cas Number	Match Quality (%)
35.5504	4.7647	1111	Anthracene-D10-	1719-06-8	98
10.4907	1.2657	295	Hydroperoxide, 1-ethylbutyl	24254-56-6	91
11.2105	0.6006	140	1-Butene, 2,3,3-trimethyl-	594-56-9	89
11.2105	0.5670	132	Cyclopropane, 1,1,2,3-tetramethyl-	74752-93-5	87
25.3620	0.4432	103	Benzoic acid, 4-ethoxy-, ethyl ester	23676-09-7	99

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Sample ID: 116078

Retention Time	Area (% of total)	Concentration (ng/L)	Library/ID	Cas Number	Match Quality (%)
35.9704	4.3069	1111	Anthracene-D10-	1517-22-2	96
10.4813	0.8266	213	Hydroperoxide, 1-ethylbutyl	24254-56-6	91
10.0900	0.6740	174	Ethane, 1,1,2,2-tetrachloro-	79-34-5	91

Concentration estimated using the response for Anthracene-d10


- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

PERFORMANCE CHAIN OF CUSTODY

TERRA ENVIRONMENTAL LABORATORIES, INC. AURA

Innovative Solutions for Nature

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16



Project Iteration ID: 1407003-488
 Client Name: Eurofins Eaton Analytical
 Project Name: RED-HILL Project # 38001111 Job # 380-84152-1
 COC Page Number: 2 of 2
 Bottle Label Color: NA

Sample Receipt Summary

Receiving Info

- Initials Received By: CR
- Date Received: 2/23/24
- Time Received: 1038
- Client Name: EUROFINS
- Courier Information: (Please circle)
 - Client
 - UPS
 - Area Fast
 - DRS
 - FedEx
 - GSO/GLS
 - Ontrac
 - PAMS
 - PHYSIS Driver:
 - Start Time: _____
 - End Time: _____
 - Total Mileage: _____
 - Number of Pickups: _____
- Container Information: (Please put the # of containers or circle none)
 - Cooler
 - Styrofoam Cooler
 - Boxes
 - None
 - Carboy(s)
 - Carboy Trash Can(s)
 - Carboy Cap(s)
 - Other _____
- What type of ice was used: (Please circle any that apply)
 - Wet Ice
 - Blue Ice
 - Dry Ice
 - Water
 - None
- Randomly Selected Samples Temperature (°C): 1.2
 Used I/R Thermometer # _____

Inspection Info

- Initials Inspected By: CR

Sample Integrity Upon Receipt:

- COC(s) included and completely filled out..... Yes / No
- All sample containers arrived intact..... Yes / No
- All samples listed on COC(s) are present..... Yes / No
- Information on containers consistent with information on COC(s)..... Yes / No
- Correct containers and volume for all analyses indicated..... Yes / No
- All samples received within method holding time..... Yes / No
- Correct preservation used for all analyses indicated..... Yes / No
- Name of sampler included on COC(s)..... Yes / No

Notes:

Monrovia, CA (Suite 100)
 750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone (626) 386-1100



Chain of Custody Record

Client Information Client Contact: Dr. Ron Fenstermacher Company: City & County of Honolulu Address: 630 South Berelania Street, Chemistry Lab City: Honolulu State/Zip: HI, 96843 Phone: 808-748-5091 (tel) Email: rfenstermacher@hbw.org Project Name: RED-HILL/HBWS sites Event Desc. RUSH Weekly Red Hill Site:		Lab PM: Arada, Rachelle E-Mail: Rachelle.Arada@et.euronisus.com Carrier Tracking No(s): 380-27984-2757 2 State of Origin:	
Due Date Requested: TAT Requested (days): Compliance Project: <input type="checkbox"/> No PO #: C20525101 exp 05312023 WO #: Project #: 38001111 SSO#:		Analysis Requested 537 1, DW, PREC - 537 1 Full List 525 2, PREC - (MOD) 525plus PLUS TICs C18 8015B_DRO_LL_CS - HNL Ranges C10-C24/C24-C36/C8- 8015B_GRO_LL - (MOD) GRO SUBCONTRACT - 625 PAH Physis LL (EAL) + TICs Perform MS/MSD (Yes or No) <input checked="" type="checkbox"/>	
Sample Identification MOANALUA WELLS AIEA GULCH WELLS PUMP2 AIEA WELLS PUMPS 1&2 (260) P2 HALAWA WELLS UNITS 1&2 P1 TB MOANALUA WELLS TB AIEA GULCH WELLS PUMP2 TB AIEA WELLS PUMPS 1&2 (260) TB HALAWA WELLS UNITS 1&2		Total Number of Containers chlorinated chlorinated 100% of 4 APPROVED BOTTLES CP 02/22/2024	
Sample Date 20-Feb-2024 20-Feb-2024 20-Feb-2024 20-Feb-2024 20-Feb-2024 20-Feb-2024 20-Feb-2024 20-Feb-2024	Sample Time 0942 1054 1131 1026 0942 1054 1131 1026	Sample Type (C=Comp, G=grab) G G G G G G G G	Matrix (W=water, S=solid, O=soil, B=biological, T=tissue, A=air) Water Water Water Water Water Water Water Water
Possible Hazard Identification <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological Deliverable Requested I, II, III, IV, Other (specify)		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months Special Instructions/OC Requirements	
Empty Kit Relinquished by Relinquished: [Redacted] Relinquished by: Relinquished by:		Method of Shipment: 1ED Ex Date/Time: 02/22/2024 10:15 Date/Time:	
Custody Seals Intact <input type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No (75A) 021° 01' = 2.0° / 029° 01' = 2.8° GEL-FROZEN	



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-84152-2

SDG Number: 625, 8015

Login Number: 84152

List Number: 1

Creator: Elyas, Matthew

List Source: Eurofins Eaton Analytical Pomona

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	False	Refer to NCM for affected item(s).
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-84152-2

SDG Number: 625, 8015

Login Number: 84152

List Number: 2

Creator: Khana, Piyush

List Source: Eurofins Calscience

List Creation: 02/23/24 01:58 PM

Question	Answer	Comment
Radioactivity wasn't checked or is < /= background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	1.5
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

