

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 4/2/2024 10:45:35 AM

JOB DESCRIPTION

RED-HILL
525.2
RUSH Weekly Red Hill

JOB NUMBER

380-88933-1

Eurofins Eaton Analytical Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Generated
4/2/2024 10:45:35 AM

Authorized for release by
Rachelle Arada, Project Manager
Rachelle.Arada@et.eurofinsus.com
(626)386-1106



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	15
Surrogate Summary	17
QC Sample Results	18
QC Association Summary	27
Lab Chronicle	28
Certification Summary	29
Method Summary	31
Sample Summary	32
Chain of Custody	33
Receipt Checklists	34

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Qualifiers

GC/MS Semi VOA

Qualifier	Qualifier Description
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-88933-1

Job ID: 380-88933-1

Eurofins Eaton Analytical Pomona

Job Narrative 380-88933-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 3/28/2024 10:44 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 0.6°C and 1.3°C.

GC/MS Semi VOA

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: MOANALUA WELLS
PWSID Number: HI0000331

Lab Sample ID: 380-88933-1

No Detections.

Client Sample ID: AIEA GULCH WELLS PUMP 2
PWSID Number: HI0000331

Lab Sample ID: 380-88933-2

No Detections.

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2
PWSID Number: HI0000331

Lab Sample ID: 380-88933-3

No Detections.

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1
PWSID Number: HI0000331

Lab Sample ID: 380-88933-4

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-88933-1
 SDG: 525.2

Client Sample ID: MOANALUA WELLS

Lab Sample ID: 380-88933-1

Date Collected: 03/25/24 09:38

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
2,4'-DDD	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
2,4'-DDE	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
2,4'-DDT	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
2,4-Dinitrotoluene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
2,6-Dinitrotoluene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
2-Methylnaphthalene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
4,4'-DDD	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
4,4'-DDE	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
4,4'-DDT	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Acenaphthene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Acenaphthylene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Acetochlor	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Alachlor	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
alpha-BHC	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
alpha-Chlordane	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Anthracene	<0.019	F1	0.019	ug/L		04/01/24 07:42	04/01/24 15:01	1
Atrazine	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Benz(a)anthracene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Benzo[a]pyrene	<0.019		0.019	ug/L		04/01/24 07:42	04/01/24 15:01	1
Benzo[b]fluoranthene	<0.019		0.019	ug/L		04/01/24 07:42	04/01/24 15:01	1
Benzo[g,h,i]perylene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Benzo[k]fluoranthene	<0.019		0.019	ug/L		04/01/24 07:42	04/01/24 15:01	1
beta-BHC	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Bis(2-ethylhexyl) phthalate	<0.58		0.58	ug/L		04/01/24 07:42	04/01/24 15:01	1
Bromacil	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Butachlor	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Butylbenzylphthalate	<0.48		0.48	ug/L		04/01/24 07:42	04/01/24 15:01	1
Chlorobenzilate	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Chloroneb	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Chlorothalonil (Draconil, Bravo)	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Chlorpyrifos	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Chrysene	<0.019		0.019	ug/L		04/01/24 07:42	04/01/24 15:01	1
delta-BHC	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Di(2-ethylhexyl)adipate	<0.58		0.58	ug/L		04/01/24 07:42	04/01/24 15:01	1
Dibenz(a,h)anthracene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Diclorvos (DDVP)	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Dieldrin	<0.19		0.19	ug/L		04/01/24 07:42	04/01/24 15:01	1
Diethylphthalate	<0.48		0.48	ug/L		04/01/24 07:42	04/01/24 15:01	1
Dimethylphthalate	<0.48		0.48	ug/L		04/01/24 07:42	04/01/24 15:01	1
Di-n-butyl phthalate	<0.97		0.97	ug/L		04/01/24 07:42	04/01/24 15:01	1
Di-n-octyl phthalate	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Endosulfan I (Alpha)	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Endosulfan II (Beta)	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Endosulfan sulfate	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Endrin	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Endrin aldehyde	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
EPTC	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Fluoranthene	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: MOANALUA WELLS

Lab Sample ID: 380-88933-1

Date Collected: 03/25/24 09:38

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
gamma-Chlordane	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Heptachlor	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:01	1
Heptachlor epoxide (isomer B)	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Hexachlorobenzene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Hexachlorocyclopentadiene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Indeno[1,2,3-cd]pyrene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Isophorone	<0.48		0.48	ug/L		04/01/24 07:42	04/01/24 15:01	1
Lindane	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:01	1
Malathion	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Methoxychlor	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Metolachlor	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Molinate	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Naphthalene	<0.29		0.29	ug/L		04/01/24 07:42	04/01/24 15:01	1
Parathion	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Pendimethalin (Penoxaline)	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Phenanthrene	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:01	1
Propachlor	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Pyrene	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Simazine	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Terbacil	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Terbutylazine	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1
Thiobencarb	<0.19		0.19	ug/L		04/01/24 07:42	04/01/24 15:01	1
Total Permethrin (mixed isomers)	<0.19		0.19	ug/L		04/01/24 07:42	04/01/24 15:01	1
trans-Nonachlor	<0.048		0.048	ug/L		04/01/24 07:42	04/01/24 15:01	1
Trifluralin	<0.097		0.097	ug/L		04/01/24 07:42	04/01/24 15:01	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	04/01/24 07:42	04/01/24 15:01	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	102		70 - 130	04/01/24 07:42	04/01/24 15:01	1
Perylene-d12	87		70 - 130	04/01/24 07:42	04/01/24 15:01	1
Triphenylphosphate	105		70 - 130	04/01/24 07:42	04/01/24 15:01	1

Client Sample ID: AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-88933-2

Date Collected: 03/25/24 10:35

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
2,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
2,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
2,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
2-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
4,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
4,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-88933-2

Date Collected: 03/25/24 10:35

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
4,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Acenaphthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Acenaphthylene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Acetochlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Alachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
alpha-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
alpha-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Anthracene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:21	1
Atrazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Benz(a)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Benzo[a]pyrene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:21	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:21	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:21	1
beta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 15:21	1
Bromacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Butachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Butylbenzylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:21	1
Chlorobenzilate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Chloroneb	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Chlorpyrifos	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Chrysene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:21	1
delta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 15:21	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Dieldrin	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 15:21	1
Diethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:21	1
Dimethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:21	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		04/01/24 07:42	04/01/24 15:21	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Endosulfan sulfate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Endrin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Endrin aldehyde	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
EPTC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Fluoranthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Fluorene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
gamma-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Heptachlor	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:21	1
Heptachlor epoxide (isomer B)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Hexachlorobenzene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Isophorone	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:21	1
Lindane	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:21	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-88933-2

Date Collected: 03/25/24 10:35

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Malathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Methoxychlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Metolachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Molinate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Naphthalene	<0.29		0.29	ug/L		04/01/24 07:42	04/01/24 15:21	1
Parathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Phenanthrene	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:21	1
Propachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Simazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Terbacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Terbutylazine	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1
Thiobencarb	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 15:21	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 15:21	1
trans-Nonachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:21	1
Trifluralin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:21	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	04/01/24 07:42	04/01/24 15:21	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	103		70 - 130	04/01/24 07:42	04/01/24 15:21	1
Perylene-d12	92		70 - 130	04/01/24 07:42	04/01/24 15:21	1
Triphenylphosphate	98		70 - 130	04/01/24 07:42	04/01/24 15:21	1

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2

Lab Sample ID: 380-88933-3

Date Collected: 03/25/24 11:02

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
2,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
2,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
2,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
2-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
4,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
4,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
4,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Acenaphthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Acenaphthylene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Acetochlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Alachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
alpha-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
alpha-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Anthracene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:41	1
Atrazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2

Lab Sample ID: 380-88933-3

Date Collected: 03/25/24 11:02

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Benz(a)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Benzo[a]pyrene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:41	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:41	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:41	1
beta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 15:41	1
Bromacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Butachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Butylbenzylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:41	1
Chlorobenzilate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Chloroneb	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Chlorpyrifos	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Chrysene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 15:41	1
delta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 15:41	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Dieldrin	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 15:41	1
Diethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:41	1
Dimethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:41	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		04/01/24 07:42	04/01/24 15:41	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Endosulfan sulfate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Endrin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Endrin aldehyde	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
EPTC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Fluoranthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Fluorene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
gamma-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Heptachlor	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:41	1
Heptachlor epoxide (isomer B)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Hexachlorobenzene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Isophorone	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 15:41	1
Lindane	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:41	1
Malathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Methoxychlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Metolachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Molinate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Naphthalene	<0.29		0.29	ug/L		04/01/24 07:42	04/01/24 15:41	1
Parathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Phenanthrene	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 15:41	1
Propachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2

Lab Sample ID: 380-88933-3

Date Collected: 03/25/24 11:02

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Simazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Terbacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Terbutylazine	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1
Thiobencarb	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 15:41	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 15:41	1
trans-Nonachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 15:41	1
Trifluralin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 15:41	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	04/01/24 07:42	04/01/24 15:41	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	105		70 - 130	04/01/24 07:42	04/01/24 15:41	1
Perylene-d12	91		70 - 130	04/01/24 07:42	04/01/24 15:41	1
Triphenylphosphate	96		70 - 130	04/01/24 07:42	04/01/24 15:41	1

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-88933-4

Date Collected: 03/25/24 10:07

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
2,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
2,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
2,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
2-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
4,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
4,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
4,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Acenaphthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Acenaphthylene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Acetochlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Alachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
alpha-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
alpha-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Anthracene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 16:01	1
Atrazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Benz(a)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Benzo[a]pyrene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 16:01	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 16:01	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 16:01	1
beta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 16:01	1
Bromacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Butachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-88933-4

Date Collected: 03/25/24 10:07

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Butylbenzylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 16:01	1
Chlorobenzilate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Chloroneb	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Chlorpyrifos	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Chrysene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 16:01	1
delta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 16:01	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Dieldrin	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 16:01	1
Diethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 16:01	1
Dimethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 16:01	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		04/01/24 07:42	04/01/24 16:01	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Endosulfan sulfate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Endrin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Endrin aldehyde	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
EPTC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Fluoranthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Fluorene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
gamma-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Heptachlor	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 16:01	1
Heptachlor epoxide (isomer B)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Hexachlorobenzene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Isophorone	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 16:01	1
Lindane	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 16:01	1
Malathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Methoxychlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Metolachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Molinate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Naphthalene	<0.29		0.29	ug/L		04/01/24 07:42	04/01/24 16:01	1
Parathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Phenanthrene	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 16:01	1
Propachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Simazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Terbacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Terbutylazine	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1
Thiobencarb	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 16:01	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 16:01	1
trans-Nonachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 16:01	1
Trifluralin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 16:01	1

Client Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-88933-1
 SDG: 525.2

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-88933-4

Date Collected: 03/25/24 10:07

Matrix: Drinking Water

Date Received: 03/28/24 10:44

PWSID Number: HI0000331

<u>Tentatively Identified Compound</u>	<u>Est. Result</u>	<u>Qualifier</u>	<u>Unit</u>	<u>D</u>	<u>RT</u>	<u>CAS No.</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
Tentatively Identified Compound	None		ug/L			N/A	04/01/24 07:42	04/01/24 16:01	1
<u>Surrogate</u>	<u>%Recovery</u>	<u>Qualifier</u>	<u>Limits</u>				<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
2-Nitro-m-xylene	104		70 - 130				04/01/24 07:42	04/01/24 16:01	1
Perylene-d12	95		70 - 130				04/01/24 07:42	04/01/24 16:01	1
Triphenylphosphate	117		70 - 130				04/01/24 07:42	04/01/24 16:01	1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: MOANALUA WELLS
PWSID Number: HI0000331

Lab Sample ID: 380-88933-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL		RL	Method	Prep Type
				Limit				
Alachlor	<0.048		ug/L	2		0.048	525.2	Total/NA
Atrazine	<0.048		ug/L	3		0.048	525.2	Total/NA
Benzo[a]pyrene	<0.019		ug/L	0.2		0.019	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.58		ug/L	6		0.58	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.58		ug/L	400		0.58	525.2	Total/NA
Endrin	<0.097		ug/L	2		0.097	525.2	Total/NA
Heptachlor	<0.039		ug/L	0.4		0.039	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.048		ug/L	0.2		0.048	525.2	Total/NA
Hexachlorobenzene	<0.048		ug/L	1		0.048	525.2	Total/NA
Hexachlorocyclopentadiene	<0.048		ug/L	50		0.048	525.2	Total/NA
Lindane	<0.039		ug/L	0.2		0.039	525.2	Total/NA
Methoxychlor	<0.097		ug/L	40		0.097	525.2	Total/NA
Simazine	<0.048		ug/L	4		0.048	525.2	Total/NA

Client Sample ID: AIEA GULCH WELLS PUMP 2
PWSID Number: HI0000331

Lab Sample ID: 380-88933-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL		RL	Method	Prep Type
				Limit				
Alachlor	<0.049		ug/L	2		0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3		0.049	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2		0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59		ug/L	6		0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400		0.59	525.2	Total/NA
Endrin	<0.098		ug/L	2		0.098	525.2	Total/NA
Heptachlor	<0.039		ug/L	0.4		0.039	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.049		ug/L	0.2		0.049	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1		0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50		0.049	525.2	Total/NA
Lindane	<0.039		ug/L	0.2		0.039	525.2	Total/NA
Methoxychlor	<0.098		ug/L	40		0.098	525.2	Total/NA
Simazine	<0.049		ug/L	4		0.049	525.2	Total/NA

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2
PWSID Number: HI0000331

Lab Sample ID: 380-88933-3

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL		RL	Method	Prep Type
				Limit				
Alachlor	<0.049		ug/L	2		0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3		0.049	525.2	Total/NA

Eurofins Eaton Analytical Pomona

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2
(Continued)

Lab Sample ID: 380-88933-3

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59		ug/L	6	0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400	0.59	525.2	Total/NA
Endrin	<0.098		ug/L	2	0.098	525.2	Total/NA
Heptachlor	<0.039		ug/L	0.4	0.039	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.049		ug/L	0.2	0.049	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.039		ug/L	0.2	0.039	525.2	Total/NA
Methoxychlor	<0.098		ug/L	40	0.098	525.2	Total/NA
Simazine	<0.049		ug/L	4	0.049	525.2	Total/NA

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-88933-4

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.049		ug/L	2	0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59		ug/L	6	0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400	0.59	525.2	Total/NA
Endrin	<0.098		ug/L	2	0.098	525.2	Total/NA
Heptachlor	<0.039		ug/L	0.4	0.039	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.049		ug/L	0.2	0.049	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.039		ug/L	0.2	0.039	525.2	Total/NA
Methoxychlor	<0.098		ug/L	40	0.098	525.2	Total/NA
Simazine	<0.049		ug/L	4	0.049	525.2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-88933-1
 SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-88933-1	MOANALUA WELLS	102	87	105
380-88933-1 MS	MOANALUA WELLS	103	92	102
380-88933-2	AIEA GULCH WELLS PUMP 2	103	92	98
380-88933-2 DU	AIEA GULCH WELLS PUMP 2	103	91	114
380-88933-3	AIEA WELLS PUMPS 1&2 (260) P2	105	91	96
380-88933-4	HALAWA WELLS UNITS 1 & 2 P1	104	95	117

Surrogate Legend

2NMX = 2-Nitro-m-xylene
 PRY = Perylene-d12
 TPP = Triphenylphosphate

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
LCS 380-83785/2-A	Lab Control Sample	102	95	104
MB 380-83785/1-A	Method Blank	103	94	111
MRL 380-83785/3-A	Lab Control Sample	98	93	94

Surrogate Legend

2NMX = 2-Nitro-m-xylene
 PRY = Perylene-d12
 TPP = Triphenylphosphate

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-88933-1
 SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 380-83785/1-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 83785

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
2,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
2,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
2,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
2-Methylnaphthalene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
4,4'-DDD	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
4,4'-DDE	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
4,4'-DDT	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Acenaphthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Acenaphthylene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Acetochlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Alachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
alpha-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
alpha-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Anthracene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 14:40	1
Atrazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Benz(a)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Benzo[a]pyrene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 14:40	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 14:40	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 14:40	1
beta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 14:40	1
Bromacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Butachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Butylbenzylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 14:40	1
Chlorobenzilate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Chloroneb	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Chlorpyrifos	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Chrysene	<0.020		0.020	ug/L		04/01/24 07:42	04/01/24 14:40	1
delta-BHC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		04/01/24 07:42	04/01/24 14:40	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Dieldrin	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 14:40	1
Diethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 14:40	1
Dimethylphthalate	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 14:40	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		04/01/24 07:42	04/01/24 14:40	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Endosulfan sulfate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Endrin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Endrin aldehyde	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
EPTC	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 380-83785/1-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 83785

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Fluorene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
gamma-Chlordane	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Heptachlor	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 14:40	1
Heptachlor epoxide (isomer B)	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Hexachlorobenzene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Isophorone	<0.49		0.49	ug/L		04/01/24 07:42	04/01/24 14:40	1
Lindane	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 14:40	1
Malathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Methoxychlor	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Metolachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Molinate	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Naphthalene	<0.29		0.29	ug/L		04/01/24 07:42	04/01/24 14:40	1
Parathion	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Phenanthrene	<0.039		0.039	ug/L		04/01/24 07:42	04/01/24 14:40	1
Propachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Pyrene	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Simazine	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Terbacil	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Terbutylazine	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1
Thiobencarb	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 14:40	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		04/01/24 07:42	04/01/24 14:40	1
trans-Nonachlor	<0.049		0.049	ug/L		04/01/24 07:42	04/01/24 14:40	1
Trifluralin	<0.098		0.098	ug/L		04/01/24 07:42	04/01/24 14:40	1

Tentatively Identified Compound	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Phenol, 2,4-bis(1,1-dimethylethyl)-	1.11	T J N	ug/L		4.29	96-76-4	04/01/24 07:42	04/01/24 14:40	1
Tetradecanamide	0.618	T J N	ug/L		6.79	638-58-4	04/01/24 07:42	04/01/24 14:40	1
9-Octadecenamide, (Z)-	11.9	T J N	ug/L		7.71	301-02-0	04/01/24 07:42	04/01/24 14:40	1
13-Docosenamide, (Z)-	2.05	T J N	ug/L		10.43	112-84-5	04/01/24 07:42	04/01/24 14:40	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	103		70 - 130	04/01/24 07:42	04/01/24 14:40	1
Perylene-d12	94		70 - 130	04/01/24 07:42	04/01/24 14:40	1
Triphenylphosphate	111		70 - 130	04/01/24 07:42	04/01/24 14:40	1

Lab Sample ID: LCS 380-83785/2-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	1.94	1.90		ug/L		98	70 - 130
2,4'-DDD	1.94	2.02		ug/L		104	70 - 130
2,4'-DDE	1.94	2.03		ug/L		105	70 - 130
2,4'-DDT	1.94	1.98		ug/L		102	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-83785/2-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2,4-Dinitrotoluene	1.94	1.86		ug/L		96	70 - 130
2,6-Dinitrotoluene	1.94	1.85		ug/L		95	70 - 130
2-Methylnaphthalene	1.94	1.94		ug/L		100	70 - 130
4,4'-DDD	1.94	1.82		ug/L		94	70 - 130
4,4'-DDE	1.94	1.89		ug/L		98	70 - 130
4,4'-DDT	1.94	2.04		ug/L		105	70 - 130
Acenaphthene	1.94	1.73		ug/L		89	70 - 130
Acenaphthylene	1.94	1.78		ug/L		92	70 - 130
Acetochlor	1.94	2.02		ug/L		104	70 - 130
Alachlor	1.94	1.97		ug/L		102	70 - 130
alpha-BHC	1.94	1.85		ug/L		95	70 - 130
alpha-Chlordane	1.94	1.79		ug/L		92	70 - 130
Anthracene	1.94	1.47		ug/L		76	70 - 130
Atrazine	1.94	1.97		ug/L		102	70 - 130
Benz(a)anthracene	1.94	1.80		ug/L		93	70 - 130
Benzo[a]pyrene	1.94	1.74		ug/L		90	70 - 130
Benzo[b]fluoranthene	1.94	2.00		ug/L		103	70 - 130
Benzo[g,h,i]perylene	1.94	2.24		ug/L		115	70 - 130
Benzo[k]fluoranthene	1.94	2.06		ug/L		106	70 - 130
beta-BHC	1.94	1.86		ug/L		96	70 - 130
Bis(2-ethylhexyl) phthalate	1.94	1.86		ug/L		96	70 - 130
Bromacil	1.94	2.01		ug/L		104	70 - 130
Butachlor	1.94	2.07		ug/L		107	70 - 130
Butylbenzylphthalate	1.94	2.18		ug/L		112	70 - 130
Chlorobenzilate	1.94	2.12		ug/L		110	70 - 130
Chloroneb	1.94	1.98		ug/L		102	70 - 130
Chlorothalonil (Draconil, Bravo)	1.94	1.91		ug/L		98	70 - 130
Chlorpyrifos	1.94	2.20		ug/L		113	70 - 130
Chrysene	1.94	1.91		ug/L		99	70 - 130
delta-BHC	1.94	1.78		ug/L		92	70 - 130
Di(2-ethylhexyl)adipate	1.94	2.14		ug/L		111	70 - 130
Dibenz(a,h)anthracene	1.94	2.16		ug/L		111	70 - 130
Diclorvos (DDVP)	1.94	1.94		ug/L		100	70 - 130
Dieldrin	1.94	1.91		ug/L		99	70 - 130
Diethylphthalate	1.94	1.94		ug/L		100	70 - 130
Dimethylphthalate	1.94	1.90		ug/L		98	70 - 130
Di-n-butyl phthalate	3.88	4.72		ug/L		122	70 - 130
Di-n-octyl phthalate	1.94	1.72		ug/L		89	70 - 130
Endosulfan I (Alpha)	1.94	1.96		ug/L		101	70 - 130
Endosulfan II (Beta)	1.94	2.35		ug/L		121	70 - 130
Endosulfan sulfate	1.94	1.99		ug/L		103	70 - 130
Endrin	1.94	1.89		ug/L		97	70 - 130
Endrin aldehyde	1.94	1.89		ug/L		98	60 - 130
EPTC	1.94	2.18		ug/L		112	70 - 130
Fluoranthene	1.94	1.93		ug/L		100	70 - 130
Fluorene	1.94	1.82		ug/L		94	70 - 130
gamma-Chlordane	1.94	1.85		ug/L		95	70 - 130
Heptachlor	1.94	2.03		ug/L		105	70 - 130
Heptachlor epoxide (isomer B)	1.94	1.95		ug/L		101	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-83785/2-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexachlorobenzene	1.94	1.90		ug/L		98	70 - 130
Hexachlorocyclopentadiene	1.94	1.73		ug/L		89	70 - 130
Indeno[1,2,3-cd]pyrene	1.94	2.18		ug/L		112	70 - 130
Isophorone	1.94	1.79		ug/L		92	70 - 130
Lindane	1.94	1.88		ug/L		97	70 - 130
Malathion	1.94	1.91		ug/L		98	70 - 130
Methoxychlor	1.94	2.01		ug/L		103	70 - 130
Metolachlor	1.94	2.06		ug/L		106	70 - 130
Molinate	1.94	2.12		ug/L		109	70 - 130
Naphthalene	1.94	1.91		ug/L		98	70 - 130
Parathion	1.94	2.14		ug/L		110	70 - 130
Pendimethalin (Penoxaline)	1.94	1.97		ug/L		101	70 - 130
Phenanthrene	1.94	1.77		ug/L		91	70 - 130
Propachlor	1.94	2.02		ug/L		104	70 - 130
Pyrene	1.94	2.09		ug/L		108	70 - 130
Simazine	1.94	1.99		ug/L		103	70 - 130
Terbacil	1.94	1.98		ug/L		102	70 - 130
Terbutylazine	1.94	1.88		ug/L		97	70 - 130
Thiobencarb	1.94	1.94		ug/L		100	70 - 130
trans-Nonachlor	1.94	1.63		ug/L		84	70 - 130
Trifluralin	1.94	1.85		ug/L		95	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2-Nitro-m-xylene	102		70 - 130
Perylene-d12	95		70 - 130
Triphenylphosphate	104		70 - 130

Lab Sample ID: MRL 380-83785/3-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.0975	0.102		ug/L		105	50 - 150
2,4'-DDD	0.0975	0.101		ug/L		103	50 - 150
2,4'-DDE	0.0975	0.0684	J	ug/L		70	50 - 150
2,4'-DDT	0.0975	0.0826	J	ug/L		85	50 - 150
2,4-Dinitrotoluene	0.0975	0.0827	J	ug/L		85	50 - 150
2,6-Dinitrotoluene	0.0975	0.0974	J	ug/L		100	50 - 150
2-Methylnaphthalene	0.0975	0.0951	J	ug/L		98	50 - 150
4,4'-DDD	0.0975	0.0809	J	ug/L		83	50 - 150
4,4'-DDE	0.0975	0.0806	J	ug/L		83	50 - 150
4,4'-DDT	0.0975	0.104		ug/L		106	50 - 150
Acenaphthene	0.0975	0.0891	J	ug/L		91	50 - 150
Acenaphthylene	0.0975	0.0943	J	ug/L		97	50 - 150
Acetochlor	0.0488	0.0454	J	ug/L		93	50 - 150
Alachlor	0.0488	0.0502		ug/L		103	50 - 150
alpha-BHC	0.0975	0.110		ug/L		113	50 - 150
alpha-Chlordane	0.0244	<0.028		ug/L		71	50 - 150

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-83785/3-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Anthracene	0.0195	<0.019		ug/L		90	50 - 150
Atrazine	0.0488	0.0488	J	ug/L		100	50 - 150
Benz(a)anthracene	0.0488	0.0546		ug/L		112	50 - 150
Benzo[a]pyrene	0.0195	0.0112	J	ug/L		58	50 - 150
Benzo[b]fluoranthene	0.0195	0.0132	J	ug/L		68	50 - 150
Benzo[g,h,i]perylene	0.0488	0.0448	J	ug/L		92	50 - 150
Benzo[k]fluoranthene	0.0195	<0.017		ug/L		62	50 - 150
beta-BHC	0.0975	0.101		ug/L		103	50 - 150
Bis(2-ethylhexyl) phthalate	0.585	0.623		ug/L		107	50 - 150
Bromacil	0.0975	0.107		ug/L		110	50 - 150
Butachlor	0.0488	0.0423	J	ug/L		87	50 - 150
Butylbenzylphthalate	0.146	0.165	J	ug/L		113	50 - 150
Chlorobenzilate	0.0975	0.0977	J	ug/L		100	50 - 150
Chloroneb	0.0975	0.111		ug/L		114	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0975	0.0943	J	ug/L		97	50 - 150
Chlorpyrifos	0.0488	0.0460	J	ug/L		94	50 - 150
Chrysene	0.0195	0.0286		ug/L		147	50 - 150
delta-BHC	0.0975	0.112		ug/L		115	50 - 150
Di(2-ethylhexyl)adipate	0.293	0.375	J	ug/L		128	50 - 150
Dibenz(a,h)anthracene	0.0488	0.0439	J	ug/L		90	50 - 150
Diclorvos (DDVP)	0.0488	0.0641		ug/L		132	50 - 150
Dieldrin	0.0975	0.0949	J	ug/L		97	50 - 150
Diethylphthalate	0.146	0.156	J	ug/L		107	50 - 150
Dimethylphthalate	0.293	0.296	J	ug/L		101	50 - 150
Di-n-butyl phthalate	0.293	0.332	J	ug/L		114	49 - 243
Di-n-octyl phthalate	0.0975	0.0652	J	ug/L		67	50 - 150
Endosulfan I (Alpha)	0.0975	0.0938	J	ug/L		96	50 - 150
Endosulfan II (Beta)	0.0975	0.107		ug/L		110	50 - 150
Endosulfan sulfate	0.0975	0.0798	J	ug/L		82	50 - 150
Endrin	0.0975	0.0693	J	ug/L		71	50 - 150
Endrin aldehyde	0.0975	<0.082		ug/L		82	50 - 150
EPTC	0.0975	0.0997		ug/L		102	50 - 150
Fluoranthene	0.0488	0.0483	J	ug/L		99	50 - 150
Fluorene	0.0488	<0.049		ug/L		97	50 - 150
gamma-Chlordane	0.0244	<0.020		ug/L		72	50 - 150
Heptachlor	0.0390	0.0423		ug/L		108	50 - 150
Heptachlor epoxide (isomer B)	0.0488	0.0485	J	ug/L		99	50 - 150
Hexachlorobenzene	0.0488	0.0439	J	ug/L		90	50 - 150
Hexachlorocyclopentadiene	0.0488	0.0459	J	ug/L		94	50 - 150
Indeno[1,2,3-cd]pyrene	0.0488	0.0473	J	ug/L		97	50 - 150
Isophorone	0.0975	0.110	J	ug/L		112	50 - 150
Lindane	0.0390	0.0465		ug/L		119	50 - 150
Malathion	0.0975	0.0947	J	ug/L		97	50 - 150
Methoxychlor	0.0975	0.0975	J	ug/L		100	50 - 150
Metolachlor	0.0488	0.0533		ug/L		109	50 - 150
Molinate	0.0975	0.0980		ug/L		100	50 - 150
Naphthalene	0.0975	0.102	J	ug/L		104	50 - 150
Parathion	0.0975	0.0754	J	ug/L		77	50 - 150
Pendimethalin (Penoxaline)	0.0975	0.0788	J	ug/L		81	50 - 150

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-83785/3-A
Matrix: Water
Analysis Batch: 83952

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Phenanthrene	0.0195	0.0202	J	ug/L		103	50 - 150
Propachlor	0.0488	0.0480	J	ug/L		99	50 - 150
Pyrene	0.0488	0.0466	J	ug/L		95	50 - 150
Simazine	0.0488	0.0492		ug/L		101	50 - 150
Terbacil	0.0975	0.0955	J	ug/L		98	50 - 150
Terbutylazine	0.0975	0.0983		ug/L		101	50 - 150
Thiobencarb	0.0975	0.105	J	ug/L		108	50 - 150
trans-Nonachlor	0.0244	<0.025		ug/L		99	50 - 150
Trifluralin	0.0975	0.0926	J	ug/L		95	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	93		70 - 130
Triphenylphosphate	94		70 - 130

Lab Sample ID: 380-88933-1 MS
Matrix: Drinking Water
Analysis Batch: 83952

Client Sample ID: MOANALUA WELLS
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.097		1.97	2.07		ug/L		105	70 - 130
2,4'-DDD	<0.097		1.97	2.01		ug/L		102	70 - 130
2,4'-DDE	<0.097		1.97	2.01		ug/L		102	70 - 130
2,4'-DDT	<0.097		1.97	1.98		ug/L		101	70 - 130
2,4-Dinitrotoluene	<0.097		1.97	2.06		ug/L		105	70 - 130
2,6-Dinitrotoluene	<0.097		1.97	1.95		ug/L		99	70 - 130
2-Methylnaphthalene	<0.097		1.97	2.11		ug/L		107	70 - 130
4,4'-DDD	<0.097		1.97	1.82		ug/L		92	70 - 130
4,4'-DDE	<0.097		1.97	1.86		ug/L		94	70 - 130
4,4'-DDT	<0.097		1.97	2.00		ug/L		102	70 - 130
Acenaphthene	<0.097		1.97	1.84		ug/L		94	70 - 130
Acenaphthylene	<0.097		1.97	1.89		ug/L		96	70 - 130
Acetochlor	<0.097		1.97	1.97		ug/L		100	70 - 130
Alachlor	<0.048		1.97	2.00		ug/L		101	70 - 130
alpha-BHC	<0.097		1.97	1.94		ug/L		99	70 - 130
alpha-Chlordane	<0.048		1.97	1.76		ug/L		89	70 - 130
Anthracene	<0.019	F1	1.97	1.02	F1	ug/L		52	70 - 130
Atrazine	<0.048		1.97	2.09		ug/L		106	70 - 130
Benz(a)anthracene	<0.048		1.97	1.68		ug/L		85	70 - 130
Benzo[a]pyrene	<0.019		1.97	1.47		ug/L		75	70 - 130
Benzo[b]fluoranthene	<0.019		1.97	2.04		ug/L		104	70 - 130
Benzo[g,h,i]perylene	<0.048		1.97	2.20		ug/L		112	70 - 130
Benzo[k]fluoranthene	<0.019		1.97	2.06		ug/L		105	70 - 130
beta-BHC	<0.097		1.97	1.95		ug/L		99	70 - 130
Bis(2-ethylhexyl) phthalate	<0.58		1.97	1.81		ug/L		92	70 - 130
Bromacil	<0.097		1.97	2.08		ug/L		106	70 - 130
Butachlor	<0.048		1.97	2.07		ug/L		105	70 - 130
Butylbenzylphthalate	<0.48		1.97	2.18		ug/L		111	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-88933-1 MS
Matrix: Drinking Water
Analysis Batch: 83952

Client Sample ID: MOANALUA WELLS
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Chlorobenzilate	<0.097		1.97	2.18		ug/L		111	70 - 130
Chloroneb	<0.097		1.97	2.06		ug/L		105	70 - 130
Chlorothalonil (Draconil, Bravo)	<0.097		1.97	2.00		ug/L		102	70 - 130
Chlorpyrifos	<0.048		1.97	2.21		ug/L		112	70 - 130
Chrysene	<0.019		1.97	2.03		ug/L		103	70 - 130
delta-BHC	<0.097		1.97	1.82		ug/L		93	70 - 130
Di(2-ethylhexyl)adipate	<0.58		1.97	2.01		ug/L		102	70 - 130
Dibenz(a,h)anthracene	<0.048		1.97	2.08		ug/L		106	70 - 130
Diclorvos (DDVP)	<0.048		1.97	2.04		ug/L		104	70 - 130
Dieldrin	<0.19		1.97	1.96		ug/L		100	70 - 130
Diethylphthalate	<0.48		1.97	2.03		ug/L		103	70 - 130
Dimethylphthalate	<0.48		1.97	2.00		ug/L		102	70 - 130
Di-n-butyl phthalate	<0.97		3.94	4.79		ug/L		119	70 - 130
Di-n-octyl phthalate	<0.097		1.97	1.65		ug/L		84	70 - 130
Endosulfan I (Alpha)	<0.097		1.97	1.96		ug/L		99	70 - 130
Endosulfan II (Beta)	<0.097		1.97	2.27		ug/L		115	70 - 130
Endosulfan sulfate	<0.097		1.97	2.01		ug/L		102	70 - 130
Endrin	<0.097		1.97	1.92		ug/L		98	70 - 130
Endrin aldehyde	<0.097		1.97	2.01		ug/L		102	60 - 130
EPTC	<0.097		1.97	2.30		ug/L		117	70 - 130
Fluoranthene	<0.097		1.97	2.00		ug/L		102	70 - 130
Fluorene	<0.048		1.97	1.94		ug/L		99	70 - 130
gamma-Chlordane	<0.048		1.97	1.80		ug/L		91	70 - 130
Heptachlor	<0.039		1.97	2.00		ug/L		102	70 - 130
Heptachlor epoxide (isomer B)	<0.048		1.97	1.93		ug/L		98	70 - 130
Hexachlorobenzene	<0.048		1.97	2.05		ug/L		104	70 - 130
Hexachlorocyclopentadiene	<0.048		1.97	1.93		ug/L		98	70 - 130
Indeno[1,2,3-cd]pyrene	<0.048		1.97	2.13		ug/L		108	70 - 130
Isophorone	<0.48		1.97	1.89		ug/L		96	70 - 130
Lindane	<0.039		1.97	1.98		ug/L		101	70 - 130
Malathion	<0.097		1.97	1.93		ug/L		98	70 - 130
Methoxychlor	<0.097		1.97	2.21		ug/L		112	70 - 130
Metolachlor	<0.048		1.97	2.06		ug/L		105	70 - 130
Molinate	<0.097		1.97	2.22		ug/L		113	70 - 130
Naphthalene	<0.29		1.97	2.03		ug/L		103	70 - 130
Parathion	<0.097		1.97	2.25		ug/L		114	70 - 130
Pendimethalin (Penoxaline)	<0.097		1.97	2.06		ug/L		105	70 - 130
Phenanthrene	<0.039		1.97	1.87		ug/L		95	70 - 130
Propachlor	<0.048		1.97	2.14		ug/L		109	70 - 130
Pyrene	<0.048		1.97	2.11		ug/L		107	70 - 130
Simazine	<0.048		1.97	2.14		ug/L		109	70 - 130
Terbacil	<0.097		1.97	2.16		ug/L		110	70 - 130
Terbutylazine	<0.097		1.97	1.98		ug/L		101	70 - 130
Thiobencarb	<0.19		1.97	1.94		ug/L		99	70 - 130
trans-Nonachlor	<0.048		1.97	1.72		ug/L		87	70 - 130
Trifluralin	<0.097		1.97	2.05		ug/L		104	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-88933-1 MS
Matrix: Drinking Water
Analysis Batch: 83952

Client Sample ID: MOANALUA WELLS
Prep Type: Total/NA
Prep Batch: 83785

<i>Surrogate</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
2-Nitro-m-xylene	103		70 - 130
Perylene-d12	92		70 - 130
Triphenylphosphate	102		70 - 130

Lab Sample ID: 380-88933-2 DU
Matrix: Drinking Water
Analysis Batch: 83952

Client Sample ID: AIEA GULCH WELLS PUMP 2
Prep Type: Total/NA
Prep Batch: 83785

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
1-Methylnaphthalene	<0.098		<0.099		ug/L		NC	20
2,4'-DDD	<0.098		<0.099		ug/L		NC	20
2,4'-DDE	<0.098		<0.099		ug/L		NC	20
2,4'-DDT	<0.098		<0.099		ug/L		NC	20
2,4-Dinitrotoluene	<0.098		<0.099		ug/L		NC	20
2,6-Dinitrotoluene	<0.098		<0.099		ug/L		NC	20
2-Methylnaphthalene	<0.098		<0.099		ug/L		NC	20
4,4'-DDD	<0.098		<0.099		ug/L		NC	20
4,4'-DDE	<0.098		<0.099		ug/L		NC	20
4,4'-DDT	<0.098		<0.099		ug/L		NC	20
Acenaphthene	<0.098		<0.099		ug/L		NC	20
Acenaphthylene	<0.098		<0.099		ug/L		NC	20
Acetochlor	<0.098		<0.099		ug/L		NC	20
Alachlor	<0.049		<0.049		ug/L		NC	20
alpha-BHC	<0.098		<0.099		ug/L		NC	20
alpha-Chlordane	<0.049		<0.049		ug/L		NC	20
Anthracene	<0.020		<0.020		ug/L		NC	20
Atrazine	<0.049		<0.049		ug/L		NC	20
Benz(a)anthracene	<0.049		<0.049		ug/L		NC	20
Benzo[a]pyrene	<0.020		<0.020		ug/L		NC	20
Benzo[b]fluoranthene	<0.020		<0.020		ug/L		NC	20
Benzo[g,h,i]perylene	<0.049		<0.049		ug/L		NC	20
Benzo[k]fluoranthene	<0.020		<0.020		ug/L		NC	20
beta-BHC	<0.098		<0.099		ug/L		NC	20
Bis(2-ethylhexyl) phthalate	<0.59		<0.59		ug/L		NC	20
Bromacil	<0.098		<0.099		ug/L		NC	20
Butachlor	<0.049		<0.049		ug/L		NC	20
Butylbenzylphthalate	<0.49		<0.49		ug/L		NC	20
Chlorobenzilate	<0.098		<0.099		ug/L		NC	20
Chloroneb	<0.098		<0.099		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	<0.098		<0.099		ug/L		NC	20
Chlorpyrifos	<0.049		<0.049		ug/L		NC	20
Chrysene	<0.020		<0.020		ug/L		NC	20
delta-BHC	<0.098		<0.099		ug/L		NC	20
Di(2-ethylhexyl)adipate	<0.59		<0.59		ug/L		NC	20
Dibenz(a,h)anthracene	<0.049		<0.049		ug/L		NC	20
Diclorvos (DDVP)	<0.049		<0.049		ug/L		NC	20
Dieldrin	<0.20		<0.20		ug/L		NC	20
Diethylphthalate	<0.49		<0.49		ug/L		NC	20

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-88933-2 DU

Matrix: Drinking Water

Analysis Batch: 83952

Client Sample ID: AIEA GULCH WELLS PUMP 2

Prep Type: Total/NA

Prep Batch: 83785

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
Dimethylphthalate	<0.49		<0.49		ug/L		NC	20
Di-n-butyl phthalate	<0.98		<0.99		ug/L		NC	20
Di-n-octyl phthalate	<0.098		<0.099		ug/L		NC	20
Endosulfan I (Alpha)	<0.098		<0.099		ug/L		NC	20
Endosulfan II (Beta)	<0.098		<0.099		ug/L		NC	20
Endosulfan sulfate	<0.098		<0.099		ug/L		NC	20
Endrin	<0.098		<0.099		ug/L		NC	20
Endrin aldehyde	<0.098		<0.099		ug/L		NC	20
EPTC	<0.098		<0.099		ug/L		NC	20
Fluoranthene	<0.098		<0.099		ug/L		NC	20
Fluorene	<0.049		<0.049		ug/L		NC	20
gamma-Chlordane	<0.049		<0.049		ug/L		NC	20
Heptachlor	<0.039		<0.039		ug/L		NC	20
Heptachlor epoxide (isomer B)	<0.049		<0.049		ug/L		NC	20
Hexachlorobenzene	<0.049		<0.049		ug/L		NC	20
Hexachlorocyclopentadiene	<0.049		<0.049		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	<0.049		<0.049		ug/L		NC	20
Isophorone	<0.49		<0.49		ug/L		NC	20
Lindane	<0.039		<0.039		ug/L		NC	20
Malathion	<0.098		<0.099		ug/L		NC	20
Methoxychlor	<0.098		<0.099		ug/L		NC	20
Metolachlor	<0.049		<0.049		ug/L		NC	20
Molinate	<0.098		<0.099		ug/L		NC	20
Naphthalene	<0.29		<0.30		ug/L		NC	20
Parathion	<0.098		<0.099		ug/L		NC	20
Pendimethalin (Penoxaline)	<0.098		<0.099		ug/L		NC	20
Phenanthrene	<0.039		<0.039		ug/L		NC	20
Propachlor	<0.049		<0.049		ug/L		NC	20
Pyrene	<0.049		<0.049		ug/L		NC	20
Simazine	<0.049		<0.049		ug/L		NC	20
Terbacil	<0.098		<0.099		ug/L		NC	20
Terbutylazine	<0.098		<0.099		ug/L		NC	20
Thiobencarb	<0.20		<0.20		ug/L		NC	20
Total Permethrin (mixed isomers)	<0.20		<0.20		ug/L		NC	20
trans-Nonachlor	<0.049		<0.049		ug/L		NC	20
Trifluralin	<0.098		<0.099		ug/L		NC	20

Surrogate	DU %Recovery	DU Qualifier	Limits
2-Nitro-m-xylene	103		70 - 130
Perylene-d12	91		70 - 130
Triphenylphosphate	114		70 - 130

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-88933-1
 SDG: 525.2

GC/MS Semi VOA

Prep Batch: 83785

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-88933-1	MOANALUA WELLS	Total/NA	Drinking Water	525.2	
380-88933-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	525.2	
380-88933-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	525.2	
380-88933-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	525.2	
MB 380-83785/1-A	Method Blank	Total/NA	Water	525.2	
LCS 380-83785/2-A	Lab Control Sample	Total/NA	Water	525.2	
MRL 380-83785/3-A	Lab Control Sample	Total/NA	Water	525.2	
380-88933-1 MS	MOANALUA WELLS	Total/NA	Drinking Water	525.2	
380-88933-2 DU	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	525.2	

Analysis Batch: 83952

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-88933-1	MOANALUA WELLS	Total/NA	Drinking Water	525.2	83785
380-88933-2	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	525.2	83785
380-88933-3	AIEA WELLS PUMPS 1&2 (260) P2	Total/NA	Drinking Water	525.2	83785
380-88933-4	HALAWA WELLS UNITS 1 & 2 P1	Total/NA	Drinking Water	525.2	83785
MB 380-83785/1-A	Method Blank	Total/NA	Water	525.2	83785
LCS 380-83785/2-A	Lab Control Sample	Total/NA	Water	525.2	83785
MRL 380-83785/3-A	Lab Control Sample	Total/NA	Water	525.2	83785
380-88933-1 MS	MOANALUA WELLS	Total/NA	Drinking Water	525.2	83785
380-88933-2 DU	AIEA GULCH WELLS PUMP 2	Total/NA	Drinking Water	525.2	83785

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-88933-1
 SDG: 525.2

Client Sample ID: MOANALUA WELLS

Lab Sample ID: 380-88933-1

Date Collected: 03/25/24 09:38

Matrix: Drinking Water

Date Received: 03/28/24 10:44

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			83785	OTM3	EA POM	04/01/24 07:42
Total/NA	Analysis	525.2		1	83952	Q8LA	EA POM	04/01/24 15:01

Client Sample ID: AIEA GULCH WELLS PUMP 2

Lab Sample ID: 380-88933-2

Date Collected: 03/25/24 10:35

Matrix: Drinking Water

Date Received: 03/28/24 10:44

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			83785	OTM3	EA POM	04/01/24 07:42
Total/NA	Analysis	525.2		1	83952	Q8LA	EA POM	04/01/24 15:21

Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2

Lab Sample ID: 380-88933-3

Date Collected: 03/25/24 11:02

Matrix: Drinking Water

Date Received: 03/28/24 10:44

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			83785	OTM3	EA POM	04/01/24 07:42
Total/NA	Analysis	525.2		1	83952	Q8LA	EA POM	04/01/24 15:41

Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1

Lab Sample ID: 380-88933-4

Date Collected: 03/25/24 10:07

Matrix: Drinking Water

Date Received: 03/28/24 10:44

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			83785	OTM3	EA POM	04/01/24 07:42
Total/NA	Analysis	525.2		1	83952	Q8LA	EA POM	04/01/24 16:01

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Laboratory: Eurofins Eaton Analytical Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	02-12-24 *

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	1-Methylnaphthalene
525.2	525.2	Drinking Water	2,4'-DDD
525.2	525.2	Drinking Water	2,4'-DDE
525.2	525.2	Drinking Water	2,4'-DDT
525.2	525.2	Drinking Water	2,4-Dinitrotoluene
525.2	525.2	Drinking Water	2,6-Dinitrotoluene
525.2	525.2	Drinking Water	2-Methylnaphthalene
525.2	525.2	Drinking Water	4,4'-DDD
525.2	525.2	Drinking Water	4,4'-DDE
525.2	525.2	Drinking Water	4,4' DDT
525.2	525.2	Drinking Water	Acenaphthene
525.2	525.2	Drinking Water	Acenaphthylene
525.2	525.2	Drinking Water	Acetochlor
525.2	525.2	Drinking Water	alpha-BHC
525.2	525.2	Drinking Water	alpha-Chlordane
525.2	525.2	Drinking Water	Anthracene
525.2	525.2	Drinking Water	Benz(a)anthracene
525.2	525.2	Drinking Water	Benzo[b]fluoranthene
525.2	525.2	Drinking Water	Benzo[g,h,i]perylene
525.2	525.2	Drinking Water	Benzo[k]fluoranthene
525.2	525.2	Drinking Water	beta-BHC
525.2	525.2	Drinking Water	Bromacil
525.2	525.2	Drinking Water	Butylbenzylphthalate
525.2	525.2	Drinking Water	Chlorobenzilate
525.2	525.2	Drinking Water	Chloroneb
525.2	525.2	Drinking Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Drinking Water	Chlorpyrifos
525.2	525.2	Drinking Water	Chrysene
525.2	525.2	Drinking Water	delta-BHC
525.2	525.2	Drinking Water	Dibenz(a,h)anthracene
525.2	525.2	Drinking Water	Diclorvos (DDVP)
525.2	525.2	Drinking Water	Diethylphthalate
525.2	525.2	Drinking Water	Dimethylphthalate
525.2	525.2	Drinking Water	Di-n-butyl phthalate
525.2	525.2	Drinking Water	Di n octyl phthalate
525.2	525.2	Drinking Water	Endosulfan I (Alpha)
525.2	525.2	Drinking Water	Endosulfan II (Beta)
525.2	525.2	Drinking Water	Endosulfan sulfate
525.2	525.2	Drinking Water	Endrin aldehyde
525.2	525.2	Drinking Water	EPTC
525.2	525.2	Drinking Water	Fluoranthene
525.2	525.2	Drinking Water	Fluorene
525.2	525.2	Drinking Water	gamma-Chlordane
525.2	525.2	Drinking Water	Indeno[1,2,3-cd]pyrene
525.2	525.2	Drinking Water	Isophorone

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Laboratory: Eurofins Eaton Analytical Pomona (Continued)

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
-----------	---------	-----------------------	-----------------

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	Malathion
525.2	525.2	Drinking Water	Molinate
525.2	525.2	Drinking Water	Naphthalene
525.2	525.2	Drinking Water	Parathion
525.2	525.2	Drinking Water	Pendimethalin (Penoxaline)
525.2	525.2	Drinking Water	Phenanthrene
525.2	525.2	Drinking Water	Pyrene
525.2	525.2	Drinking Water	Terbacil
525.2	525.2	Drinking Water	Terbutylazine
525.2	525.2	Drinking Water	Thiobencarb
525.2	525.2	Drinking Water	Total Permethrin (mixed isomers)
525.2	525.2	Drinking Water	trans-Nonachlor
525.2	525.2	Drinking Water	Trifluralin

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-88933-1
SDG: 525.2

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
525.2	Extraction of Semivolatile Compounds	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL


Job ID: 380-88933-1
SDG: 525.2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-88933-1	MOANALUA WELLS	Drinking Water	03/25/24 09:38	03/28/24 10:44	HI0000331
380-88933-2	AIEA GULCH WELLS PUMP 2	Drinking Water	03/25/24 10:35	03/28/24 10:44	HI0000331
380-88933-3	AIEA WELLS PUMPS 1&2 (260) P2	Drinking Water	03/25/24 11:02	03/28/24 10:44	HI0000331
380-88933-4	HALAWA WELLS UNITS 1 & 2 P1	Drinking Water	03/25/24 10:07	03/28/24 10:44	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Monrovia, CA (Suite 100)
 750 Royal Oaks Drive Suite 100
 Monrovia CA 91016
 Phone (626) 386-1100

Chain of Custody Record



Client Information		Lab PM Arada, Rachelle		Carrier Tracking Note(s)		COC No 380-27984-2757.2	
Client Contact Dr Ron Fenstermacher		E-Mail Rachelle.Arada@et.euronisus.com		State of Origin		Page Page 1 of 1	
Company City & County of Honolulu		PWSID		Analysis Requested		Job #	
Address 630 South Beretania Street; Chemistry Lab		Due Date Requested		533 - All Analytes		Preservation Codes	
City Honolulu		TAT Requested (days)		537 1_DW_PREC - 537 1 Full List		M - Hexane N - None O - AsNaO2 P - Na2O4S Q - Na2SO3 R - NaHSO4 S - H2SO4 T - TSP Dodecahydrate	
State Zip HI, 96843		Compliance Project Δ No		525 2_PREC - (MOD) 525plus PLUS TICs		U - Acetone V - MCAA W - pH 4-5 Y - Trizma Z - other (specify)	
Phone 808-748-5091 (tel)		C20525101 exp 05312023		8015B_DRO_LL_CS HNL Ranges C10 C24/C24-C36/C8		Other	
Email rfenstermacher@hbws.org		WO #		8015B_GRO_LL - (MOD) GRO			
Project Name RED-HILL/HBWS sites Event Desc. RUSH Weekly Red Hill		Project # 38001111		SUBCONTRACT 825 PAH Physis LL (EAL) + TICs			
Site		SSOW#		Perform MS/MSD (Yes or No)			
				Field Filtered Sample (Yes or No)			
Sample Identification		Sample Date		Sample Time		Sample Type (C=Comp, G=grab)	
MOANALUA WELLS		25-Mar-2024		0938		G	
AIEA GULCH WELLS PUMP2		25-Mar-2024		1088		G	
AIEA WELLS PUMPS 1&2 (260)P2		25-Mar-2024		1102		G	
HALAWA WELLS UNITS 1&2 P1		25-Mar-2024		1007		G	
TB MOANALUA WELLS		25-Mar-2024		0938			
TB AIEA GULCH WELLS PUMP2		25-Mar-2024		1088			
TB AIEA WELLS PUMPS 1&2 (260)		25-Mar-2024		1102			
TB HALAWA WELLS UNITS 1&2		25-Mar-2024		1007			
Possible Hazard Identification		<input type="checkbox"/> Non-Hazard		<input type="checkbox"/> Flammable		<input type="checkbox"/> Skin Irritant	
Deliverable Requested		<input type="checkbox"/> I, II, III		<input type="checkbox"/> IV		Other (specify)	
Empty Kit Relinquished by		Date		Time		Company	
Relinquished by		21 March 2024		1400		HBWS	
Relinquished by		Date/Time		Date/Time		Company	
Relinquished by		Date/Time		Date/Time		Company	
Custody Seals Intact		Custody Seal No		Color Temperature(s) °C and Other Remarks			
Δ Yes Δ No				75.1A 0.04 - 0.1 °C / 0.1 °C		① 1.4 °C - 0.1 °C - 1.3 °C GEL-FROZEN	



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-88933-1

SDG Number: 525.2

Login Number: 88933

List Number: 1

Creator: Elyas, Matthew

List Source: Eurofins Eaton Analytical Pomona

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	False	Refer to NCM for affected items.
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

