

## ANALYTICAL REPORT

Eurofins Eaton Monrovia  
750 Royal Oaks Drive  
Suite 100  
Monrovia, CA 91016  
Tel: (626)386-1100

Laboratory Job ID: 380-13074-1  
Client Project/Site: RED-HILL  
Sampling Event: RUSH Weekly Red Hill

For:  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 308  
Honolulu, Hawaii 96843

Attn: Mr. Erwin Kawata



Authorized for release by:  
10/17/2022 4:30:28 PM  
Kathleen Robb, Client Program Manager  
(949)261-1022  
[Kathleen.Robb@et.eurofinsus.com](mailto:Kathleen.Robb@et.eurofinsus.com)

Designee for  
Rachelle Arada, Manager of Project Management  
(626)386-1106  
[Rachelle.Arada@et.eurofinsus.com](mailto:Rachelle.Arada@et.eurofinsus.com)

### LINKS

Review your project  
results through



Have a Question?



Visit us at:

[www.eurofinsus.com/Env](http://www.eurofinsus.com/Env)

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW,Water matrices)



---

Kathleen Robb  
Client Program Manager  
10/17/2022 4:30:29 PM





# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	10
Surrogate Summary . . . . .	11
QC Sample Results . . . . .	14
QC Association Summary . . . . .	29
Lab Chronicle . . . . .	31
Certification Summary . . . . .	32
Method Summary . . . . .	34
Sample Summary . . . . .	35
Subcontract Data . . . . .	36
Chain of Custody . . . . .	83
Receipt Checklists . . . . .	87

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Qualifiers

### GC/MS Semi VOA

Qualifier	Qualifier Description
*+	LCS and/or LCSD is outside acceptance limits, high biased.
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

### GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

### Subcontract

Qualifier	Qualifier Description
U	This analyte was not detected.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
⌘	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count



# Case Narrative

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

---

## Job ID: 380-13074-1

---

### Laboratory: Eurofins Eaton Monrovia

#### Narrative

---

#### Job Narrative 380-13074-1

#### Comments

No additional comments.

#### Receipt

The sample was received on 7/29/2022 10:00 AM. Unless otherwise noted below, the sample arrived in good condition, and where required, properly preserved and on ice. The temperature of the cooler at receipt was 4.9° C.

#### GC/MS Semi VOA

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

#### Subcontract non-Sister

See attached subcontract report.

#### Organic Prep

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

#### Subcontract Work

Methods 8015 Diesel LL (EAL) and Motor Oil, 8015 Gas (Purgeable) LL (EAL): These methods were subcontracted to EMAX Laboratories Inc. The subcontract laboratory certifications are different from that of the facility issuing the final report.

Method 625 PAH Physis LL (EAL) + TICs: This method was subcontracted to Physis Environmental Laboratories. The subcontract laboratory certification is different from that of the facility issuing the final report.



# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

**Client Sample ID: AIEA GULCH WELLS PUMP 2  
(331-202-TP072)**  
**PWSID Number: HI0000331**

**Lab Sample ID: 380-13074-1**

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Monrovia

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

**Client Sample ID: AIEA GULCH WELLS PUMP 2  
(331-202-TP072)**

**Lab Sample ID: 380-13074-1**

**Date Collected: 07/27/22 11:36**

**Matrix: Drinking Water**

**Date Received: 07/29/22 10:00**

**PWSID Number: HI0000331**

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
2,4'-DDD	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
2,4'-DDE	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
2,4'-DDT	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
2,4-Dinitrotoluene	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
2,6-Dinitrotoluene	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
4,4'-DDD	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
4,4'-DDE	ND	^3+	0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
4,4'-DDT	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Acenaphthene	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Acenaphthylene	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Acetochlor	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Alachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
alpha-BHC	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
alpha-Chlordane	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Anthracene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 11:00	1
Atrazine	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Benz(a)anthracene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Benzo[a]pyrene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 11:00	1
Benzo[b]fluoranthene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 11:00	1
Benzo[g,h,i]perylene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Benzo[k]fluoranthene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 11:00	1
beta-BHC	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Bromacil	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Butachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Butylbenzylphthalate	ND	^3+	0.49	ug/L		08/01/22 08:52	08/19/22 11:00	1
Caffeine	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Chlorobenzilate	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Chloroneb	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Chlorothalonil (Draconil, Bravo)	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Chlorpyrifos	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Chrysene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 11:00	1
delta-BHC	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Di(2-ethylhexyl)adipate	ND	^3+ **	0.59	ug/L		08/01/22 08:52	08/19/22 11:00	1
Bis(2-ethylhexyl) phthalate	ND		0.59	ug/L		08/01/22 08:52	08/19/22 11:00	1
Diazinon (Qualitative)	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Dibenz(a,h)anthracene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Diclorvos (DDVP)	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Dieldrin	ND		0.20	ug/L		08/01/22 08:52	08/19/22 11:00	1
Diethylphthalate	ND		0.49	ug/L		08/01/22 08:52	08/19/22 11:00	1
Dimethoate	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Dimethylphthalate	ND		0.49	ug/L		08/01/22 08:52	08/19/22 11:00	1
Di-n-butyl phthalate	ND		0.98	ug/L		08/01/22 08:52	08/19/22 11:00	1
Di-n-octyl phthalate	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Endosulfan I (Alpha)	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Endosulfan II (Beta)	ND	^3+	0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Endosulfan sulfate	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Endrin	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Endrin aldehyde	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1

Eurofins Eaton Monrovia

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

**Client Sample ID: AIEA GULCH WELLS PUMP 2  
(331-202-TP072)**

**Lab Sample ID: 380-13074-1**

**Date Collected: 07/27/22 11:36**

**Matrix: Drinking Water**

**Date Received: 07/29/22 10:00**

**PWSID Number: HI0000331**

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
EPTC	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Fluoranthene	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Fluorene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
gamma-Chlordane	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Heptachlor	ND		0.039	ug/L		08/01/22 08:52	08/19/22 11:00	1
Heptachlor epoxide (isomer B)	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Hexachlorobenzene	ND	^3+	0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Hexachlorocyclopentadiene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Indeno[1,2,3-cd]pyrene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Isophorone	ND		0.49	ug/L		08/01/22 08:52	08/19/22 11:00	1
Lindane	ND		0.039	ug/L		08/01/22 08:52	08/19/22 11:00	1
Malathion	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Methoxychlor	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Metolachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Metribuzin	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Molinate	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Naphthalene	ND		0.29	ug/L		08/01/22 08:52	08/19/22 11:00	1
Parathion	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Pendimethalin (Penoxaline)	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Total Permethrin (mixed isomers)	ND		0.20	ug/L		08/01/22 08:52	08/19/22 11:00	1
Phenanthrene	ND		0.039	ug/L		08/01/22 08:52	08/19/22 11:00	1
Propachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Pyrene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Simazine	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Terbacil	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Terbutylazine	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1
Thiobencarb	ND		0.20	ug/L		08/01/22 08:52	08/19/22 11:00	1
trans-Nonachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 11:00	1
Trifluralin	ND		0.098	ug/L		08/01/22 08:52	08/19/22 11:00	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
D-Limonene	0.71	T J N	ug/L		2.57	5989-27-5	08/01/22 08:52	08/19/22 11:00	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	94		70 - 130	08/01/22 08:52	08/19/22 11:00	1
Triphenylphosphate	101		70 - 130	08/01/22 08:52	08/19/22 11:00	1
Perylene-d12	101		70 - 130	08/01/22 08:52	08/19/22 11:00	1

**Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Acenaphthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Acenaphthylene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Anthracene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1

Eurofins Eaton Monrovia

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

**Client Sample ID: AIEA GULCH WELLS PUMP 2  
(331-202-TP072)**

**Lab Sample ID: 380-13074-1**

**Date Collected: 07/27/22 11:36**

**Matrix: Drinking Water**

**Date Received: 07/29/22 10:00**

**PWSID Number: HI0000331**

**Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Benzo[a]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Biphenyl	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Chrysene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Dibenzothiophene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		08/02/22 00:00	08/07/22 06:47	1
Fluoranthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Fluorene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Naphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Perylene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Phenanthrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1
Pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/07/22 06:47	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	87		45 - 118	08/02/22 00:00	08/07/22 06:47	1
(d10-Phenanthrene)	90		56 - 123	08/02/22 00:00	08/07/22 06:47	1
(d12-Chrysene)	84		36 - 142	08/02/22 00:00	08/07/22 06:47	1
(d12-Perylene)	87		36 - 161	08/02/22 00:00	08/07/22 06:47	1
(d8-Naphthalene)	81		20 - 112	08/02/22 00:00	08/07/22 06:47	1

**Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
DIESEL	ND	U	0.028		mg/L			08/05/22 19:12	1
MOTOR OIL	ND	U	0.056		mg/L			08/05/22 19:12	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
BROMOBENZENE	81		60 - 130		08/05/22 19:12	1
HEXACOSANE	98		60 - 130		08/05/22 19:12	1

**Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
GASOLINE	ND	U	0.02		mg/L			08/02/22 15:16	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
BROMOFLUOROBENZENE	90		60 - 140		08/02/22 15:16	1

# Action Limit Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-13074-1

**Client Sample ID: AIEA GULCH WELLS PUMP 2  
 (331-202-TP072)  
 PWSID Number: HI0000331**

**Lab Sample ID: 380-13074-1**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	ND		ug/L	2	0.049	525.2	Total/NA
Atrazine	ND		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	ND		ug/L	0.2	0.020	525.2	Total/NA
Di(2-ethylhexyl)adipate	ND	^3+ **	ug/L	400	0.59	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	ND		ug/L	6	0.59	525.2	Total/NA
Endrin	ND		ug/L	2	0.098	525.2	Total/NA
Heptachlor	ND		ug/L	0.4	0.039	525.2	Total/NA
Heptachlor epoxide (isomer B)	ND		ug/L	0.2	0.049	525.2	Total/NA
Hexachlorobenzene	ND	^3+	ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	ND		ug/L	50	0.049	525.2	Total/NA
Lindane	ND		ug/L	0.2	0.039	525.2	Total/NA
Methoxychlor	ND		ug/L	40	0.098	525.2	Total/NA
Simazine	ND		ug/L	4	0.049	525.2	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	TPP (70-130)	PRY (70-130)
380-13074-1	AIEA GULCH WELLS PUMP 2 (	94	101	101

**Surrogate Legend**  
 2NMX = 2-Nitro-m-xylene  
 TPP = Triphenylphosphate  
 PRY = Perylene-d12

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	TPP (70-130)	PRY (70-130)
380-12317-I-1-A DU	Duplicate	94	100	97
380-12602-S-6-A MS	Matrix Spike	96	103	102
LCS 380-10968/3-A	Lab Control Sample	96	100	100
LCS 380-10968/4-A	Lab Control Sample Dup	94	100	101
MB 380-10968/1-A	Method Blank	93	106	94
MRL 380-10968/2-A	Lab Control Sample	96	101	93

**Surrogate Legend**  
 2NMX = 2-Nitro-m-xylene  
 TPP = Triphenylphosphate  
 PRY = Perylene-d12

## Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)				
		ANT (45-118)	CRY (36-142)	NPT (20-112)	PHN (56-123)	PRY (36-161)
380-13074-1	AIEA GULCH WELLS PUMP 2 (	87	84	81	90	87

**Surrogate Legend**  
 ANT = (d10-Acenaphthene)  
 CRY = (d12-Chrysene)  
 NPT = (d8-Naphthalene)  
 PHN = (d10-Phenanthrene)  
 PRY = (d12-Perylene)

## Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Matrix: water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)				
		ANT (65-113)	CRY (60-139)	NPT (44-119)	PHN (80-111)	PRY (36-161)
98866-B1	Method Blank	95	95	91	98	93
98866-BS1	Lab Control Sample	92	95	86	95	97
98866-BS2	Lab Control Sample Dup	93	98	87	96	99

**Surrogate Legend**  
 ANT = (d10-Acenaphthene)  
 CRY = (d12-Chrysene)

Eurofins Eaton Monrovia

# Surrogate Summary

Client: City & County of Honolulu

Job ID: 380-13074-1

Project/Site: RED-HILL

NPT = (d8-Naphthalene)

PHN = (d10-Phenanthrene)

PRY = (d12-Perylene)

## Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BB (60-130)	XACOSAI (60-130)
380-13074-1	AIEA GULCH WELLS PUMP 2 (	81	98

#### Surrogate Legend

BB = BROMOBENZENE

HEXACOSANE = HEXACOSANE

## Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Matrix: WATER

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BB	XACOSAI
22DSH009WB	Method Blank		

#### Surrogate Legend

BB = BROMOBENZENE

HEXACOSANE = HEXACOSANE

## Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Matrix: WATER

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BB (60-130)	XACOSAI (60-130)
22DSH009WC	LCD	70	113
22DSH009WL	Lab Control Sample	78	110

#### Surrogate Legend

BB = BROMOBENZENE

HEXACOSANE = HEXACOSANE

## Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB (60-140)
380-13074-1	AIEA GULCH WELLS PUMP 2 (	90

#### Surrogate Legend

BFB = BROMOFLUOROBENZENE

## Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: WATER

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB
22VG39H01B	Method Blank	

#### Surrogate Legend

Eurofins Eaton Monrovia



# Surrogate Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL  
BFB = BROMOFLUOROBENZENE

Job ID: 380-13074-1

## Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: WATER

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB (70-130)
22VG39H01C	LCD	113
22VG39H01L	Lab Control Sample	113

#### Surrogate Legend

BFB = BROMOFLUOROBENZENE

## Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: WATER

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB (60-140)
22H004-01M	Matrix Spike	110
22H004-01S	Matrix Spike Duplicate	113

#### Surrogate Legend

BFB = BROMOFLUOROBENZENE

# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 380-10968/1-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
2,4'-DDD	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
2,4'-DDE	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
2,4'-DDT	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
2,4-Dinitrotoluene	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
2,6-Dinitrotoluene	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
4,4'-DDD	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
4,4'-DDE	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
4,4'-DDT	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Acenaphthene	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Acenaphthylene	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Acetochlor	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Alachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
alpha-BHC	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
alpha-Chlordane	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Anthracene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 10:40	1
Atrazine	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Benz(a)anthracene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Benzo[a]pyrene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 10:40	1
Benzo[b]fluoranthene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 10:40	1
Benzo[g,h,i]perylene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Benzo[k]fluoranthene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 10:40	1
beta-BHC	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Bromacil	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Butachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Butylbenzylphthalate	ND		0.49	ug/L		08/01/22 08:52	08/19/22 10:40	1
Caffeine	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Chlorobenzilate	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Chloroneb	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Chlorothalonil (Draconil, Bravo)	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Chlorpyrifos	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Chrysene	ND		0.020	ug/L		08/01/22 08:52	08/19/22 10:40	1
delta-BHC	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Di(2-ethylhexyl)adipate	ND		0.59	ug/L		08/01/22 08:52	08/19/22 10:40	1
Bis(2-ethylhexyl) phthalate	ND		0.59	ug/L		08/01/22 08:52	08/19/22 10:40	1
Diazinon (Qualitative)	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Dibenz(a,h)anthracene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Diclorvos (DDVP)	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Dieldrin	ND		0.20	ug/L		08/01/22 08:52	08/19/22 10:40	1
Diethylphthalate	ND		0.49	ug/L		08/01/22 08:52	08/19/22 10:40	1
Dimethoate	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Dimethylphthalate	ND		0.49	ug/L		08/01/22 08:52	08/19/22 10:40	1
Di-n-butyl phthalate	ND		0.99	ug/L		08/01/22 08:52	08/19/22 10:40	1
Di-n-octyl phthalate	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Endosulfan I (Alpha)	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Endosulfan II (Beta)	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Endosulfan sulfate	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Endrin	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Endrin aldehyde	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1

Eurofins Eaton Monrovia

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MB 380-10968/1-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
EPTC	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Fluoranthene	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Fluorene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
gamma-Chlordane	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Heptachlor	ND		0.039	ug/L		08/01/22 08:52	08/19/22 10:40	1
Heptachlor epoxide (isomer B)	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Hexachlorobenzene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Hexachlorocyclopentadiene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Indeno[1,2,3-cd]pyrene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Isophorone	ND		0.49	ug/L		08/01/22 08:52	08/19/22 10:40	1
Lindane	ND		0.039	ug/L		08/01/22 08:52	08/19/22 10:40	1
Malathion	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Methoxychlor	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Metolachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Metribuzin	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Molinate	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Naphthalene	ND		0.30	ug/L		08/01/22 08:52	08/19/22 10:40	1
Parathion	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Pendimethalin (Penoxaline)	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Total Permethrin (mixed isomers)	ND		0.20	ug/L		08/01/22 08:52	08/19/22 10:40	1
Phenanthrene	ND		0.039	ug/L		08/01/22 08:52	08/19/22 10:40	1
Propachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Pyrene	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Simazine	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Terbacil	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Terbutylazine	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1
Thiobencarb	ND		0.20	ug/L		08/01/22 08:52	08/19/22 10:40	1
trans-Nonachlor	ND		0.049	ug/L		08/01/22 08:52	08/19/22 10:40	1
Trifluralin	ND		0.099	ug/L		08/01/22 08:52	08/19/22 10:40	1

<i>Tentatively Identified Compound</i>	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Decane	2.03	T J N	ug/L		2.41	124-18-5	08/01/22 08:52	08/19/22 10:40	1
Tetradecanoic acid	1.09	T J N	ug/L		5.83	544-63-8	08/01/22 08:52	08/19/22 10:40	1
Octadecanoic acid	0.684	T J N	ug/L		6.51	57-11-4	08/01/22 08:52	08/19/22 10:40	1
Hexadecanamide	1.55	T J N	ug/L		6.66	629-54-9	08/01/22 08:52	08/19/22 10:40	1
Octadecanamide	1.06	T J N	ug/L		7.60	124-26-5	08/01/22 08:52	08/19/22 10:40	1
13-Docosenamide, (Z)-	1.02	T J N	ug/L		10.10	112-84-5	08/01/22 08:52	08/19/22 10:40	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	93		70 - 130	08/01/22 08:52	08/19/22 10:40	1
Triphenylphosphate	106		70 - 130	08/01/22 08:52	08/19/22 10:40	1
Perylene-d12	94		70 - 130	08/01/22 08:52	08/19/22 10:40	1

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 380-10968/3-A**

**Matrix: Water**

**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample**

**Prep Type: Total/NA**

**Prep Batch: 10968**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	1.97	2.14		ug/L		109	70 - 130
2,4'-DDE	1.97	1.92		ug/L		98	70 - 130
2,4'-DDT	1.97	1.86		ug/L		94	70 - 130
2,4-Dinitrotoluene	1.97	2.05		ug/L		104	70 - 130
2,6-Dinitrotoluene	1.97	2.01		ug/L		102	70 - 130
4,4'-DDD	1.97	1.86		ug/L		94	70 - 130
4,4'-DDE	1.97	1.94		ug/L		99	70 - 130
4,4'-DDT	1.97	2.04		ug/L		103	70 - 130
Acenaphthene	1.97	1.85		ug/L		94	70 - 130
Acenaphthylene	1.97	1.91		ug/L		97	70 - 130
Acetochlor	1.97	2.01		ug/L		102	70 - 130
Alachlor	1.97	1.88		ug/L		96	70 - 130
alpha-BHC	1.97	1.83		ug/L		93	70 - 130
alpha-Chlordane	1.97	1.86		ug/L		94	70 - 130
Anthracene	1.97	1.90		ug/L		97	70 - 130
Atrazine	1.97	2.09		ug/L		106	70 - 130
Benz(a)anthracene	1.97	1.88		ug/L		96	70 - 130
Benzo[a]pyrene	1.97	1.69		ug/L		86	70 - 130
Benzo[b]fluoranthene	1.97	1.84		ug/L		93	70 - 130
Benzo[g,h,i]perylene	1.97	1.81		ug/L		92	70 - 130
Benzo[k]fluoranthene	1.97	1.84		ug/L		93	70 - 130
beta-BHC	1.97	1.79		ug/L		91	70 - 130
Bromacil	1.97	1.96		ug/L		99	70 - 130
Butachlor	1.97	2.06		ug/L		104	70 - 130
Butylbenzylphthalate	1.97	2.31		ug/L		118	70 - 130
Caffeine	1.97	1.62		ug/L		82	45 - 137
Chlorobenzilate	1.97	2.07		ug/L		105	70 - 130
Chloroneb	1.97	1.84		ug/L		93	70 - 130
Chlorothalonil (Draconil, Bravo)	1.97	2.26		ug/L		115	70 - 130
Chlorpyrifos	1.97	1.95		ug/L		99	70 - 130
Chrysene	1.97	1.65		ug/L		84	70 - 130
delta-BHC	1.97	1.75		ug/L		89	70 - 130
Di(2-ethylhexyl)adipate	1.97	2.68	*+	ug/L		136	70 - 130
Bis(2-ethylhexyl) phthalate	1.97	2.16		ug/L		110	70 - 130
Diazinon (Qualitative)	1.97	1.91		ug/L		97	15 - 132
Dibenz(a,h)anthracene	1.97	1.90		ug/L		97	70 - 130
Diclorvos (DDVP)	1.97	2.34		ug/L		119	70 - 130
Dieldrin	1.97	2.04		ug/L		104	70 - 130
Diethylphthalate	1.97	2.03		ug/L		103	70 - 130
Dimethoate	1.97	1.70		ug/L		87	35 - 100
Dimethylphthalate	1.97	2.09		ug/L		106	70 - 130
Di-n-butyl phthalate	3.94	4.06		ug/L		103	70 - 130
Di-n-octyl phthalate	1.97	1.83		ug/L		93	70 - 130
Endosulfan I (Alpha)	1.97	1.67		ug/L		85	70 - 130
Endosulfan II (Beta)	1.97	1.97		ug/L		100	70 - 130
Endosulfan sulfate	1.97	2.22		ug/L		113	70 - 130
Endrin	1.97	1.95		ug/L		99	70 - 130
Endrin aldehyde	1.97	1.77		ug/L		90	70 - 130

Eurofins Eaton Monrovia

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 380-10968/3-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
EPTC	1.97	2.21		ug/L		112	70 - 130
Fluoranthene	1.97	1.86		ug/L		94	70 - 130
Fluorene	1.97	2.07		ug/L		105	70 - 130
gamma-Chlordane	1.97	1.92		ug/L		97	70 - 130
Heptachlor	1.97	1.97		ug/L		100	70 - 130
Heptachlor epoxide (isomer B)	1.97	2.02		ug/L		102	70 - 130
Hexachlorobenzene	1.97	1.91		ug/L		97	70 - 130
Hexachlorocyclopentadiene	1.97	1.86		ug/L		94	70 - 130
Indeno[1,2,3-cd]pyrene	1.97	1.84		ug/L		94	70 - 130
Isophorone	1.97	1.97		ug/L		100	70 - 130
Lindane	1.97	1.80		ug/L		91	70 - 130
Malathion	1.97	2.10		ug/L		107	70 - 130
Methoxychlor	1.97	2.07		ug/L		105	70 - 130
Metolachlor	1.97	1.93		ug/L		98	70 - 130
Metribuzin	1.97	1.97		ug/L		100	70 - 130
Molinate	1.97	2.24		ug/L		114	70 - 130
Naphthalene	1.97	2.05		ug/L		104	70 - 130
Parathion	1.97	2.25		ug/L		114	70 - 130
Pendimethalin (Penoxaline)	1.97	2.34		ug/L		119	70 - 130
Phenanthrene	1.97	1.93		ug/L		98	70 - 130
Propachlor	1.97	2.27		ug/L		115	70 - 130
Pyrene	1.97	1.96		ug/L		99	70 - 130
Simazine	1.97	2.13		ug/L		108	70 - 130
Terbacil	1.97	2.25		ug/L		114	70 - 130
Terbutylazine	1.97	1.94		ug/L		98	70 - 130
Thiobencarb	1.97	2.06		ug/L		105	70 - 130
trans-Nonachlor	1.97	1.87		ug/L		95	70 - 130
Trifluralin	1.97	2.47		ug/L		125	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2-Nitro-m-xylene	96		70 - 130
Triphenylphosphate	100		70 - 130
Perylene-d12	100		70 - 130

**Lab Sample ID: LCSD 380-10968/4-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
2,4'-DDD	1.97	2.14		ug/L		108	70 - 130	0	20
2,4'-DDE	1.97	1.90		ug/L		97	70 - 130	1	20
2,4'-DDT	1.97	1.83		ug/L		93	70 - 130	1	20
2,4-Dinitrotoluene	1.97	1.97		ug/L		100	70 - 130	4	20
2,6-Dinitrotoluene	1.97	1.93		ug/L		98	70 - 130	4	20
4,4'-DDD	1.97	1.82		ug/L		93	70 - 130	2	20
4,4'-DDE	1.97	1.94		ug/L		99	70 - 130	0	20
4,4'-DDT	1.97	2.00		ug/L		101	70 - 130	2	20
Acenaphthene	1.97	1.82		ug/L		93	70 - 130	1	20

Eurofins Eaton Monrovia

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCSD 380-10968/4-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec		RPD	RPD Limit
							Limits	RPD		
Acenaphthylene	1.97	1.91		ug/L		97	70 - 130	0	20	
Acetochlor	1.97	2.03		ug/L		103	70 - 130	1	20	
Alachlor	1.97	1.85		ug/L		94	70 - 130	2	20	
alpha-BHC	1.97	1.81		ug/L		92	70 - 130	1	20	
alpha-Chlordane	1.97	1.88		ug/L		95	70 - 130	1	20	
Anthracene	1.97	1.88		ug/L		95	70 - 130	1	20	
Atrazine	1.97	2.10		ug/L		107	70 - 130	1	20	
Benz(a)anthracene	1.97	1.89		ug/L		96	70 - 130	0	20	
Benzo[a]pyrene	1.97	1.65		ug/L		84	70 - 130	2	20	
Benzo[b]fluoranthene	1.97	1.75		ug/L		89	70 - 130	5	20	
Benzo[g,h,i]perylene	1.97	1.70		ug/L		86	70 - 130	6	20	
Benzo[k]fluoranthene	1.97	1.79		ug/L		91	70 - 130	3	20	
beta-BHC	1.97	1.73		ug/L		88	70 - 130	3	20	
Bromacil	1.97	1.90		ug/L		96	70 - 130	3	20	
Butachlor	1.97	2.06		ug/L		104	70 - 130	0	20	
Butylbenzylphthalate	1.97	2.27		ug/L		115	70 - 130	2	20	
Caffeine	1.97	1.60		ug/L		81	45 - 137	1	20	
Chlorobenzilate	1.97	2.10		ug/L		106	70 - 130	1	20	
Chloroneb	1.97	1.78		ug/L		90	70 - 130	3	20	
Chlorothalonil (Draconil, Bravo)	1.97	2.27		ug/L		115	70 - 130	0	20	
Chlorpyrifos	1.97	1.98		ug/L		100	70 - 130	2	20	
Chrysene	1.97	1.65		ug/L		84	70 - 130	0	20	
delta-BHC	1.97	1.74		ug/L		88	70 - 130	1	20	
Di(2-ethylhexyl)adipate	1.97	2.67	+	ug/L		135	70 - 130	0	20	
Bis(2-ethylhexyl) phthalate	1.97	2.05		ug/L		104	70 - 130	6	20	
Diazinon (Qualitative)	1.97	1.90		ug/L		96	15 - 132	1	20	
Dibenz(a,h)anthracene	1.97	1.73		ug/L		88	70 - 130	9	20	
Diclorvos (DDVP)	1.97	2.25		ug/L		114	70 - 130	4	20	
Dieldrin	1.97	2.08		ug/L		105	70 - 130	2	20	
Diethylphthalate	1.97	1.97		ug/L		100	70 - 130	3	20	
Dimethoate	1.97	1.75		ug/L		89	35 - 100	3	20	
Dimethylphthalate	1.97	1.99		ug/L		101	70 - 130	5	20	
Di-n-butyl phthalate	3.94	3.95		ug/L		100	70 - 130	3	20	
Di-n-octyl phthalate	1.97	1.69		ug/L		86	70 - 130	8	20	
Endosulfan I (Alpha)	1.97	1.67		ug/L		84	70 - 130	0	20	
Endosulfan II (Beta)	1.97	1.98		ug/L		101	70 - 130	1	20	
Endosulfan sulfate	1.97	2.24		ug/L		114	70 - 130	1	20	
Endrin	1.97	1.97		ug/L		100	70 - 130	1	20	
Endrin aldehyde	1.97	1.86		ug/L		94	70 - 130	5	20	
EPTC	1.97	2.21		ug/L		112	70 - 130	0	20	
Fluoranthene	1.97	1.87		ug/L		95	70 - 130	1	20	
Fluorene	1.97	2.01		ug/L		102	70 - 130	3	20	
gamma-Chlordane	1.97	1.95		ug/L		99	70 - 130	2	20	
Heptachlor	1.97	1.98		ug/L		100	70 - 130	1	20	
Heptachlor epoxide (isomer B)	1.97	2.02		ug/L		102	70 - 130	0	20	
Hexachlorobenzene	1.97	1.90		ug/L		96	70 - 130	1	20	
Hexachlorocyclopentadiene	1.97	1.85		ug/L		94	70 - 130	0	20	
Indeno[1,2,3-cd]pyrene	1.97	1.76		ug/L		89	70 - 130	5	20	
Isophorone	1.97	1.88		ug/L		96	70 - 130	4	20	

Eurofins Eaton Monrovia

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCSD 380-10968/4-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lindane	1.97	1.77		ug/L		90	70 - 130	1	20
Malathion	1.97	2.06		ug/L		105	70 - 130	2	20
Methoxychlor	1.97	2.01		ug/L		102	70 - 130	3	20
Metolachlor	1.97	1.90		ug/L		96	70 - 130	2	20
Metribuzin	1.97	1.97		ug/L		100	70 - 130	0	20
Molinate	1.97	2.15		ug/L		109	70 - 130	4	20
Naphthalene	1.97	1.99		ug/L		101	70 - 130	3	20
Parathion	1.97	2.20		ug/L		111	70 - 130	2	20
Pendimethalin (Penoxaline)	1.97	2.30		ug/L		117	70 - 130	2	20
Phenanthrene	1.97	1.84		ug/L		93	70 - 130	5	20
Propachlor	1.97	2.24		ug/L		113	70 - 130	1	20
Pyrene	1.97	1.92		ug/L		97	70 - 130	2	20
Simazine	1.97	2.13		ug/L		108	70 - 130	0	20
Terbacil	1.97	2.11		ug/L		107	70 - 130	6	20
Terbutylazine	1.97	1.93		ug/L		98	70 - 130	0	20
Thiobencarb	1.97	2.09		ug/L		106	70 - 130	1	20
trans-Nonachlor	1.97	1.91		ug/L		97	70 - 130	2	20
Trifluralin	1.97	2.48		ug/L		126	70 - 130	1	20

Surrogate	LCSD %Recovery	LCSD Qualifier	LCSD Limits
2-Nitro-m-xylene	94		70 - 130
Triphenylphosphate	100		70 - 130
Perylene-d12	101		70 - 130

**Lab Sample ID: MRL 380-10968/2-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	0.0987	0.130		ug/L		132	50 - 150
2,4'-DDE	0.0987	0.0869	J	ug/L		88	50 - 150
2,4'-DDT	0.0987	0.0805	J	ug/L		82	50 - 150
2,4-Dinitrotoluene	0.0987	0.123		ug/L		125	50 - 150
2,6-Dinitrotoluene	0.0987	0.109		ug/L		110	50 - 150
4,4'-DDD	0.0987	0.0833	J	ug/L		84	50 - 150
4,4'-DDE	0.0987	0.151	^3+	ug/L		153	50 - 150
4,4'-DDT	0.0987	0.0855	J	ug/L		87	50 - 150
Acenaphthene	0.0987	0.0942	J	ug/L		95	50 - 150
Acenaphthylene	0.0987	0.0762	J	ug/L		77	50 - 150
Acetochlor	0.0494	0.0448	J	ug/L		91	50 - 150
Alachlor	0.0494	0.0590		ug/L		120	50 - 150
alpha-BHC	0.0987	0.101		ug/L		102	50 - 150
alpha-Chlordane	0.0494	0.0453	J	ug/L		92	50 - 150
Anthracene	0.0197	0.0201		ug/L		102	50 - 150
Atrazine	0.0494	0.0489	J	ug/L		99	50 - 150
Benz(a)anthracene	0.0494	0.0398	J	ug/L		81	50 - 150
Benzo[a]pyrene	0.0197	0.0133	J	ug/L		67	50 - 150
Benzo[b]fluoranthene	0.0197	0.0166	J	ug/L		84	50 - 150

Eurofins Eaton Monrovia



# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MRL 380-10968/2-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Benzo[g,h,i]perylene	0.0494	0.0320	J	ug/L		65	50 - 150
Benzo[k]fluoranthene	0.0197	ND		ug/L		75	50 - 150
beta-BHC	0.0987	0.102		ug/L		103	50 - 150
Bromacil	0.0987	0.109		ug/L		110	50 - 150
Butachlor	0.0494	0.0648		ug/L		131	50 - 150
Butylbenzylphthalate	0.148	0.577	^3+	ug/L		389	50 - 150
Caffeine	0.0494	0.0451	J	ug/L		91	50 - 150
Chlorobenzilate	0.0987	0.0831	J	ug/L		84	50 - 150
Chloroneb	0.0987	0.0917	J	ug/L		93	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0987	0.0912	J	ug/L		92	50 - 150
Chlorpyrifos	0.0494	0.0486	J	ug/L		98	50 - 150
Chrysene	0.0197	0.0156	J	ug/L		79	50 - 150
delta-BHC	0.0987	0.113		ug/L		114	50 - 150
Di(2-ethylhexyl)adipate	0.296	0.646	^3+	ug/L		218	50 - 150
Bis(2-ethylhexyl) phthalate	0.592	0.725		ug/L		122	50 - 150
Diazinon (Qualitative)	0.0987	0.0997		ug/L		101	15 - 132
Dibenz(a,h)anthracene	0.0494	ND		ug/L		60	50 - 150
Diclorvos (DDVP)	0.0494	0.0570		ug/L		116	50 - 150
Dieldrin	0.0987	0.124	J	ug/L		126	50 - 150
Diethylphthalate	0.148	0.171	J	ug/L		116	50 - 150
Dimethoate	0.0987	0.0808	J	ug/L		82	35 - 100
Dimethylphthalate	0.296	0.299	J	ug/L		101	50 - 150
Di-n-butyl phthalate	0.296	0.361	J	ug/L		122	49 - 243
Di-n-octyl phthalate	0.0987	0.0986	J	ug/L		100	50 - 150
Endosulfan I (Alpha)	0.0987	0.0896	J	ug/L		91	50 - 150
Endosulfan II (Beta)	0.0987	0.445	^3+	ug/L		451	50 - 150
Endosulfan sulfate	0.0987	0.113		ug/L		114	50 - 150
Endrin	0.0987	0.138		ug/L		139	50 - 150
Endrin aldehyde	0.0987	ND		ug/L		81	50 - 150
EPTC	0.0987	0.0971	J	ug/L		98	50 - 150
Fluoranthene	0.0494	0.0481	J	ug/L		97	50 - 150
Fluorene	0.0494	0.0537		ug/L		109	50 - 150
gamma-Chlordane	0.0494	0.0492		ug/L		100	50 - 150
Heptachlor	0.0395	0.0418		ug/L		106	50 - 150
Heptachlor epoxide (isomer B)	0.0494	0.0554		ug/L		112	50 - 150
Hexachlorobenzene	0.0494	0.0846	^3+	ug/L		171	50 - 150
Hexachlorocyclopentadiene	0.0494	0.0393	J	ug/L		80	50 - 150
Indeno[1,2,3-cd]pyrene	0.0494	0.0344	J	ug/L		70	50 - 150
Isophorone	0.0987	0.0942	J	ug/L		95	50 - 150
Lindane	0.0494	0.0434		ug/L		88	50 - 150
Malathion	0.0987	0.104		ug/L		105	50 - 150
Methoxychlor	0.0987	0.0860	J	ug/L		87	50 - 150
Metolachlor	0.0494	0.0533		ug/L		108	50 - 150
Metribuzin	0.0494	0.0455	J	ug/L		92	50 - 150
Molinate	0.0987	0.104		ug/L		106	50 - 150
Naphthalene	0.0987	0.101	J	ug/L		102	50 - 150
Parathion	0.0987	0.143		ug/L		145	50 - 150
Pendimethalin (Penoxaline)	0.0987	0.0890	J	ug/L		90	50 - 150
Phenanthrene	0.0197	0.0246	J	ug/L		124	50 - 150

Eurofins Eaton Monrovia



# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MRL 380-10968/2-A**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Propachlor	0.0494	0.0536		ug/L		109	50 - 150
Pyrene	0.0494	0.0502		ug/L		102	50 - 150
Simazine	0.0494	0.0567		ug/L		115	50 - 150
Terbacil	0.0987	0.111		ug/L		112	50 - 150
Terbutylazine	0.0987	0.0883	J	ug/L		89	50 - 150
Thiobencarb	0.0987	0.108	J	ug/L		109	50 - 150
trans-Nonachlor	0.0494	0.0452	J	ug/L		92	50 - 150
Trifluralin	0.0987	0.0918	J	ug/L		93	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	96		70 - 130
Triphenylphosphate	101		70 - 130
Perylene-d12	93		70 - 130

**Lab Sample ID: 380-12602-S-6-A MS**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	ND		1.97	2.17		ug/L		110	70 - 130
2,4'-DDE	ND		1.97	1.88		ug/L		95	70 - 130
2,4'-DDT	ND		1.97	1.90		ug/L		96	70 - 130
2,4-Dinitrotoluene	ND		1.97	1.97		ug/L		100	70 - 130
2,6-Dinitrotoluene	ND		1.97	1.93		ug/L		98	70 - 130
4,4'-DDD	ND		1.97	1.94		ug/L		98	70 - 130
4,4'-DDE	ND	^3+	1.97	1.94		ug/L		99	70 - 130
4,4'-DDT	ND		1.97	2.01		ug/L		102	70 - 130
Acenaphthene	ND		1.97	1.86		ug/L		94	70 - 130
Acenaphthylene	ND		1.97	1.95		ug/L		99	70 - 130
Acetochlor	ND		1.97	2.05		ug/L		104	70 - 130
Alachlor	ND		1.97	1.90		ug/L		96	70 - 130
alpha-BHC	ND		1.97	1.83		ug/L		93	70 - 130
alpha-Chlordane	ND		1.97	1.90		ug/L		96	70 - 130
Anthracene	ND		1.97	1.45		ug/L		73	70 - 130
Atrazine	ND		1.97	2.16		ug/L		110	70 - 130
Benz(a)anthracene	ND		1.97	1.96		ug/L		99	70 - 130
Benzo[a]pyrene	ND		1.97	1.62		ug/L		82	70 - 130
Benzo[b]fluoranthene	ND		1.97	1.85		ug/L		93	70 - 130
Benzo[g,h,i]perylene	ND		1.97	1.72		ug/L		87	70 - 130
Benzo[k]fluoranthene	ND		1.97	1.94		ug/L		98	70 - 130
beta-BHC	ND		1.97	1.80		ug/L		91	70 - 130
Bromacil	ND		1.97	1.96		ug/L		99	70 - 130
Butachlor	ND		1.97	2.07		ug/L		105	70 - 130
Butylbenzylphthalate	ND	^3+	1.97	2.10		ug/L		106	70 - 130
Caffeine	ND		1.97	1.66		ug/L		84	46 - 144
Chlorobenzilate	ND		1.97	2.11		ug/L		107	70 - 130
Chloroneb	ND		1.97	1.90		ug/L		96	70 - 130
Chlorothalonil (Draconil, Bravo)	ND		1.97	2.29		ug/L		116	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-12602-S-6-A MS**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Chlorpyrifos	ND		1.97	2.01		ug/L		102	70 - 130
Chrysene	ND		1.97	1.73		ug/L		88	70 - 130
delta-BHC	ND		1.97	1.77		ug/L		90	70 - 130
Di(2-ethylhexyl)adipate	ND	^3+ *+	1.97	2.18		ug/L		89	70 - 130
Bis(2-ethylhexyl) phthalate	ND		1.97	1.83		ug/L		93	70 - 130
Diazinon (Qualitative)	ND		1.97	1.97		ug/L		100	15 - 132
Dibenz(a,h)anthracene	ND		1.97	1.74		ug/L		88	70 - 130
Diclorvos (DDVP)	ND		1.97	2.38		ug/L		121	70 - 130
Dieldrin	ND		1.97	2.05		ug/L		104	70 - 130
Diethylphthalate	ND		1.97	2.02		ug/L		102	70 - 130
Dimethoate	ND		1.97	1.62		ug/L		82	34 - 111
Dimethylphthalate	ND		1.97	2.01		ug/L		102	70 - 130
Di-n-butyl phthalate	ND		3.95	4.02		ug/L		102	70 - 130
Di-n-octyl phthalate	ND		1.97	1.59		ug/L		81	70 - 130
Endosulfan I (Alpha)	ND		1.97	1.72		ug/L		87	70 - 130
Endosulfan II (Beta)	ND	^3+	1.97	2.22		ug/L		112	70 - 130
Endosulfan sulfate	ND		1.97	2.16		ug/L		110	70 - 130
Endrin	ND		1.97	1.97		ug/L		100	70 - 130
Endrin aldehyde	ND		1.97	1.50		ug/L		76	70 - 130
EPTC	ND		1.97	2.28		ug/L		116	70 - 130
Fluoranthene	ND		1.97	1.94		ug/L		98	70 - 130
Fluorene	ND		1.97	2.05		ug/L		104	70 - 130
gamma-Chlordane	ND		1.97	1.96		ug/L		99	70 - 130
Heptachlor	ND		1.97	1.96		ug/L		99	70 - 130
Heptachlor epoxide (isomer B)	ND		1.97	2.08		ug/L		105	70 - 130
Hexachlorobenzene	ND	^3+	1.97	1.95		ug/L		99	70 - 130
Hexachlorocyclopentadiene	ND		1.97	1.99		ug/L		101	70 - 130
Indeno[1,2,3-cd]pyrene	ND		1.97	1.80		ug/L		91	70 - 130
Isophorone	ND		1.97	1.97		ug/L		100	70 - 130
Lindane	ND		1.97	1.83		ug/L		93	70 - 130
Malathion	ND		1.97	2.15		ug/L		109	70 - 130
Methoxychlor	ND		1.97	2.17		ug/L		110	70 - 130
Metolachlor	ND		1.97	1.96		ug/L		99	70 - 130
Metribuzin	ND		1.97	2.03		ug/L		103	70 - 130
Molinate	ND		1.97	2.30		ug/L		116	70 - 130
Naphthalene	ND		1.97	2.03		ug/L		103	70 - 130
Parathion	ND		1.97	2.29		ug/L		116	70 - 130
Pendimethalin (Penoxaline)	ND		1.97	2.40		ug/L		122	70 - 130
Phenanthrene	ND		1.97	1.92		ug/L		97	70 - 130
Propachlor	ND		1.97	2.32		ug/L		117	70 - 130
Pyrene	ND		1.97	2.01		ug/L		102	70 - 130
Simazine	ND		1.97	2.25		ug/L		114	70 - 130
Terbacil	ND		1.97	2.16		ug/L		110	70 - 130
Terbutylazine	ND		1.97	2.09		ug/L		106	70 - 130
Thiobencarb	ND		1.97	2.04		ug/L		103	70 - 130
trans-Nonachlor	ND		1.97	1.88		ug/L		95	70 - 130
Trifluralin	ND		1.97	2.58		ug/L		130	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-12602-S-6-A MS**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Surrogate	%Recovery	MS MS Qualifier	Limits
2-Nitro-m-xylene	96		70 - 130
Triphenylphosphate	103		70 - 130
Perylene-d12	102		70 - 130

**Lab Sample ID: 380-12317-I-1-A DU**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
2,4'-DDD	ND		ND		ug/L		NC	20
2,4'-DDE	ND		ND		ug/L		NC	20
2,4'-DDT	ND		ND		ug/L		NC	20
2,4-Dinitrotoluene	ND		ND		ug/L		NC	20
2,6-Dinitrotoluene	ND		ND		ug/L		NC	20
4,4'-DDD	ND		ND		ug/L		NC	20
4,4'-DDE	ND	^3+	ND		ug/L		NC	20
4,4'-DDT	ND		ND		ug/L		NC	20
Acenaphthene	ND		ND		ug/L		NC	20
Acenaphthylene	ND		ND		ug/L		NC	20
Acetochlor	ND		ND		ug/L		NC	20
Alachlor	ND		ND		ug/L		NC	20
alpha-BHC	ND		ND		ug/L		NC	20
alpha-Chlordane	ND		ND		ug/L		NC	20
Anthracene	ND		ND		ug/L		NC	20
Atrazine	ND		ND		ug/L		NC	20
Benz(a)anthracene	ND		ND		ug/L		NC	20
Benzo[a]pyrene	ND		ND		ug/L		NC	20
Benzo[b]fluoranthene	ND		ND		ug/L		NC	20
Benzo[g,h,i]perylene	ND		ND		ug/L		NC	20
Benzo[k]fluoranthene	ND		ND		ug/L		NC	20
beta-BHC	ND		ND		ug/L		NC	20
Bromacil	ND		ND		ug/L		NC	20
Butachlor	ND		ND		ug/L		NC	20
Butylbenzylphthalate	ND	^3+	ND		ug/L		NC	20
Caffeine	ND		ND		ug/L		NC	20
Chlorobenzilate	ND		ND		ug/L		NC	20
Chloroneb	ND		ND		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	ND		ND		ug/L		NC	20
Chlorpyrifos	ND		ND		ug/L		NC	20
Chrysene	ND		ND		ug/L		NC	20
delta-BHC	ND		ND		ug/L		NC	20
Di(2-ethylhexyl)adipate	ND	^3+ *+	ND	*+	ug/L		NC	20
Bis(2-ethylhexyl) phthalate	ND		ND		ug/L		NC	20
Diazinon (Qualitative)	ND		ND		ug/L		NC	20
Dibenz(a,h)anthracene	ND		ND		ug/L		NC	20
Diclorvos (DDVP)	ND		ND		ug/L		NC	20
Dieldrin	ND		ND		ug/L		NC	20
Diethylphthalate	ND		ND		ug/L		NC	20

# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-12317-I-1-A DU**  
**Matrix: Water**  
**Analysis Batch: 14003**

**Client Sample ID: Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 10968**

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
Dimethoate	ND		ND		ug/L		NC	20
Dimethylphthalate	ND		ND		ug/L		NC	20
Di-n-butyl phthalate	ND		ND		ug/L		NC	20
Di-n-octyl phthalate	ND		ND		ug/L		NC	20
Endosulfan I (Alpha)	ND		ND		ug/L		NC	20
Endosulfan II (Beta)	ND	^3+	ND		ug/L		NC	20
Endosulfan sulfate	ND		ND		ug/L		NC	20
Endrin	ND		ND		ug/L		NC	20
Endrin aldehyde	ND		ND		ug/L		NC	20
EPTC	ND		ND		ug/L		NC	20
Fluoranthene	ND		ND		ug/L		NC	20
Fluorene	ND		ND		ug/L		NC	20
gamma-Chlordane	ND		ND		ug/L		NC	20
Heptachlor	ND		ND		ug/L		NC	20
Heptachlor epoxide (isomer B)	ND		ND		ug/L		NC	20
Hexachlorobenzene	ND	^3+	ND		ug/L		NC	20
Hexachlorocyclopentadiene	ND		ND		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	ND		ND		ug/L		NC	20
Isophorone	ND		ND		ug/L		NC	20
Lindane	ND		ND		ug/L		NC	20
Malathion	ND		ND		ug/L		NC	20
Methoxychlor	ND		ND		ug/L		NC	20
Metolachlor	ND		ND		ug/L		NC	20
Metribuzin	ND		ND		ug/L		NC	20
Molinate	ND		ND		ug/L		NC	20
Naphthalene	ND		ND		ug/L		NC	20
Parathion	ND		ND		ug/L		NC	20
Pendimethalin (Penoxaline)	ND		ND		ug/L		NC	20
Total Permethrin (mixed isomers)	ND		ND		ug/L		NC	20
Phenanthrene	ND		ND		ug/L		NC	20
Propachlor	ND		ND		ug/L		NC	20
Pyrene	ND		ND		ug/L		NC	20
Simazine	ND		ND		ug/L		NC	20
Terbacil	ND		ND		ug/L		NC	20
Terbutylazine	ND		ND		ug/L		NC	20
Thiobencarb	ND		ND		ug/L		NC	20
trans-Nonachlor	ND		ND		ug/L		NC	20
Trifluralin	ND		ND		ug/L		NC	20

Surrogate	%Recovery	DU Qualifier	DU Limits
2-Nitro-m-xylene	94		70 - 130
Triphenylphosphate	100		70 - 130
Perylene-d12	97		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

**Lab Sample ID: 98866-B1**  
**Matrix: water**  
**Analysis Batch: O-38076**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: O-38076\_P**

Analyte	Blank Result	Blank Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Acenaphthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Acenaphthylene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Anthracene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Biphenyl	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Chrysene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Dibenzothiophene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		08/02/22 00:00	08/06/22 16:58	1
Fluoranthene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Fluorene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Naphthalene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Perylene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Phenanthrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1
Pyrene	ND		0.005	0.001	µg/L		08/02/22 00:00	08/06/22 16:58	1

Surrogate	Blank %Recovery	Blank Qualifier	Limits	Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	95		65 - 113	08/02/22 00:00	08/06/22 16:58	1
(d10-Phenanthrene)	98		80 - 111	08/02/22 00:00	08/06/22 16:58	1
(d12-Chrysene)	95		60 - 139	08/02/22 00:00	08/06/22 16:58	1
(d12-Perylene)	93		36 - 161	08/02/22 00:00	08/06/22 16:58	1
(d8-Naphthalene)	91		44 - 119	08/02/22 00:00	08/06/22 16:58	1

**Lab Sample ID: 98866-BS1**  
**Matrix: water**  
**Analysis Batch: O-38076**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: O-38076\_P**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.5	0.438		µg/L		88	49 - 117
1-Methylphenanthrene	0.5	0.437		µg/L		87	66 - 127
2,3,5-Trimethylnaphthalene	0.5	0.426		µg/L		85	57 - 120
2,6-Dimethylnaphthalene	0.5	0.44		µg/L		88	54 - 117
2-Methylnaphthalene	0.5	0.452		µg/L		90	47 - 130
Acenaphthene	0.5	0.433		µg/L		87	53 - 131
Acenaphthylene	0.5	0.432		µg/L		86	43 - 140
Anthracene	0.5	0.442		µg/L		88	58 - 135

Eurofins Eaton Monrovia

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

**Lab Sample ID: 98866-BS1**  
**Matrix: water**  
**Analysis Batch: O-38076**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: O-38076\_P**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Benz[a]anthracene	0.5	0.491		µg/L		98	55 - 145
Benzo[a]pyrene	0.5	0.397		µg/L		79	51 - 143
Benzo[b]fluoranthene	0.5	0.579		µg/L		116	46 - 165
Benzo[e]pyrene	0.5	0.516		µg/L		103	42 - 152
Benzo[g,h,i]perylene	0.5	0.445		µg/L		89	63 - 133
Benzo[k]fluoranthene	0.5	0.503		µg/L		101	56 - 145
Biphenyl	0.5	0.438		µg/L		88	56 - 119
Chrysene	0.5	0.418		µg/L		84	56 - 141
Dibenz[a,h]anthracene	0.5	0.552		µg/L		110	55 - 150
Dibenzo[a,l]pyrene	0.5	0.502		µg/L		100	50 - 150
Dibenzothiophene	0.5	0.446		µg/L		89	75 - 113
Disalicylidenepropanediamine	50	34.7		µg/L		69	50 - 150
Fluoranthene	0.5	0.449		µg/L		90	60 - 146
Fluorene	0.5	0.427		µg/L		85	58 - 131
Indeno[1,2,3-cd]pyrene	0.5	0.581		µg/L		116	50 - 151
Naphthalene	0.5	0.417		µg/L		83	41 - 126
Perylene	0.5	0.476		µg/L		95	48 - 141
Phenanthrene	0.5	0.446		µg/L		89	67 - 127
Pyrene	0.5	0.451		µg/L		90	54 - 156

Surrogate	LCS %Recovery	LCS Qualifier	Limits
(d10-Acenaphthene)	92		65 - 113
(d10-Phenanthrene)	95		80 - 111
(d12-Chrysene)	95		60 - 139
(d12-Perylene)	97		36 - 161
(d8-Naphthalene)	86		44 - 119

**Lab Sample ID: 98866-BS2**  
**Matrix: water**  
**Analysis Batch: O-38076**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: O-38076\_P**

Analyte	Spike Added	LCS DUP Result	LCS DUP Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	0.5	0.437		µg/L		87	49 - 117	1	30
1-Methylphenanthrene	0.5	0.455		µg/L		91	66 - 127	4	30
2,3,5-Trimethylnaphthalene	0.5	0.435		µg/L		87	57 - 120	2	30
2,6-Dimethylnaphthalene	0.5	0.444		µg/L		89	54 - 117	1	30
2-Methylnaphthalene	0.5	0.45		µg/L		90	47 - 130	0	30
Acenaphthene	0.5	0.436		µg/L		87	53 - 131	0	30
Acenaphthylene	0.5	0.437		µg/L		87	43 - 140	1	30
Anthracene	0.5	0.45		µg/L		90	58 - 135	2	30
Benz[a]anthracene	0.5	0.523		µg/L		105	55 - 145	7	30
Benzo[a]pyrene	0.5	0.405		µg/L		81	51 - 143	2	30
Benzo[b]fluoranthene	0.5	0.601		µg/L		120	46 - 165	3	30
Benzo[e]pyrene	0.5	0.525		µg/L		105	42 - 152	2	30
Benzo[g,h,i]perylene	0.5	0.452		µg/L		90	63 - 133	1	30
Benzo[k]fluoranthene	0.5	0.527		µg/L		105	56 - 145	4	30
Biphenyl	0.5	0.446		µg/L		89	56 - 119	1	30
Chrysene	0.5	0.434		µg/L		87	56 - 141	4	30

Eurofins Eaton Monrovia

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

**Lab Sample ID: 98866-BS2**  
**Matrix: water**  
**Analysis Batch: O-38076**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: O-38076\_P**

Analyte	Spike Added	LCS DUP Result	LCS DUP Qualifier	Unit	D	%Rec	%Rec		RPD	Limit
							Limits	RPD		
Dibenz[a,h]anthracene	0.5	0.57		µg/L		114	55 - 150	4	30	
Dibenzo[a,i]pyrene	0.5	0.536		µg/L		107	50 - 150	7	30	
Dibenzothiophene	0.5	0.446		µg/L		89	75 - 113	0	30	
Disalicylidenepropanediamine	50	39.1		µg/L		78	50 - 150	12	30	
Fluoranthene	0.5	0.472		µg/L		94	60 - 146	4	30	
Fluorene	0.5	0.435		µg/L		87	58 - 131	2	30	
Indeno[1,2,3-cd]pyrene	0.5	0.601		µg/L		120	50 - 151	3	30	
Naphthalene	0.5	0.425		µg/L		85	41 - 126	2	30	
Perylene	0.5	0.491		µg/L		98	48 - 141	3	30	
Phenanthrene	0.5	0.456		µg/L		91	67 - 127	2	30	
Pyrene	0.5	0.47		µg/L		94	54 - 156	4	30	

Surrogate	LCS DUP		Limits
	%Recovery	Qualifier	
(d10-Acenaphthene)	93		65 - 113
(d10-Phenanthrene)	96		80 - 111
(d12-Chrysene)	98		60 - 139
(d12-Perylene)	99		36 - 161
(d8-Naphthalene)	87		44 - 119

## Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

**Lab Sample ID: 22DSH009WB**  
**Matrix: WATER**  
**Analysis Batch: 22DSH009W**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB MB		RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
DIESEL	ND	U	0.025		mg/L			08/05/22 18:16	1
MOTOR OIL	ND	U	0.05		mg/L			08/05/22 18:16	1

Surrogate	MB MB		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
BROMOBENZENE					08/05/22 18:16	1
HEXACOSANE					08/05/22 18:16	1

**Lab Sample ID: 22DSH009WL**  
**Matrix: WATER**  
**Analysis Batch: 22DSH009W**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec	
							Limits	RPD
DIESEL	2.5	2.23		mg/L		89	50 - 130	

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
BROMOBENZENE	78		60 - 130
HEXACOSANE	110		60 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

**Lab Sample ID: 22VG39H01B**  
**Matrix: WATER**  
**Analysis Batch: 22VG39H01**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
GASOLINE	ND	U	0.02		mg/L			08/02/22 12:53	1
Surrogate	MB %Recovery	MB Qualifier	Limits				Prepared	Analyzed	Dil Fac
BROMOFLUOROBENZENE								08/02/22 12:53	1

**Lab Sample ID: 22VG39H01L**  
**Matrix: WATER**  
**Analysis Batch: 22VG39H01**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
GASOLINE	0.5	0.497		mg/L		99	60 - 130
Surrogate	LCS %Recovery	LCS Qualifier	Limits				
BROMOFLUOROBENZENE	113		70 - 130				

**Lab Sample ID: 22H004-01M**  
**Matrix: WATER**  
**Analysis Batch: 22VG39H01**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
GASOLINE	ND		0.5	0.458		mg/L		92	50 - 130
Surrogate	MS %Recovery	MS Qualifier	Limits						
BROMOFLUOROBENZENE	110		60 - 140						

**Lab Sample ID: 22H004-01S**  
**Matrix: WATER**  
**Analysis Batch: 22VG39H01**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
GASOLINE	ND		0.5	0.471		mg/L		94	50 - 130	3	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
BROMOFLUOROBENZENE	113		60 - 140								



# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## GC/MS Semi VOA

### Prep Batch: 10968

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	525.2	
MB 380-10968/1-A	Method Blank	Total/NA	Water	525.2	
LCS 380-10968/3-A	Lab Control Sample	Total/NA	Water	525.2	
LCSD 380-10968/4-A	Lab Control Sample Dup	Total/NA	Water	525.2	
MRL 380-10968/2-A	Lab Control Sample	Total/NA	Water	525.2	
380-12602-S-6-A MS	Matrix Spike	Total/NA	Water	525.2	
380-12317-I-1-A DU	Duplicate	Total/NA	Water	525.2	

### Analysis Batch: 14003

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	525.2	10968
MB 380-10968/1-A	Method Blank	Total/NA	Water	525.2	10968
LCS 380-10968/3-A	Lab Control Sample	Total/NA	Water	525.2	10968
LCSD 380-10968/4-A	Lab Control Sample Dup	Total/NA	Water	525.2	10968
MRL 380-10968/2-A	Lab Control Sample	Total/NA	Water	525.2	10968
380-12602-S-6-A MS	Matrix Spike	Total/NA	Water	525.2	10968
380-12317-I-1-A DU	Duplicate	Total/NA	Water	525.2	10968

## Subcontract

### Analysis Batch: O-38076

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	625 PAH Physis LL (EAL) + TICs	O-38076_P
98866-B1	Method Blank	Total/NA	water	625 PAH Physis LL (EAL) + TICs	O-38076_P
98866-BS1	Lab Control Sample	Total/NA	water	625 PAH Physis LL (EAL) + TICs	O-38076_P
98866-BS2	Lab Control Sample Dup	Total/NA	water	625 PAH Physis LL (EAL) + TICs	O-38076_P

### Analysis Batch: 22DSH009W

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	8015 Diesel LL (EAL) and Motor Oil	
22DSH009WB	Method Blank	Total/NA	WATER	8015 Diesel LL (EAL) and Motor Oil	
22DSH009WL	Lab Control Sample	Total/NA	WATER	8015 Diesel LL (EAL) and Motor Oil	

### Analysis Batch: 22VG39H01

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	8015 Gas (Purgeable) LL (EAL)	
22VG39H01B	Method Blank	Total/NA	WATER	8015 Gas (Purgeable) LL (EAL)	
22VG39H01L	Lab Control Sample	Total/NA	WATER	8015 Gas (Purgeable) LL (EAL)	

Eurofins Eaton Monrovia

# QC Association Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-13074-1

## Subcontract (Continued)

### Analysis Batch: 22VG39H01 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
22H004-01M	Matrix Spike	Total/NA	WATER	8015 Gas (Purgeable) LL (EAL)	
22H004-01S	Matrix Spike Duplicate	Total/NA	WATER	8015 Gas (Purgeable) LL (EAL)	

### Prep Batch: O-38076\_P

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	EPA_625	
98866-B1	Method Blank	Total/NA	water	EPA_625	
98866-BS1	Lab Control Sample	Total/NA	water	EPA_625	
98866-BS2	Lab Control Sample Dup	Total/NA	water	EPA_625	



# Lab Chronicle

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

**Client Sample ID: AIEA GULCH WELLS PUMP 2  
(331-202-TP072)**

**Lab Sample ID: 380-13074-1**

**Date Collected: 07/27/22 11:36**

**Matrix: Drinking Water**

**Date Received: 07/29/22 10:00**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			10968	G9MN	EA MON	08/01/22 08:52
Total/NA	Analysis	525.2		1	14003	UPAC	EA MON	08/19/22 11:00
Total/NA	Prep	EPA_625		1	O-38076_P			08/02/22 00:00
Total/NA	Analysis	625 PAH Physis LL (EAL) + TICs		1	O-38076	YC		08/07/22 06:47
Total/NA	Analysis	8015 Diesel LL (EAL) and Motor Oil		1	22DSH009W	SDees		08/05/22 19:12
Total/NA	Analysis	8015 Gas (Purgeable) LL (EAL)		1	22VG39H01	SCerva		08/02/22 15:16

**Laboratory References:**

= Physis Environmental Laboratories, 1904 Wright Circle, Anaheim, CA 92806

EA MON = Eurofins Eaton Monrovia, 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016, TEL (626)386-1100

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

## Laboratory: Eurofins Eaton Monrovia

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-23

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	2,4'-DDD
525.2	525.2	Drinking Water	2,4'-DDE
525.2	525.2	Drinking Water	2,4'-DDT
525.2	525.2	Drinking Water	2,4-Dinitrotoluene
525.2	525.2	Drinking Water	2,6-Dinitrotoluene
525.2	525.2	Drinking Water	4,4'-DDD
525.2	525.2	Drinking Water	4,4'-DDE
525.2	525.2	Drinking Water	4,4'-DDT
525.2	525.2	Drinking Water	Acenaphthene
525.2	525.2	Drinking Water	Acenaphthylene
525.2	525.2	Drinking Water	Acetochlor
525.2	525.2	Drinking Water	alpha-BHC
525.2	525.2	Drinking Water	alpha-Chlordane
525.2	525.2	Drinking Water	Anthracene
525.2	525.2	Drinking Water	Benz(a)anthracene
525.2	525.2	Drinking Water	Benzo[b]fluoranthene
525.2	525.2	Drinking Water	Benzo[g,h,i]perylene
525.2	525.2	Drinking Water	Benzo[k]fluoranthene
525.2	525.2	Drinking Water	beta-BHC
525.2	525.2	Drinking Water	Bromacil
525.2	525.2	Drinking Water	Butylbenzylphthalate
525.2	525.2	Drinking Water	Caffeine
525.2	525.2	Drinking Water	Chlorobenzilate
525.2	525.2	Drinking Water	Chloroneb
525.2	525.2	Drinking Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Drinking Water	Chlorpyrifos
525.2	525.2	Drinking Water	Chrysene
525.2	525.2	Drinking Water	delta-BHC
525.2	525.2	Drinking Water	Diazinon (Qualitative)
525.2	525.2	Drinking Water	Dibenz(a,h)anthracene
525.2	525.2	Drinking Water	Diclorvos (DDVP)
525.2	525.2	Drinking Water	Diethylphthalate
525.2	525.2	Drinking Water	Dimethoate
525.2	525.2	Drinking Water	Dimethylphthalate
525.2	525.2	Drinking Water	Di-n-butyl phthalate
525.2	525.2	Drinking Water	Di-n-octyl phthalate
525.2	525.2	Drinking Water	Endosulfan I (Alpha)
525.2	525.2	Drinking Water	Endosulfan II (Beta)
525.2	525.2	Drinking Water	Endosulfan sulfate
525.2	525.2	Drinking Water	Endrin aldehyde
525.2	525.2	Drinking Water	EPTC
525.2	525.2	Drinking Water	Fluoranthene
525.2	525.2	Drinking Water	Fluorene
525.2	525.2	Drinking Water	gamma-Chlordane
525.2	525.2	Drinking Water	Indeno[1,2,3-cd]pyrene

# Accreditation/Certification Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-13074-1

## Laboratory: Eurofins Eaton Monrovia (Continued)

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
-----------	---------	-----------------------	-----------------

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	Isophorone
525.2	525.2	Drinking Water	Malathion
525.2	525.2	Drinking Water	Molinate
525.2	525.2	Drinking Water	Naphthalene
525.2	525.2	Drinking Water	Parathion
525.2	525.2	Drinking Water	Pendimethalin (Penoxaline)
525.2	525.2	Drinking Water	Phenanthrene
525.2	525.2	Drinking Water	Pyrene
525.2	525.2	Drinking Water	Terbacil
525.2	525.2	Drinking Water	Terbutylazine
525.2	525.2	Drinking Water	Thiobencarb
525.2	525.2	Drinking Water	Total Permethrin (mixed isomers)
525.2	525.2	Drinking Water	trans-Nonachlor
525.2	525.2	Drinking Water	Trifluralin



# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA MON
625	EPA 625 Base/Neutral and Acid Organics i	EPA	
8015	8015 - TPH DRO/ORO	EPA	
8015B	SW846 8015B Gasoline Range Organics	SW846	
525.2	Extraction of Semivolatile Compounds	EPA	EA MON

#### Protocol References:

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

#### Laboratory References:

= Physis Environmental Laboratories, 1904 Wright Circle, Anaheim, CA 92806

EA MON = Eurofins Eaton Monrovia, 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016, TEL (626)386-1100

# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-13074-1

---

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-13074-1	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Drinking Water	07/27/22 11:36	07/29/22 10:00	HI0000331

---

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17



3051 Fujita Street  
Torrance, CA 90505  
Tel: (310)-618-8889

Date: 08-12-2022  
EMAX Batch No.: 22H004

Attn: Jackie Contreras

Eurofins Eaton Analytical  
750 Royal Oaks Dr., Suite 100  
Monrovia, CA 91016-3629

Subject: Laboratory Report  
Project: 380-13074


Enclosed is the Laboratory report for samples received on 08/01/22.  
The data reported relate only to samples listed below :

Sample ID	Control #	Col Date	Matrix	Analysis
380-13074-1	H004-01	07/27/22	WATER	TPH GASOLINE TPH DIESEL & MOTOR OIL
380-13074-1MS	H004-01M	07/27/22	WATER	TPH GASOLINE
380-13074-1MSD	H004-01S	07/27/22	WATER	TPH GASOLINE

The results are summarized on the following pages.

Please feel free to call if you have any questions concerning these results.

Sincerely yours,

  
Caspar J. Pang  
Laboratory Director

This report is confidential and intended solely for the use of the individual or entity to whom it is addressed. This report shall not be reproduced except in full or without the written approval of EMAX.

EMAX certifies that results included in this report meets all TNI & DOD requirements unless noted in the Case Narrative.

NELAP Accredited Certificate Number CA002912022-22  
ANAB Accredited DoD ELAP and ISO/IEC 17025 Certificate Number L2278 Testing  
California ELAP Accredited Certificate Number 2672







Type of Delivery <input type="checkbox"/> Fedex <input type="checkbox"/> UPS <input type="checkbox"/> GSO <input type="checkbox"/> Others <input type="checkbox"/> EMAX Courier <input checked="" type="checkbox"/> Client Delivery	Airbill / Tracking Number	ECN <u>22H004</u> Recipient <u>Maria Rivera</u> Date <u>08/01/22</u> Time <u>10:20</u>
---	---------------------------	--

**COC INSPECTION**

<input checked="" type="checkbox"/> Client Name	<input checked="" type="checkbox"/> Client PM/FC	<input type="checkbox"/> Sampler Name	<input checked="" type="checkbox"/> Sampling Date/Time	<input checked="" type="checkbox"/> Sample ID	<input checked="" type="checkbox"/> Matrix
<input checked="" type="checkbox"/> Address	<input type="checkbox"/> Tel # / Fax #	<input type="checkbox"/> Courier Signature	<input checked="" type="checkbox"/> Analysis Required	<input type="checkbox"/> Preservative (if any)	<input checked="" type="checkbox"/> TAT
Safety Issues (if any)	<input type="checkbox"/> High concentrations expected	<input type="checkbox"/> From Superfund Site	<input type="checkbox"/> Rad screening required		

Note: \_\_\_\_\_

**PACKAGING INSPECTION**

Container	<input checked="" type="checkbox"/> Cooler	<input type="checkbox"/> Box	<input type="checkbox"/> Other
Condition	<input type="checkbox"/> Custody Seal	<input type="checkbox"/> Intact	<input type="checkbox"/> Damaged
Packaging	<input checked="" type="checkbox"/> Bubble Pack	<input type="checkbox"/> Styrofoam	<input type="checkbox"/> Popcorn
Temperatures (Cool, ≤6 °C but not frozen)	<input checked="" type="checkbox"/> Cooler 1 <u>2.1</u> °C	<input type="checkbox"/> Cooler 2 _____ °C	<input type="checkbox"/> Cooler 3 _____ °C
	<input type="checkbox"/> Cooler 6 _____ °C	<input type="checkbox"/> Cooler 7 _____ °C	<input type="checkbox"/> Cooler 8 _____ °C
Thermometer:	A - S/N <u>210583479</u>	B - S/N _____	C - S/N <u>210271399</u>
			D - S/N <u>210760272</u>

Comments:  Temperature is out of range. PM was informed IMMEDIATELY.  
 Note: \_\_\_\_\_

**DISCREPANCIES**

LabSampleID	LabSampleContainerID	Code	ClientSample Label ID / Information	Corrective Action
<i>[Large diagonal scribble across the table]</i>				

pH holding time requirement for water samples is 15 mins. Water samples for pH analysis are received beyond 15 minutes from sampling time. *MB 8/5/22*

NOTES/OBSERVATIONS:  
 SAMPLE MATRIX IS DRINKING WATER?  YES  NO

- LEGEND:**
- |   |   |   |
|---|---|---|
| <ul style="list-style-type: none"> <li>D1 Analysis is not indicated in _____</li> <li>D2 Analysis mismatch COC vs label</li> <li>D3 Sample ID mismatch COC vs label</li> <li>D4 Sample ID is not indicated in _____</li> <li>D5 Container -[inproper] [leaking] [broken]</li> <li>D6 Date/Time is not indicated in _____</li> <li>D7 Date/Time mismatch COC vs label</li> <li>D8 Sample listed in COC is not received</li> <li>D9 Sample received is not listed in COC</li> <li>D10 No initial/date on corrections in COC/label</li> <li>D11 Container count mismatch COC vs received</li> <li>D12 Container size mismatch COC vs received</li> </ul> | <ul style="list-style-type: none"> <li>D13 Out of Holding Time</li> <li>D14 Bubble is &gt;6mm</li> <li>D15 No trip blank in cooler</li> <li>D16 Preservation not indicated in _____</li> <li>D17 Preservation mismatch COC vs label</li> <li>D18 Insufficient chemical preservative</li> <li>D19 Insufficient Sample</li> <li>D20 No filtration info for dissolved analysis</li> <li>D21 No sample for moisture determination</li> <li>D22 _____</li> <li>D23 _____</li> <li>D24 _____</li> </ul> | <ul style="list-style-type: none"> <li><input type="checkbox"/> Continue to next page.</li> <li>R1 Proceed as indicated in <input type="checkbox"/> COC <input type="checkbox"/> Label</li> <li>R2 Refer to attached instruction</li> <li>R3 Cancel the analysis</li> <li>R4 Use vial with smallest bubble first</li> <li>R5 Log-in with latest sampling date and time+1 min</li> <li>R6 Adjust pH as necessary</li> <li>R7 Filter and preserved as necessary</li> <li>R8 _____</li> <li>R9 _____</li> <li>R10 _____</li> <li>R11 _____</li> <li>R12 _____</li> </ul> |
|---|---|---|

REVIEWS:

Sample Labeling <i>Jocelyne</i>	SRF <i>[Signature]</i>	PM <i>MB</i>
Date <u>08/02/22</u>	Date <u>8/2/22</u>	Date <u>8/5/22</u>

## REPORTING CONVENTIONS

### DATA QUALIFIERS:

Lab Qualifier	AFCEE Qualifier	Description
J	F	Indicates that the analyte is positively identified and the result is less than RL but greater than MDL.
N		Indicates presumptive evidence of a compound.
B	B	Indicates that the analyte is found in the associated method blank as well as in the sample at above QC level.
E	J	Indicates that the result is above the maximum calibration range or estimated value.
*	*	Out of QC limit.

**Note:** The above qualifiers are used to flag the results unless the project requires a different set of qualification criteria.

### ACRONYMS AND ABBREVIATIONS:

CRDL	Contract Required Detection Limit
RL	Reporting Limit
MRL	Method Reporting Limit
PQL	Practical Quantitation Limit
MDL	Method Detection Limit
DO	Diluted out

### DATES

The date and time information for leaching and preparation reflect the beginning date and time of the procedure unless the method, protocol, or project specifically requires otherwise.

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

380-13074

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

SDG#: 22H004



CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 380-13074

SDG : 22H004

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

One(1) water sample was received on 08/01/22 to be analyzed for Total Petroleum Hydrocarbons by Purge and Trap in accordance with Method 5030B/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. VG39H01B - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. VG39H01L/VG39H01C were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Gasoline was within MS QC limits in H004-01M/H004-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogate was added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL
Project     : 380-13074
=====
SDG NO.    : 22H004
Instrument ID : GCT039
=====

```

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
									WATER
MBLK1W	VG39H01B	1	NA	08/02/2212:53	08/02/2212:53	EH02005A	EH02003A	22VG39H01	Method Blank
LCS1W	VG39H01L	1	NA	08/02/2213:29	08/02/2213:29	EH02006A	EH02003A	22VG39H01	Lab Control Sample (LCS)
LCD1W	VG39H01C	1	NA	08/02/2214:04	08/02/2214:04	EH02007A	EH02003A	22VG39H01	LCS Duplicate
380-13074-1	H004-01	1	NA	08/02/2215:16	08/02/2215:16	EH02009A	EH02003A	22VG39H01	Field Sample
380-13074-1MS	H004-01M	1	NA	08/02/2215:53	08/02/2215:53	EH02010A	EH02003A	22VG39H01	Matrix Spike Sample (MS)
380-13074-1MSD	H004-01S	1	NA	08/02/2216:29	08/02/2216:29	EH02011A	EH02003A	22VG39H01	MS Duplicate (MSD)

FN - Filename  
% Moist - Percent Moisture



1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# SAMPLE RESULTS



METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 07/27/22 11:36
Project     : 380-13074                   Date Received: 08/01/22
Batch No.   : 22H004                       Date Extracted: 08/02/22 15:16
Sample ID   : 380-13074-1                 Date Analyzed: 08/02/22 15:16
Lab Samp ID: H004-01                       Dilution Factor: 1
Lab File ID: EH02009A                       Matrix: WATER
Ext Btch ID: 22VG39H01                     % Moisture: NA
Calib. Ref.: EH02003A                       Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
GASOLINE	ND	0.020	0.010	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0361	0.0400	90	60-140

Notes:

Parameter H-C Range  
Gasoline C6-C10  
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
Sample Amount : 5ml Final Volume : 5ml  
Prepared by : SCerva Analyzed by : SCerva



1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# QC SUMMARIES

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 08/02/22 12:53
Project     : 380-13074                   Date Received: 08/02/22
Batch No.   : 22H004                       Date Extracted: 08/02/22 12:53
Sample ID   : MBLK1W                       Date Analyzed: 08/02/22 12:53
Lab Samp ID: VG39H01B                      Dilution Factor: 1
Lab File ID: EH02005A                      Matrix: WATER
Ext Btch ID: 22VG39H01                     % Moisture: NA
Calib. Ref.: EH02003A                     Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
GASOLINE	ND	0.020	0.010

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0354	0.0400	88	60-140

Notes:

Parameter H-C Range  
Gasoline C6-C10  
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
Sample Amount : 5ml Final Volume : 5ml  
Prepared by : SCerva Analyzed by : SCerva

EMAX QUALITY CONTROL DATA  
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 380-13074  
BATCH NO. : 22H004  
METHOD : 5030B/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : MBLK1W                             LCS1W         LCD1W
LAB SAMPLE ID : VG39H01B                         VG39H01L     VG39H01C
LAB FILE ID  : EH02005A                         EH02006A     EH02007A
DATE PREPARED : 08/02/22 12:53                 08/02/22 13:29 08/02/22 14:04
DATE ANALYZED : 08/02/22 12:53                 08/02/22 13:29 08/02/22 14:04
PREP BATCH   : 22VG39H01                       22VG39H01    22VG39H01
CALIBRATION REF: EH02003A                       EH02003A     EH02003A
    
```

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.497	99	0.500	0.477	95	4	60-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0451	113	0.0400	0.0453	113	70-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA  
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 380-13074  
BATCH NO. : 22H004  
METHOD : 5030B/8015B

MATRIX : WATER		% MOISTURE:NA
DILUTION FACTOR: 1	1	1
SAMPLE ID : 380-13074-1	380-13074-1MS	380-13074-1MSD
LAB SAMPLE ID : H004-01	H004-01M	H004-01S
LAB FILE ID : EH02009A	EH02010A	EH02011A
DATE PREPARED : 08/02/22 15:16	08/02/22 15:53	08/02/22 16:29
DATE ANALYZED : 08/02/22 15:16	08/02/22 15:53	08/02/22 16:29
PREP BATCH : 22VG39H01	22VG39H01	22VG39H01
CALIBRATION REF: EH02003A	EH02003A	EH02003A

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.458	92	0.500	0.471	94	3	50-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0441	110	0.0400	0.0453	113	60-140

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

380-13074

METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG#: 22H004



CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 380-13074

SDG : 22H004

METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 08/01/22 to be analyzed for Total Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSH009WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. DSH009WL/DSH009WC were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Diesel was within MS QC limits in 22H059-01M/22H059-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

Client : EUROFINS EATON ANALYTICAL  
 Project : 380-13074  
 SDG NO. : 22H004  
 Instrument ID : D5

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
									WATER
	DSH009WB	1	NA	08/05/2218:16	08/04/2214:45	LH05021A	LH05017A	22DSH009W	Method Blank
	LCS1W	1	NA	08/05/2218:35	08/04/2214:45	LH05022A	LH05017A	22DSH009W	Lab Control Sample (LCS)
	LCD1W	1	NA	08/05/2218:53	08/04/2214:45	LH05023A	LH05017A	22DSH009W	LCS Duplicate
380-13074-1	H004-01	1	NA	08/05/2219:12	08/04/2214:45	LH05024A	LH05017A	22DSH009W	Field Sample

FN - Filename  
 % Moist - Percent Moisture



1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# SAMPLE RESULTS



METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 07/27/22 11:36
Project     : 380-13074                 Date Received: 08/01/22
Batch No.   : 22H004                    Date Extracted: 08/04/22 14:45
Sample ID   : 380-13074-1              Date Analyzed: 08/05/22 19:12
Lab Samp ID: 22H004-01                 Dilution Factor: 1
Lab File ID: LH05024A                  Matrix: WATER
Ext Btch ID: 22DSH009W                 % Moisture: NA
Calib. Ref.: LH05017A                  Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.028	0.014	
Motor Oil	ND	0.056	0.028	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.447	0.555	81	60-130
Hexacosane	0.136	0.139	98	60-130

Notes:

Parameter H-C Range  
Diesel C10-C24  
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 900ml Final Volume : 5ml  
Prepared by : POrto Analyzed by : SDeeso

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# QC SUMMARIES

METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 08/04/22 14:45
Project    : 380-13074                   Date Received: 08/04/22
Batch No.  : 22H004                       Date Extracted: 08/04/22 14:45
Sample ID  : MBLK1W                       Date Analyzed: 08/05/22 18:16
Lab Samp ID: DSH009WB                     Dilution Factor: 1
Lab File ID: LH05021A                     Matrix: WATER
Ext Btch ID: 22DSH009W                    % Moisture: NA
Calib. Ref.: LH05017A                    Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.025	0.012	
Motor Oil	ND	0.050	0.025	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.346	0.500	69	60-130
Hexacosane	0.132	0.125	106	60-130

Notes:

Parameter      H-C Range  
Diesel            C10-C24  
Motor Oil        C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml                      Final Volume : 5ml  
Prepared by    : P0reto                         Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA  
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 380-13074  
BATCH NO. : 22H004  
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : MBLK1W                             LCS1W
LAB SAMPLE ID : DSH009WB                         DSH009WL
LAB FILE ID  : LH05021A                         LH05022A
DATE PREPARED : 08/04/22 14:45                 08/04/22 14:45
DATE ANALYZED : 08/05/22 18:16                 08/05/22 18:35
PREP BATCH   : 22DSH009W                       22DSH009W
CALIBRATION REF: LH05017A                      LH05017A
    
```

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Diesel	ND	2.50	2.23	89	2.50	2.24	90	0	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromobenzene	0.500	0.389	78	0.500	0.351	70	60-130
Hexacosane	0.125	0.137	110	0.125	0.141	113	60-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA  
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 380-14125  
BATCH NO. : 22H059  
METHOD : 3520C/8015B

```

=====
MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : 380-14125-1 380-14125-1MS 380-14125-1MSD
LAB SAMPLE ID : 22H059-01 22H059-01M 22H059-01s
LAB FILE ID : LH05025A LH05026A LH05027A
DATE PREPARED : 08/04/22 14:45 08/04/22 14:45 08/04/22 14:45
DATE ANALYZED : 08/05/22 19:31 08/05/22 19:49 08/05/22 20:08
PREP BATCH : 22DSH009W 22DSH009W 22DSH009W
CALIBRATION REF: LH05017A LH05017A LH05017A
  
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Diesel	ND	2.72	2.59	95	2.70	2.35	87	10	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.545	0.413	76	0.540	0.377	70	60-130
Hexacosane	0.136	0.152	112	0.135	0.142	105	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

August 09, 2022

Debbie Frank  
Eurofins Eaton Analytical  
750 Royal Oaks Drive  
Suite 100  
Monrovia, CA 91016-

Project Name: RED\_HILL Project # 38001111 Job # 380-13074-1  
Physis Project ID: 1407003-262

Dear Debbie,


Enclosed are the analytical results for the sample submitted to PHYSIS Environmental Laboratories, Inc. (PHYSIS) on 8/1/2022. A total of 1 sample was received for analysis in accordance with the attached chain of custody (COC). Per the COC, the sample was analyzed for:

Organics
Polynuclear Aromatic Hydrocarbons by EPA 625.1
Disalicylidenepropanediamine by EPA 625.1
Dibenzo [a,l] Pyrene w/ PAHs by EPA 625.1

Analytical results in this report apply only to samples submitted to PHYSIS in accordance with the COC and are intended to be considered in their entirety.

Please feel free to contact me at any time with any questions. PHYSIS appreciates the opportunity to provide you with our analytical and support services.

Regards,

  
Misty Mercier  
714 602-5320  
Extension 202  
mistymercier@physislabs.com

## PROJECT SAMPLE LIST

Eurofins Eaton Analytical

PHYSIS Project ID: 1407003-262

RED\_HILL Project # 38001111 Job # 380-13074-1

Total Samples: 1

PHYSIS ID	Sample ID	Description	Date	Time	Matrix	Sample Type
98867	AIEA GULCH WELLS PUMP	331-202-TP072 (380-13074-1)	7/27/2022	11:36	Samplewater	Not Specified

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

## ABBREVIATIONS and ACRONYMS

QM	Quality Manual
QA	Quality Assurance
QC	Quality Control
MDL	method detection limit
RL	reporting limit
R1	project sample
R2	project sample replicate
MS1	matrix spike
MS2	matrix spike replicate
B1	procedural blank
B2	procedural blank replicate
BS1	blank spike
BS2	blank spike replicate
LCS1	laboratory control spike
LCS2	laboratory control spike replicate
LCM1	laboratory control material
LCM2	laboratory control material replicate
CRM1	certified reference material
CRM2	certified reference material replicate
RPD	relative percent difference
LMW	low molecular weight
HMW	high molecular weight



## QUALITY ASSURANCE SUMMARY

**LABORATORY BATCH:** Physis' QM defines a laboratory batch as a group of 20 or fewer project samples of similar matrix, processed together under the same conditions and with the same reagents. QC samples are associated with each batch and were used to assess the validity of the sample analyses.

**PROCEDURAL BLANK:** Laboratory contamination introduced during method use is assessed through the preparation and analysis of procedural blanks is provided at a minimum frequency of one per batch.

**ACCURACY:** Accuracy of analytical measurements is the degree of closeness based on percent recovery calculations between measured values and the actual or true value and includes a combination of reproducibility error and systematic bias due to sampling and analytical operations. Accuracy of the project data was indicated by analysis of MS, BS, LCS, LCM, CRM, and/or surrogate spikes on a minimum frequency of one per batch. Physis' QM requires that 95% of the target compounds greater than 10 times the MDL be within the specified acceptance limits.

**PRECISION:** Precision is the agreement among a set of replicate measurements without assumption of knowledge of the true value and is based on RPD calculations between repeated values. Precision of the project data was determined by analysis of replicate MS<sub>1</sub>/MS<sub>2</sub>, BS<sub>1</sub>/BS<sub>2</sub>, LCS<sub>1</sub>/LCS<sub>2</sub>, LCM<sub>1</sub>/LCM<sub>2</sub>, CRM<sub>1</sub>/CRM<sub>2</sub>, surrogate spikes and/or replicate project sample analysis (R<sub>1</sub>/R<sub>2</sub>) on a minimum frequency of one per batch. Physis' QM requires that for 95% of the compounds greater than 10 times the MDL, the percent RPD should be within the specified acceptance range.

**BLANK SPIKES:** BS is the introduction of a known concentration of analyte into the procedural blank. BS demonstrates performance of the preparation and analytical methods on a clean matrix void of potential matrix related interferences. The BS is performed in laboratory deionized water, making these recoveries a better indicator of the efficiency of the laboratory method per se.

**MATRIX SPIKES:** MS is the introduction of a known concentration of analyte into a sample. MS samples demonstrate the effect a particular project sample matrix has on the accuracy of a measurement. Individually, MS samples also indicate the bias of analytical measurements due to chemical interferences inherent in the in the specific project sample spiked. Intrinsic target analyte concentration in the specific project sample can also significantly impact MS recovery.

**CERTIFIED REFERENCE MATERIALS:** CRMs are materials of various matrices for which analytical information has been determined and certified by a recognized authority. These are used to provide a quantitative assessment of the accuracy of an analytical method. CRMs provide evidence that the laboratory preparation and analysis produces results that are comparable to those obtained by an independent organization.

**LABORATORY CONTROL MATERIAL:** LCM is provided because a suitable natural seawater CRM is not available and can be used to indicate accuracy of the method. Physis' internal LCM is seawater collected at ~800 meters in the Southern California San Pedro Basin and can be used as a reference for background concentrations in clean, natural seawater for comparison to project samples.

**LABORATORY CONTROL SPIKES:** LCS is the introduction of a known concentration of analyte into Physis' LCM. LCS samples were employed to assess the effect the seawater matrix has on the accuracy of a measurement. LCS also indicate the bias of this method due to chemical interferences inherent in the in the seawater matrix. Intrinsic LCM concentration can also significantly impact LCS recovery.

**SURROGATES:** A surrogate is a pure analyte unlikely to be found in any project sample, behaves similarly to

the target analyte and most often used with organic analytical procedures. Surrogates are added in known concentration to all samples and are measured to indicate overall efficiency of the method including processing and analyses.

**HOLDING TIME:** Method recommended holding times are the length of time a project sample can be stored under specific conditions after collection and prior to analysis without significantly affecting the analyte's concentration. Holding times can be extended if preservation techniques are employed to reduce biodegradation, volatilization, oxidation, sorption, precipitation, and other physical and chemical processes.

**SAMPLE STORAGE/RETENTION:** In order to maintain chemical integrity prior to analysis, all samples submitted to Physis are refrigerated (liquids) or frozen (solids) upon receipt unless otherwise recommended by applicable methods. Solid samples are retained for 1 year from collection while liquid samples are retained until method recommended holding times elapse.

**TOTAL/DISSOLVED FRACTION:** In some instances, the results for the dissolved fraction may be higher than the total fraction for a particular analyte (e.g. trace metals). This is typically caused by the analytical variation for each result and indicates that the target analyte is primarily in the dissolved phase, within the sample.

## PHYSIS QUALIFIER CODES

CODE	DEFINITION
#	see Case Narrative
ND	analyte not detected at or above the MDL
B	analyte was detected in the procedural blank greater than 10 times the MDL
E	analyte concentration exceeds the upper limit of the linear calibration range, reported value is estimated
H	sample received and/or analyzed past the recommended holding time
J	analyte was detected at a concentration below the RL and above the MDL, reported value is estimated
N	insufficient sample, analysis could not be performed
M	analyte was outside the specified accuracy and/or precision acceptance limits due to matrix interference. The associated B/BS were within limits, therefore the sample data was reported without further clarification
SH	analyte concentration in the project sample exceeded the spike concentration, therefore accuracy and/or precision acceptance limits do not apply
SL	analyte results were lower than 10 times the MDL, therefore accuracy and/or precision acceptance limits do not apply
NH	project sample was heterogeneous and sample homogeneity could not be readily achieved using routine laboratory practices, therefore accuracy and/or precision acceptance limits do not apply
Q	analyte was outside the specified QAPP acceptance limits for precision and/or accuracy but within Physis derived acceptance limits, therefore the sample data was reported without further clarification
R	Physis' QM allows for 5% of the target compounds greater than 10 times the MDL to be outside the specified acceptance limits for precision and/or accuracy. This is often due to random error and does not indicate any significant problems with the analysis of these project samples

---

## CASE NARRATIVE

### QUALIFIER NOTES

In addition to the use of analyte specific Physis Qualifier Codes where applicable, the following were also noted.

#### **ND**

MDL is listed due to report format restrictions; it is not used in reporting. Analytical results reported are ND at the RL.

# BIANALYTICALS

## REPORT

TERRA AURA  
ENVIRONMENTAL LABORATORIES, INC.

*Innovative Solutions for Nature*

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

## Base/Neutral Extractable Compounds

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
<b>Sample ID: 98867-R1</b>	<b>AIEA GULCH WELLS PUMP 2 331-20 Matrix: Samplewater</b>						<b>Sampled:</b>	<b>27-Jul-22 11:36</b>	<b>Received:</b>	<b>01-Aug-22</b>	
Disalicylidenepranediamine	EPA 625.1	µg/L	ND	1	0.05	0.1	Total		O-38076	02-Aug-22	07-Aug-22



## Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed	
<b>Sample ID: 98867-R1</b>	<b>AIEA GULCH WELLS PUMP 2 331-20 Matrix: Samplewater</b>						<b>Sampled:</b>	<b>27-Jul-22 11:36</b>	<b>Received:</b>	<b>01-Aug-22</b>		
(d10-Acenaphthene)	EPA 625.1	% Recovery	87	1			Total		O-38076	02-Aug-22	07-Aug-22	
(d10-Phenanthrene)	EPA 625.1	% Recovery	90	1			Total		O-38076	02-Aug-22	07-Aug-22	
(d12-Chrysene)	EPA 625.1	% Recovery	84	1			Total		O-38076	02-Aug-22	07-Aug-22	
(d12-Perylene)	EPA 625.1	% Recovery	87	1			Total		O-38076	02-Aug-22	07-Aug-22	
(d8-Naphthalene)	EPA 625.1	% Recovery	81	1			Total		O-38076	02-Aug-22	07-Aug-22	
1-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
1-Methylphenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
2,3,5-Trimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
2,6-Dimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
2-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Acenaphthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Acenaphthylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Benz[a]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Benzo[a]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Benzo[b]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Benzo[e]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Benzo[g,h,i]perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Benzo[k]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Biphenyl	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
Chrysene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
D benz[a,h]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
D benzo[a,l]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	
D benzothiophene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22	

### Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22
Fluorene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22
Indeno[1,2,3-cd]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22
Naphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22
Perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22
Phenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22
Pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38076	02-Aug-22	07-Aug-22





# QUALITY CONTROL REPORT

TERRA ENVIRONMENTAL LABORATORIES, INC. AURA

*Innovative Solutions for Nature*

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

## Base/Neutral Extractable Compounds

## QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE SOURCE		ACCURACY		PRECISION		QA CODEc
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
<b>Sample ID: 98866-B1</b>		<b>QAQC Procedural Blank</b>			<b>Matrix: BlankMatrix</b>		<b>Sampled:</b>		<b>Received:</b>				
		Method: EPA 625.1			Batch ID: O-38076		Prepared: 02-Aug-22		Analyzed: 06-Aug-22				
Disalicylideneprapanediamin	Total	ND	1	0.05	0.1	µg/L							
<b>Sample ID: 98866-BS1</b>		<b>QAQC Procedural Blank</b>			<b>Matrix: BlankMatrix</b>		<b>Sampled:</b>		<b>Received:</b>				
		Method: EPA 625.1			Batch ID: O-38076		Prepared: 02-Aug-22		Analyzed: 06-Aug-22				
Disalicylideneprapanediamin	Total	34.7	1	0.05	0.1	µg/L	50	0	69	50 - 150%	PASS		
<b>Sample ID: 98866-BS2</b>		<b>QAQC Procedural Blank</b>			<b>Matrix: BlankMatrix</b>		<b>Sampled:</b>		<b>Received:</b>				
		Method: EPA 625.1			Batch ID: O-38076		Prepared: 02-Aug-22		Analyzed: 06-Aug-22				
Disalicylideneprapanediamin	Total	39.1	1	0.05	0.1	µg/L	50	0	78	50 - 150%	PASS	12	30 PASS

**Polynuclear Aromatic Hydrocarbons**

**QUALITY CONTROL REPORT**

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY	PRECISION	QA CODEc
							LEVEL	RESULT	% LIMITS	% LIMITS	
<b>Sample ID: 98866-B1</b>		<b>QAQC Procedural Blank</b>			<b>Matrix: BlankMatrix</b>		<b>Sampled:</b>		<b>Received:</b>		
	Method: EPA 625.1					Batch ID: O-38076	Prepared: 02-Aug-22	Analyzed: 06-Aug-22			
(d10-Acenaphthene)	Total	95	1			% Recovery	100	95	65 - 113%	PASS	
(d10-Phenanthrene)	Total	98	1			% Recovery	100	98	80 - 111%	PASS	
(d12-Chrysene)	Total	95	1			% Recovery	100	95	60 - 139%	PASS	
(d12-Perylene)	Total	93	1			% Recovery	100	93	36 - 161%	PASS	
(d8-Naphthalene)	Total	91	1			% Recovery	100	91	44 - 119%	PASS	
1-Methylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
1-Methylphenanthrene	Total	ND	1	0.001	0.005	µg/L					
2,3,5-Trimethylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
2,6-Dimethylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
2-Methylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
Acenaphthene	Total	ND	1	0.001	0.005	µg/L					
Acenaphthylene	Total	ND	1	0.001	0.005	µg/L					
Anthracene	Total	ND	1	0.001	0.005	µg/L					
Benz[a]anthracene	Total	ND	1	0.001	0.005	µg/L					
Benzo[a]pyrene	Total	ND	1	0.001	0.005	µg/L					
Benzo[b]fluoranthene	Total	ND	1	0.001	0.005	µg/L					
Benzo[e]pyrene	Total	ND	1	0.001	0.005	µg/L					
Benzo[g,h,i]perylene	Total	ND	1	0.001	0.005	µg/L					
Benzo[k]fluoranthene	Total	ND	1	0.001	0.005	µg/L					
Biphenyl	Total	ND	1	0.001	0.005	µg/L					
Chrysene	Total	ND	1	0.001	0.005	µg/L					
Dibenz[a,h]anthracene	Total	ND	1	0.001	0.005	µg/L					
Dibenzo[a,l]pyrene	Total	ND	1	0.001	0.005	µg/L					

## Polynuclear Aromatic Hydrocarbons

## QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Dibenzothiophene	Total	ND	1	0.001	0.005	µg/L							
Fluoranthene	Total	ND	1	0.001	0.005	µg/L							
Fluorene	Total	ND	1	0.001	0.005	µg/L							
Indeno[1,2,3-cd]pyrene	Total	ND	1	0.001	0.005	µg/L							
Naphthalene	Total	ND	1	0.001	0.005	µg/L							
Perylene	Total	ND	1	0.001	0.005	µg/L							
Phenanthrene	Total	ND	1	0.001	0.005	µg/L							
Pyrene	Total	ND	1	0.001	0.005	µg/L							



## Polynuclear Aromatic Hydrocarbons

## QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION	QA CODE
							LEVEL	RESULT	%	LIMITS	%	LIMITS
<b>Sample ID: 98866-BS1</b>		<b>QAQC Procedural Blank</b>			<b>Matrix: BlankMatrix</b>			<b>Sampled:</b>		<b>Received:</b>		
Method: EPA 625.1		Batch ID: O-38076			Prepared: 02-Aug-22		Analyzed: 06-Aug-22					
(d10-Acenaphthene)	Total	92	1			% Recovery	100	0	92	65 - 113%	PASS	
(d10-Phenanthrene)	Total	95	1			% Recovery	100	0	95	80 - 111%	PASS	
(d12-Chrysene)	Total	95	1			% Recovery	100	0	95	60 - 139%	PASS	
(d12-Perylene)	Total	97	1			% Recovery	100	0	97	36 - 161%	PASS	
(d8-Naphthalene)	Total	86	1			% Recovery	100	0	86	44 - 119%	PASS	
1-Methylnaphthalene	Total	0.438	1	0.001	0.005	µg/L	0.5	0	88	49 - 117%	PASS	
1-Methylphenanthrene	Total	0.437	1	0.001	0.005	µg/L	0.5	0	87	66 - 127%	PASS	
2,3,5-Trimethylnaphthalene	Total	0.426	1	0.001	0.005	µg/L	0.5	0	85	57 - 120%	PASS	
2,6-Dimethylnaphthalene	Total	0.44	1	0.001	0.005	µg/L	0.5	0	88	54 - 117%	PASS	
2-Methylnaphthalene	Total	0.452	1	0.001	0.005	µg/L	0.5	0	90	47 - 130%	PASS	
Acenaphthene	Total	0.433	1	0.001	0.005	µg/L	0.5	0	87	53 - 131%	PASS	
Acenaphthylene	Total	0.432	1	0.001	0.005	µg/L	0.5	0	86	43 - 140%	PASS	
Anthracene	Total	0.442	1	0.001	0.005	µg/L	0.5	0	88	58 - 135%	PASS	
Benz[a]anthracene	Total	0.491	1	0.001	0.005	µg/L	0.5	0	98	55 - 145%	PASS	
Benzo[a]pyrene	Total	0.397	1	0.001	0.005	µg/L	0.5	0	79	51 - 143%	PASS	
Benzo[b]fluoranthene	Total	0.579	1	0.001	0.005	µg/L	0.5	0	116	46 - 165%	PASS	
Benzo[e]pyrene	Total	0.516	1	0.001	0.005	µg/L	0.5	0	103	42 - 152%	PASS	
Benzo[g,h,i]perylene	Total	0.445	1	0.001	0.005	µg/L	0.5	0	89	63 - 133%	PASS	
Benzo[k]fluoranthene	Total	0.503	1	0.001	0.005	µg/L	0.5	0	101	56 - 145%	PASS	
Biphenyl	Total	0.438	1	0.001	0.005	µg/L	0.5	0	88	56 - 119%	PASS	
Chrysene	Total	0.418	1	0.001	0.005	µg/L	0.5	0	84	56 - 141%	PASS	
Dibenz[a,h]anthracene	Total	0.552	1	0.001	0.005	µg/L	0.5	0	110	55 - 150%	PASS	
Dibenzo[a,l]pyrene	Total	0.502	1	0.001	0.005	µg/L	0.5	0	100	50 - 150%	PASS	

**Polynuclear Aromatic Hydrocarbons**

**QUALITY CONTROL REPORT**

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODE <sub>c</sub>
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Dibenzothiophene	Total	0.446	1	0.001	0.005	µg/L	0.5	0	89	75 - 113%	PASS		
Fluoranthene	Total	0.449	1	0.001	0.005	µg/L	0.5	0	90	60 - 146%	PASS		
Fluorene	Total	0.427	1	0.001	0.005	µg/L	0.5	0	85	58 - 131%	PASS		
Indeno[1,2,3-cd]pyrene	Total	0.581	1	0.001	0.005	µg/L	0.5	0	116	50 - 151%	PASS		
Naphthalene	Total	0.417	1	0.001	0.005	µg/L	0.5	0	83	41 - 126%	PASS		
Perylene	Total	0.476	1	0.001	0.005	µg/L	0.5	0	95	48 - 141%	PASS		
Phenanthrene	Total	0.446	1	0.001	0.005	µg/L	0.5	0	89	67 - 127%	PASS		
Pyrene	Total	0.451	1	0.001	0.005	µg/L	0.5	0	90	54 - 156%	PASS		



**Polynuclear Aromatic Hydrocarbons**

**QUALITY CONTROL REPORT**

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc		
							LEVEL	RESULT	%	LIMITS	%	LIMITS			
<b>Sample ID: 98866-BS2</b>		<b>QAQC Procedural Blank</b>			<b>Matrix: BlankMatrix</b>			<b>Sampled:</b>			<b>Received:</b>				
Method: EPA 625.1		Batch ID: O-38076			Prepared: 02-Aug-22			Analyzed: 06-Aug-22							
(d10-Acenaphthene)	Total	93	1				% Recovery	100	0	93	65 - 113%	PASS	1	30	PASS
(d10-Phenanthrene)	Total	96	1				% Recovery	100	0	96	80 - 111%	PASS	1	30	PASS
(d12-Chrysene)	Total	98	1				% Recovery	100	0	98	60 - 139%	PASS	3	30	PASS
(d12-Perylene)	Total	99	1				% Recovery	100	0	99	36 - 161%	PASS	2	30	PASS
(d8-Naphthalene)	Total	87	1				% Recovery	100	0	87	44 - 119%	PASS	1	30	PASS
1-Methylnaphthalene	Total	0.437	1	0.001	0.005	µg/L		0.5	0	87	49 - 117%	PASS	1	30	PASS
1-Methylphenanthrene	Total	0.455	1	0.001	0.005	µg/L		0.5	0	91	66 - 127%	PASS	4	30	PASS
2,3,5-Trimethylnaphthalene	Total	0.435	1	0.001	0.005	µg/L		0.5	0	87	57 - 120%	PASS	2	30	PASS
2,6-Dimethylnaphthalene	Total	0.444	1	0.001	0.005	µg/L		0.5	0	89	54 - 117%	PASS	1	30	PASS
2-Methylnaphthalene	Total	0.45	1	0.001	0.005	µg/L		0.5	0	90	47 - 130%	PASS	0	30	PASS
Acenaphthene	Total	0.436	1	0.001	0.005	µg/L		0.5	0	87	53 - 131%	PASS	0	30	PASS
Acenaphthylene	Total	0.437	1	0.001	0.005	µg/L		0.5	0	87	43 - 140%	PASS	1	30	PASS
Anthracene	Total	0.45	1	0.001	0.005	µg/L		0.5	0	90	58 - 135%	PASS	2	30	PASS
Benz[a]anthracene	Total	0.523	1	0.001	0.005	µg/L		0.5	0	105	55 - 145%	PASS	7	30	PASS
Benzo[a]pyrene	Total	0.405	1	0.001	0.005	µg/L		0.5	0	81	51 - 143%	PASS	2	30	PASS
Benzo[b]fluoranthene	Total	0.601	1	0.001	0.005	µg/L		0.5	0	120	46 - 165%	PASS	3	30	PASS
Benzo[e]pyrene	Total	0.525	1	0.001	0.005	µg/L		0.5	0	105	42 - 152%	PASS	2	30	PASS
Benzo[g,h,i]perylene	Total	0.452	1	0.001	0.005	µg/L		0.5	0	90	63 - 133%	PASS	1	30	PASS
Benzo[k]fluoranthene	Total	0.527	1	0.001	0.005	µg/L		0.5	0	105	56 - 145%	PASS	4	30	PASS
Biphenyl	Total	0.446	1	0.001	0.005	µg/L		0.5	0	89	56 - 119%	PASS	1	30	PASS
Chrysene	Total	0.434	1	0.001	0.005	µg/L		0.5	0	87	56 - 141%	PASS	4	30	PASS
Dibenz[a,h]anthracene	Total	0.57	1	0.001	0.005	µg/L		0.5	0	114	55 - 150%	PASS	4	30	PASS
Dibenzo[a,l]pyrene	Total	0.536	1	0.001	0.005	µg/L		0.5	0	107	50 - 150%	PASS	7	30	PASS



## Polynuclear Aromatic Hydrocarbons

## QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODE <sub>c</sub>	
							LEVEL	RESULT	%	LIMITS	%	LIMITS		
Dibenzothiophene	Total	0.446	1	0.001	0.005	µg/L	0.5	0	89	75 - 113%	PASS	0	30	PASS
Fluoranthene	Total	0.472	1	0.001	0.005	µg/L	0.5	0	94	60 - 146%	PASS	4	30	PASS
Fluorene	Total	0.435	1	0.001	0.005	µg/L	0.5	0	87	58 - 131%	PASS	2	30	PASS
Indeno[1,2,3-cd]pyrene	Total	0.601	1	0.001	0.005	µg/L	0.5	0	120	50 - 151%	PASS	3	30	PASS
Naphthalene	Total	0.425	1	0.001	0.005	µg/L	0.5	0	85	41 - 126%	PASS	2	30	PASS
Perylene	Total	0.491	1	0.001	0.005	µg/L	0.5	0	98	48 - 141%	PASS	3	30	PASS
Phenanthrene	Total	0.456	1	0.001	0.005	µg/L	0.5	0	91	67 - 127%	PASS	2	30	PASS
Pyrene	Total	0.47	1	0.001	0.005	µg/L	0.5	0	94	54 - 156%	PASS	4	30	PASS



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

**PHYSIS**

**TENTATIVELY**

**IDENTIFIED COMPOUNDS**

ENVIRONMENTAL LABORATORIES, INC.

*Innovative Solutions for Nature*

Sample ID: 98867

RT	Area Pct	Concentration (ng/L)	Library/ID	Cas Number	Match Qual
32.6331	8.0745	1111	Anthracene-D10	1517-22-2	96
43.1648	1.7795	245	Terephthalic acid, isobutyl butyl ester	1000323-56-2	95
64.5577	1.0491	144	1,4-Benzenedicarboxylic acid, bis(2-ethylhexyl) ester	6422-86-2	96
14.9651	1.0119	139	Cyclohexane, 1,2,4,5-tetraethyl-, (1.alpha.,2.alpha.,4.alpha.,5.alpha.)-	61142-24-3	83

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Sample ID: Lab Blank Batch O-38076

RT	Area Pct	Concentration (ng/L)	Library/ID	Cas Number	Match Qual
32.6362	6.4994	1111	Anthracene-D10-	1719-06-8	97
43.1684	2.4327	416	Terephthalic acid, isobutyl butyl ester	1000323-56-2	95
64.5707	1.3441	230	1,4-Benzenedicarboxylic acid, bis(2-ethylhexyl) ester	6422-86-2	97
14.9678	1.3108	224	3-Hexene, 3-ethyl-2,5-dimethyl-	62338-08-3	84
14.8062	0.6922	118	Cyclohexane, 1,2,4,5-tetraethyl-, (1.alpha.,2.alpha.,4.alpha.,5.alpha.)-	61142-24-3	85

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

# PERFORMANCE CHAIN OF CUSTODY


TERRA ENVIRONMENTAL LABORATORIES, INC. AURA

*Innovative Solutions for Nature*

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

**Monrovia, CA (Suite 100)**  
 750 Royal Oaks Drive Suite 100  
 Monrovia, CA 91016  
 Phone: 626-386-1100

**Chain of Custody Record**



<b>Client Information (Sub Contract Lab)</b>	<b>Sampler:</b>	<b>Lab PM:</b>	<b>COC No:</b>
Client Contact: <b>Shipping/Receiving</b>	Phone:	Frank, Debbie L	380-14995-1
Company: <b>Physis Environmental Laboratories</b>	Address: <b>1904 Wright Circle,</b>	State of Origin: <b>Hawaii</b>	Page: <b>1 of 1</b>
City: <b>Anaheim</b>	Due Date Requested: <b>8/5/2022</b>	Carrier Tracking No(s):	Job #: <b>380-13074-1</b>
State Zip: <b>CA, 92808</b>	TAT Requested (days):	Analysis Requested:	Preservation Codes: A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - NaHSO4 F - MeOH G - Acetic Acid H - Ascorbic Acid I - Ice J - DI Water K - EDTA L - EDA M - Hexane N - None O - AsNAO2 P - Na2CO3 Q - Na2SO3 R - Na2S2O3 S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4.5 Y - Triana Z - other (specify) Other:

Project Name:	Project #:	Project #:		
RED-HILL	38001111			
Sile:	SSOW#:			
Honolulu BWS Sites				

Sample Identification - Client ID (Lab ID)	Sample Date	Sample Time	Sample Type (C=Comp, G=Grab)	Preservation Code	MATRIX (W=Water, S=Soil, C=C-wastefact, BT=Tea, A=Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	SUB (625 PAH Physis LL (EAL) + TICs; 626 PAH Physis LL (EAL) + TICs)	Total Number of containers	Special Instructions/Note:
AIEA GULCH WELLS PUMP 2 (331-202-TPO72) (380-13074-1)	7/27/22	11:36	Water				X	SUB (625 PAH Physis LL (EAL) + TICs; 626 PAH Physis LL (EAL) + TICs)	4	See Attached Instructions

**Possible Hazard Identification**

**Unconfirmed**

Deliverable Requested: I, II, III, IV, Other (specify) \_\_\_\_\_ Primary Deliverable Rank: 2

Empty Kit Reinquished by: \_\_\_\_\_ Date: \_\_\_\_\_

Reinquished by: *G. RETNER* Date/Time: *08/01/2022* Company: *ETA*

Reinquished by: *Ann. T* Date/Time: *8-1-22* 16:45 Company: *ETA*

Reinquished by: \_\_\_\_\_ Date/Time: \_\_\_\_\_ Company: \_\_\_\_\_

Custody Seals Intact:  Yes  No Custody Seal No.:

Cooler Temperature(s) °C and Other Remarks:

**Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)**

Return To Client  Disposal By Lab  Archive For \_\_\_\_\_ Months

Special Instructions/OC Requirements:

Received by: *Ann. T* Date/Time: *8-1-22* Company: *ETA*

Received by: *Yvonney W. Cheeking* Date/Time: *8/1/22* 16:45 Company: *PHYSIS*



Project Iteration ID: 1407003-262  
 Client Name: Eurofins Eaton Analytical  
 Project Name: RED\_HILL Project # 38001111  
 Job # 380-13074-1  
 COC Page Number: 2 of 2  
 Bottle Label Color: NA

**Sample Receipt Summary**

Receiving Info

1. Initials Received By: XC
2. Date Received: 8/1/22
3. Time Received: 1645
4. Client Name: Eurofins
5. Courier Information: (Please circle)
  - Client
  - UPS
  - Area Fast
  - DRS
  - FedEx
  - GSO/GLS
  - Ontrac
  - PAMS
  - PHYSIS Driver:
  - i. Start Time: \_\_\_\_\_
  - ii. End Time: \_\_\_\_\_
  - iii. Total Mileage: \_\_\_\_\_
  - iv. Number of Pickups: \_\_\_\_\_
6. Container Information: (Please put the # of containers or circle none)
  - Cooler
  - Styrofoam Cooler
  - Boxes
  - None
  - Carboy(s)
  - Carboy Trash Can(s)
  - Carboy Cap(s)
  - Other \_\_\_\_\_
7. What type of ice was used: (Please circle any that apply)
  - Wet Ice
  - Blue Ice
  - Dry Ice
  - Water
  - None
8. Randomly Selected Samples Temperature (°C): 6.6  
 Used I/R Thermometer # 1-2

Inspection Info

1. Initials Inspected By: RGH

Sample Integrity Upon Receipt:

1. COC(s) included and completely filled out.....  Yes /  No
2. All sample containers arrived intact.....  Yes /  No
3. All samples listed on COC(s) are present.....  Yes /  No
4. Information on containers consistent with information on COC(s).....  Yes /  No
5. Correct containers and volume for all analyses indicated.....  Yes /  No
6. All samples received within method holding time.....  Yes /  No
7. Correct preservation used for all analyses indicated.....  Yes /  No
8. Name of sampler included on COC(s).....  Yes /  No

Notes:

*See temp*

**Monrovia, CA (Suite 100)**  
 750 Royal Oaks Drive Suite 100  
 Monrovia, CA 91016  
 Phone: 626-386-1100

# Chain of Custody Record



<b>Client Information</b>		Sampler: <b>E. JUAGDAN</b>		Lab PM: Frank, Debbie L		Carrier Tracking No(s): 380-9619-2757.2	
Client Contact Dr. Ron Fenstermacher		Phone: <b>808-748-5840</b>		E-Mail: Debbie.Frank@eurofins.com		Page: Page 2 of 3	
Company City & County of Honolulu		Due Date Requested:		Analysis Requested		Job #	
Address 630 South Beretania Street Chemistry Lab		TAT Requested (days):		SUBCONTRACT - 8015 Gas (Purgeable) LL (EAL)		Preservation Codes:	
City Honolulu		Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No		SUBCONTRACT - (MOD) 525plus Plus TICs		M - Hexane	
State/Zip HI, 96843		PO #: C20525101 exp 05312023		SUBCONTRACT - 8015 Diesel LL (EAL) and Motor Oil		N - None	
Phone: 808-748-5091(Tel)		WO #:		SUBCONTRACT - 8015 Gas (Purgeable) LL (EAL) + TICs		O - AsNaO2	
Email: RFENSTEMACHER@hbws.org		Project #: 38001111		SUBCONTRACT - 825 PAH Physis LL (EAL) + TICs		P - Na2OAS	
Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill		SSOW#:		Perform MS/MSD (Yes or No)		Q - Na2SO3	
Site: Hawaii		Sample Date		Field Filtered Sample (Yes or No)		R - Na2SO3	
Sample Identification		Sample Time		Matrix (V=water, S=solid, O=wastefoil, B=T-tissue, A=air)		S - H2SO4	
MOANALUA WELLS (331-223-TP202)				Water		T - TSP Dodecahydrate	
AIEA GULCH WELLS PUMP 1 (331-201-TP071)				Water		U - Acetone	
AIEA GULCH WELLS PUMP 2 (331-202-TP072)		7/27/22 1136 G		Water		V - MCAA	
AIEA WELLS PUMPS1&2(260)331-203-TP400				Water		W - pH 4-5	
HALAWA SHAFT (331-241-TP401)				Water		Y - Trizma	
HALAWA WELLS UNITS1&2(331-206-TP065)				Water		Z - other (specify)	
MOANALUA WELLS (331-223-TP202)				Water		Other:	
AIEA GULCH WELLS PUMP 1 (331-201-TP071)				Water		Total Number of containers	
AIEA GULCH WELLS PUMP 2 (331-202-TP072)				Water		Special Instructions/Note:	
AIEA WELLS PUMPS1&2(260)331-203-TP400				Water			
HALAWA SHAFT (331-241-TP401)				Water			
HALAWA WELLS UNITS1&2(331-206-TP065)				Water			
MOANALUA WELLS (331-223-TP202)				Water			
AIEA GULCH WELLS PUMP 1 (331-201-TP071)				Water			
AIEA GULCH WELLS PUMP 2 (331-202-TP072)				Water			
AIEA WELLS PUMPS1&2(260)331-203-TP400				Water			
HALAWA SHAFT (331-241-TP401)				Water			
<b>Possible Hazard Identification</b>				380-13074 COC			
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant							
<input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological							
Deliverable Requested: I, II, III, IV, Other (specify)							
Empty Kit Relinquished by:		Date:		Time:		Special Instructions/QC Requirements:	
Relinquished by:		7/28/22 1200				Sample Disposal ( A fee may be assessed if samples are retained longer than 1 month )	
Relinquished by:						<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months	
Custody Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks:			





**Bottle Order Information**

Bottle Order: RUSH RED-HILL WEEKLY  
 Bottle Order #: 2757  
 Request From Client: 7/20/2022  
 Date Order Posted: 7/20/2022 11:12:54AM  
 Order Status: Ready To Process  
 Prepared By: Davis Haley  
**Deliver By Date: 7/22/2022 11:59:00PM**  
 Lab Project Number: 38001111  
 PWSID: HI0000331

**Order Completion Information**

Creator: Davis Haley  
 Filled by:  
 Sent Date:  
 Sent Via:  
 Tracking #:

Sets	Bottles/Set	Qty	Bottle Type Description	Preservative	Method	Matrix	Sample Type	Comments	Lot #
6	4	24	Amber Glass 1 liter - Sodium Thiosulfate	Sodium Thiosulfate	SUBCONTRACT - 625 PAH Physis LL (EAL) + TICs	Water	Normal	625 PAH + MS/MSD Volume	
6	4	24	Voa Vial 40ml - SodiumThio w/HCl-dropper	Sodium Thiosulfate	SUBCONTRACT - 8015 Gas (Purgeable) LL (EAL)	Water	Normal		
6	2	12	Amber Glass 1 L - NaThiosulfate 8mL HCL	Sodium Thiosulfate/Hydrochloric Acid	SUBCONTRACT - 8015 Diesel LL (EAL) and Motor Oil	Water	Normal		
6	2	12	Amber Glass 1 Liter- Sodium Sulfite/HCl	Sodium Sulfite w/HCl	525.2_PREC - (MOD) 525plus Plus TICs	Water	Normal		
6	0	0	VOA Vial 40mL - NaThiosulfate/HCL	Sodium Thiosulfate/Hydrochloric Acid	SUBCONTRACT - 8015 Gas (Purgeable) LL (EAL)	Water	Trip Blank		

<b>Total Bottle Summary</b>		
Bottle Type Description	Preservative	Bottle Count
Amber Glass 1 L - NaThiosulfate 8mL HCL	Sodium Thiosulfate/Hydrochloric Acid	12
Amber Glass 1 liter - Sodium Thiosulfate	Sodium Thiosulfate	24
Amber Glass 1 Liter- Sodium Sulfite/HCl	Sodium Sulfite w/HCl	12
VOA Vial 40mL - NaThiosulfate/HCL	Sodium Thiosulfate/Hydrochloric Acid	0
Voa Vial 40ml - SodiumThio w/HCl-dropper	Sodium Thiosulfate	24
Total Bottles:		<u>72</u>

Please notify your PM immediately if an error is found in shipment. When returning samples, please return all provided QC samples.



# INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number:

**SAMPLE TEMP RECEIVED:**  
 Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.  
**SAMPLES REC'D DAY OF COLLECTION? Yes / No**

IR Gun ID = 649A (Observation = 5.2 °C) (Corr. Factor -0.3 °C) (Final = 4.9 °C)

TYPE OF ICE: Real  Synthetic  No Ice  CONDITION OF ICE: Frozen  Partially Frozen  Thawed  N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: \_\_\_\_\_

**Compliance Acceptance Criteria:**

- 1) Chemistry: >0, ≤ 6°C, not frozen (NELAP) (If received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥ 10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)

4 Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (If received after 24 hrs of sample collection)

5) pH Check. Manufacturer: \_\_\_\_\_ Lot Number: \_\_\_\_\_ pH strip type: 0 - 14 or \_\_\_\_\_ Expiration Date \_\_\_\_\_ Results: \_\_\_\_\_

6) Chlorine check. Manufacturer: Sansafe. Lot No.: \_\_\_\_\_ Expiration Date: \_\_\_\_\_ Results: \_\_\_\_\_

7) VOA and Radon Headspace:

No Samples with Headspace:


Samples with Headspace (see below):

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Example from headspace concerns: Methods 815.4, HAA(8251,882), 806, SPME, @CH, 832LCMS, 888, 838, Anatoxin, LCMS methods using 40 ml vials, International clients:

None/<8 mm				None/<8 mm				None/<8 mm				None/<8 mm			
Samp ID	Bottle #	>8mm	Test	Samp ID	Bottle #	>8mm	Test	Samp ID	Bottle #	>8mm	Test	Samp ID	Bottle #	>8mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): \_\_\_\_\_

RECEIVED BY:	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
		G. REUTNER	Eurofins Eaton Analytical	07/29/2022	10:00
SAMPLES CHECKED AGAINST COC BY:	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
			Eurofins Eaton Analytical		

ORIGIN ID:HIKA (808) 748-5840  
 BWS CHEMLAB  
 HONOLULU BOARD OF WATER SUPPLY  
 630 S. BERETANIA ST  
 CHEMICAL LABORATORY  
 HONOLULU, HI 96843  
 UNITED STATES US

SHIP DATE: 28JUL22  
 ACTWGT: 50.00 LB  
 CAD: 100205419/MNET4490


BILL RECIPIENT

TO **C CHUCK**  
**EUROFINS EATON ANALYTICAL, INC**  
**750 ROYAL OAKS DR**  
**SUITE 100**  
**MONROVIA CA 91016**

INV (626) 386-1178 REF

58120492FE4A

DEPT:



TRK# 7775 1635 3737

PRIORITY OVERNIGHT

FRI - 29 JUL 10:30A

**WZ WHPA**

91016  
 CA-US  
 BUR



**Warning:** Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.  
 Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

After printing this label:



# Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-13074-1

**Login Number: 13074**  
**List Number: 1**  
**Creator: Ngo, Theodore**

**List Source: Eurofins Eaton Monrovia**

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	