


750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Report

for

Honolulu Board of Water Supply
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843
Attention: Erwin Kawata
Fax: 808-550-5018



Utah ELCP CA00006

DEB: Debbie L Frank
Project Manager

Report: 988199
Project: RED-HILL
Group: Red-Hill Expanded List (Albuquerque+)

* Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.

* Laboratory certifies that the test results meet all **TNI 2016 and ISO/IEC 17025:2017** requirements unless noted under the individual analysis.

* As applicable, this report consists of the cover page, State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms.

* Test results relate only to the sample(s) tested.

* Test results apply to the sample(s) as received, unless otherwise noted in the comments report (ISO/IEC 17025:2017).

* This report shall not be reproduced except in full, without the written approval of the laboratory.

* This report includes ISO/IEC 17025 and non-ISO 17025 accredited methods.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Montana	Cert 0035
Arizona	AZ0778	Nebraska	NE-OS-21-13
Arkansas	CA00006	Nevada	CA00006
California	2813	New Hampshire *	2959
Colorado	CA00006	New Jersey *	CA 008
Connecticut	PH-0107	New Mexico	CA00006
Delaware	CA 006	New York *	11320
Florida *	E871024	North Carolina	06701
Georgia	947	North Dakota	R-009
Guam	21-008R	Ohio - 537.1	87786
Hawaii	CA00006	Oregon *	4034
Idaho	CA00006	Pennsylvania *	68-00565
Illinois	200033	Puerto Rico	CA00006
Indiana	C-CA-01	Rhode Island	LAO00326
Iowa – Asbestos	413	South Carolina	87016
Kansas *	E-10268	South Dakota	CA11320
Kentucky	90107	Tennessee	TN02839
Louisiana *	LA008	Texas *	T104704230-20-18
Maine	CA00006	Utah (Primary AB) *	CA00006
Maryland	224	Vermont	VT0114
Marianas Islands	MP0004	Virginia *	460260
Massachusetts	M-CA006	Washington	C838
Michigan	9906	EPA Region 5	CA00006
Mississippi	CA00006	Los Angeles County Sanitation Districts	10264

* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025:2917 Accredited Method List

The test listed below are accredited and met the requirements of ISO/IEC 17025 as verify by A2LA.

Refer to our certificates and scope of accreditations (no. 5890-1 and 5890-2) found at:

<https://www.eurofinsus.com/Eaton>

Test(s)	Method(s)	Potable Water *	Waste Water
Enterococci	Enterolert	x	x
Escherichia coli (Enumeration)	SM 9221 B.1 SM 9221 F	x	
Fecal Coliform (P/A and Enumeration)	SM 9221 C (MTF/EC), SM 9221 E (MTF/EC)	x	x
Fecal Streptococci and Enterococci	SM 9230 B	x	x
Heterotrophic Bacteria	SM 9215 B	x	
Legionella	Legiolert®	x	
Pseudomonas aeruginosa	Idexx Pseudalart	x	
Total Coliform (P/A and Enumeration)	SM 9221A, SM 9221B, SM 9221 C	x	x
Total Coliform, Total Coliform with Chlorine Present	SM 9221 B	x	x
Total Coliform/E. coli (P/A and Enumeration, Idexx Colilert, Idexx Colilert 18, Colisure)	SM 9223	x	
Total Microcystins and Nodularins	EPA 546	X	
Yeast and Mold	SM 9610	x	
1,2,3-Trichloropropane (TCP) at 5 PPT	CA SRL 524M-TCP	x	
1,4-Dioxane	EPA 522	x	
2,3,7,8-TCDD	Modified EPA 1613 B	x	
Acrylamide	+ LCMS 2440)	x	
Algal Toxins/Microcystin	+ LCMS 3570	x	
Alkalinity	SM 2320B	x	x
Ammonia	EPA 350.1, SM 4500-NH3 H		x
Asbestos	EPA 100.2	x	x
Bicarbonate Alkalinity as HCO3	SM 2330 B	x	x
BOD/CBOD	SM 5210 B		x
Bromate	+ LCMS- 2447	x	
Carbonate as CO3	SM 2330 B	x	x
Carbonyls	EPA 556	x	x
Chemical Oxygen Demand	EPA 410.4, SM 5220D		x
Chlorinated Acids	EPA 515.4	x	
Chlorine Dioxide	Palin Test Chlordio X Plus, SM 4500-CLO2 D	x	
Chlorine, Free, Combined, Total Residual, Chloramines	SM 4500-Cl G	x	
Color	SM2120B	x	
Conductivity	EPA 120.1, SM 2510B	x	x
Corrosivity (Langelier Index), Carbonate as CO3, Hydroxide as OH Calculated	SM 2330 B	x	
Cyanide (Amenable)	SM 4500-CN G	x	x
Cyanide (Free)	SM 4500CN F	x	x
Cyanide (Total)	EPA 335.4	x	x
Cyanogen Chloride (Screen)	+ 335 Mod (WC-24467)	x	
Diquat and Paraquat	EPA 549.2	x	
DBP and HAA	SM 6251 B	x	
Dissolved Organic Carbon	SM 5310 C	x	
Dissolved Oxygen	SM 4500-O G		x
EDB/DCBP/TCP	EPA 504.1	x	
EDB/DBCP and Disinfection Byproducts	EPA 551.1	x	
EDTA and NTA	+ WC-2454	x	
Endothall	EPA 548.1, +(LCMS-2445)	x	
Fluoride	SM 4500F C	x	x
Glyphosate	EPA 547	x	
Glyphosate and AMPA	+ LCMS-3618	x	
Gross Alpha and Gross Beta	EPA 900.0	x	x
Gross Alpha coprecipitation	SM 7110 C	x	x
Hardness	SM 2340 B	x	x
Hexavalent Chromium	EPA 218.6,	x	x
Hexavalent Chromium	EPA 218.7,	x	
Hexavalent Chromium	SM 3500-Cr B		x
Inorganic Anions and DBPs	EPA 300.0	x	x
Norganic Anions and DBPs	EPA 300.1	x	
Kjeldahl Nitrogen	EPA 351.2		x
Metals	EPA 200.7, EPA200.8	x	x
Nitrosamines	EEA-Agilent 521.1 (GCMS-24250)	x	
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x
Odor	SM2150B	x	
Organohalide Pesticides and PCB	EPA 505	x	
Ortho Phosphate	SM 4500P E	x	
Oxyhalides Disinfection Byproducts	EPA 317.0	x	
Perchlorate	EPA 331.0	x	
Perchlorate (Low and High Levels)	EPA 314.0	x	
Perfluorinated Alkyl Acids	EPA 533, EPA 537, EPA 537.1	x	
PPCP and EDC	+ LCMS-2443	x	
pH	EPA 150.1 SM 4500-H+ B	x	x
Phenolics – Low Level	+WC 2493 (EPA 420.2 and EPA 420.4 MOD)	x	x
Phenylurea Pesticides/Herbicides	+ LCMS-2448	x	
Radium-226, Radium-228	GA Tech (Rad-2374)	x	
Radon-222	SM 7500RN	x	
Residue (Filterable)	SM 2540C	x	x
Residue (Non-Filterable)	SM 2540D		x
Residue (Total)	SM 2540B		x
Residue (Volatile)	EPA 160.4		x
Semi-Volatile Compounds	EPA 525.2	x	
Silica	SM 4500-SiO2 C	x	x
Sulfide	SM 4500-S D		x
Sulfite	SM 4500-SO3 B	x	x
Surfactants	SM 5540C	x	x
Taste and Odor	SM 6040 E	x	
Total Organic Carbon	SM 5310 C	x	x
Total Phenols	EPA 420.1		x
Total Phenols	EPA 420.4	x	x
Triazine Pesticides and their Degradates	+ LCMS-3617	x	
Turbidity	EPA 180.1	x	x
Uranium by ICP/MS	EPA 200.8	x	
UV 254 Organic Constituents	SM 5910B	x	
VOCs	EPA 524.2	x	
VOCs	+(GCMS 2412) by EPA 524.2 modified	x	

(*) includes: Bottled Water, Drinking Water and Water as Component of Food & Beverage.

(+) In-House Method

Acknowledgement of Samples Received

Addr: **Honolulu Board of Water Supply**
 630 South Beretania Street
 Public Service Bldg." Room 308
 Honolulu, HI 96843

Attn: Erwin Kawata
 Phone: 808-748-5091

Client ID: HONOLULU
 Folder #: 988199
 Project: RED-HILL
 Sample Group: Red-Hill Expanded List
 (Albuquerque+)
 Project Manager: Debbie L Frank
 Phone: (626) 386-1149
 PO #: C20525101 exp 05312023

The following samples were received from you on **February 17, 2022** at **1637**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
<u>202202170722</u>	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	02/15/2022 0915
	(SUB)Gas Fraction Hydrocarbons TPH 8015 Diesel and Motor Oil TPH 8015 Jet Fuel 5 TPH 8015 Jef Fuel 8	
<u>202202170723</u>	TRAVEL BLANK::AIEA GULCH WELLS PUMP 1 (331-201-TP071)	02/15/2022 0915
	(SUB)Gas Fraction Hydrocarbons	
<u>202202170724</u>	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	02/15/2022 0915
	(SUB)Gas Fraction Hydrocarbons TPH 8015 Diesel and Motor Oil TPH 8015 Jet Fuel 5 TPH 8015 Jef Fuel 8	
<u>202202170725</u>	TRAVEL BLANK::AIEA GULCH WELLS PUMP 2 -331-202-TP072	02/15/2022 0915
	(SUB)Gas Fraction Hydrocarbons	

Test Description



Eaton Analytical

CHAIN OF CUSTODY RECORD

EUROFINS EATON ANALYTICAL USE ONLY:

LOGIN COMMENTS: _____

SAMPLES CHECKED AGAINST COC BY: VP

SAMPLES LOGGED IN BY: _____

SAMPLE TEMP RECEIVED AT: _____ °C (check for yes)

Colton / No. California / Arizona

Monrovia

5.7 °C (Compliance: 4 ± 2 °C)

CONDITION OF BLUE ICE: Frozen Partially-Frozen _____ Thawed _____ No Ice _____

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

750 Royal Oaks Drive, Suite 100
 Monrovia, CA 91016-3629
 Phone: 626 386 1100
 Fax: 626 386 1101
 800 566 LABS (800 566 5227)

TO BE COMPLETED BY SAMPLER: _____ (check for yes)

COMPANY/AGENCY NAME: BWS HONOLULU

PROJECT CODE: Red Hill

COMPLIANCE SAMPLES NON-COMPLIANCE SAMPLES

- Requires state forms REGULATION INVOLVED: _____

EEA CLIENT CODE: Honolulu

COC ID: _____

Type of samples (circle one): ROUTINE SPECIAL CONFIRMATION (eg. SDWA, Phase V, NPDES, FDA,...)

SEE ATTACHED BOTTLE ORDER FOR ANALYSES (check for yes), OR

list ANALYSES REQUIRED (enter number of bottles sent for each test for each sample)

SAMPLE DATE	SAMPLE TIME	SAMPLE ID	CLIENT LAB ID	MATRIX	FIELD DATA	FIELD DATA	SAMPLER COMMENTS
2-15-22	0915	AIEA GULCH WELLS PUMP 1	HI0000331-201	CFW			
2-15-22	0915	AIEA GULCH WELLS PUMP 2	HI0000331-202	CFW			
		Temperature Blank					

Temp Blank: 13.0 °C

* MATRIX TYPES: RSW = Raw Surface Water CFW = Chlor(am)inated Finished Water SEAW = Sea Water BW = Bottled Water SO = Soil O = Other - Please Identify
 RGW = Raw Ground Water FW = Other Finished Water WW = Waste Water SW = Storm Water SL = Sludge

SAMPLED BY:	REINQUISHED BY:	RECEIVED BY:	REINQUISHED BY:	RECEIVED BY:	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
<u>[Signature]</u>	<u>[Signature]</u>	<u>[Signature]</u>	<u>[Signature]</u>	<u>[Signature]</u>	Derek Dotson	Derek Dotson	Honolulu Board of Water Supply	2-15-2022	1200
					<u>[Signature]</u>	<u>[Signature]</u>	Honolulu Board of Water Supply	2-16-2022	1637



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EBA Folder Number: 98899

SAMPLE TEMP RECEIVED:

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = 630 (Observation = 5.9 °C) (Corr.Factor = -0.2 °C) (Final = 5.7 °C)

TYPE OF ICE: Real Synthetic No Ice Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

4) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check. Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____

6) Chlorine check. Manufacturer: Sansafe. Lot No.: _____ Expiration Date: _____ Results: _____

7) VOA and Radon Headspace: No Samples with Headspace: Samples with Headspace (see below):

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Exempt from headspace concerns: Methods 615.4, HAA(6251,652), 505, SPME, @CH, 632LCMS, 556, 536, Anatoxin, LCMS methods using 40 ml vials, international clients:

Samp ID	Bottle #	None/<6	>6mm	Test	Samp ID	Bottle #	None/<6	>6mm	Test	Samp ID	Bottle #	None/<6	>6mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): _____

RECEIVED BY: [Signature] SIGNATURE Victor Vasconcelos PRINT NAME Eurofins Eaton Analytical COMPANY/TITLE 2-17-22 DATE 1639 TIME

SAMPLES CHECKED AGAINST COC BY: [Signature] SIGNATURE Eurofins Eaton Analytical COMPANY/TITLE _____ DATE _____ TIME _____



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: 988194

SAMPLE TEMP RECEIVED:

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes NO

IR Gun ID = 630 (Observation = 5.4 °C) (Corr. Factor = -0.2 °C) (Final = 5.2 °C)

TYPE OF ICE: Real Synthetic No Ice Condition of Ice: Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C)

4) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check. Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____

6) Chlorine check. Manufacturer: Sansafe. Lot No.: _____ Expiration Date: _____ Results: _____

7) VOA and Radon Headspace: _____

No Samples with Headspace: _____

Samples with Headspace (see below): _____

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Exempt from headspace concerns: Methods 515.4, HAA(6251,662), 505, SPME, @CH, 532LCMS, 556, 538, Anatoxin, LCMS methods using 40 ml vials, International clients:

Samp ID	Bottle #	None/<6	>6mm	Test	Samp ID	Bottle #	None/<6	>6mm	Test	Samp ID	Bottle #	None/<6	>6mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): _____

RECEIVED BY: [Signature] SIGNATURE: [Signature] PRINT NAME: Victor Vasenin COMPANY/TITLE: Eurofins Eaton Analytical DATE: 2.17.22 TIME: 1630

SAMPLES CHECKED AGAINST COG BY: _____ SIGNATURE: _____ PRINT NAME: Eurofins Eaton Analytical COMPANY/TITLE: _____ DATE: _____ TIME: _____

ORIGIN ID: HKA (808) 748-5840
BWS CHEM LAB
HONOLULU BOARD OF WATER SUPPLY
630 S. BERETANIA ST.
CHEMICAL LABORATORY
HONOLULU, HI 96843
UNITED STATES US

SHIP DATE: 16FEB22
ACTWGT: 60.00 LB
CAD: 100205419/NET 4460
BILL RECIPIENT


TO C CHUCK

EUROFINS EATON ANALYTICAL, INC
750 ROYAL OAKS DR
SUITE 100

MONROVIA CA 91016

REF: (820) 386-1178
INV: PO: DEPT:

56D.J2027CfE4A



JZ21022010581uv

1 of 2


TRK# 7760 6219 6012

0201
MASTER

THU - 17 FEB 10:30A
PRIORITY OVERNIGHT

WZ WHPA

91016
CA-US BUR



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

ORIGIN: DHIKA (808) 748-5840
BWS CHEM/LAB
HONOLULU BOARD OF WATER SUPPLY
630 S. BERETANIA ST.
CHEMICAL LABORATORY
HONOLULU, HI 96843
UNITED STATES US

SHIP DATE: 16FEB22
ACTWGT: 80.00 LB
CAD: 100205419/NET4460
BILL RECIPIENT

TO C CHUCK

EUROFINS EATON ANALYTICAL, INC


750 ROYAL OAKS DR

SUITE 100

MONROVIA CA 91016

REF: (626) 386-1178

DEPT: INV: PO:



56DJ2027C/FE4A

2 of 2

MPS# 7760 6219 5932


Mstr# 7760 6219 6012

0201

THU - 17 FEB 10:30A
PRIORITY OVERNIGHT

WZ WHPA

CA-US 91016
BUR



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 988199
Project: RED-HILL
Group: Red-Hill Expanded List
(Albuquerque+)

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Folder Comments

Results for TPH Gas, Diesel, Motor Oil and Jet Fuels are submitted by Emax in Torrance CA



Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 988199
Project: RED-HILL
Group: Red-Hill Expanded List
(Albuquerque+)

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Samples Received on:
02/17/2022 1637

Analyzed	Analyte	Sample ID	Result	HI Limit	Units	MRL
----------	---------	-----------	--------	----------	-------	-----

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 988199
Project: RED-HILL
Group: Red-Hill Expanded List
 (Albuquerque+)

Honolulu Board of Water Supply
 Erwin Kawata
 630 South Beretania Street
 Public Service Bldg.” Room 308
 Honolulu, HI 96843

Samples Received on:
 02/17/2022 1637

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
<u>AIEA GULCH WELLS PUMP 1 (331-201-TP071) (202202170722)</u>						Sampled on 02/15/2022 0915			
SW 8015B - (SUB)Gas Fraction Hydrocarbons									
02/18/22	02/18/22 18:10			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1
SW 8015B - TPH 8015 Diesel and Motor Oil									
02/21/22	02/23/22 02:10			(SW 8015B)	TPH Diesel	ND	mg/L	0.028	1
02/21/22	02/23/22 02:10			(SW 8015B)	TPH Motor Oil	ND	mg/L	0.055	1
EPA 8015 - Jet Fuel 5 C8-C18									
02/21/22	02/23/22 02:10			(EPA 8015)	Jet Fuel 5	ND	mg/L	0.055	1
EPA 8015 - Jet Fuel 8 C8-C18									
	02/23/22 02:10			(EPA 8015)	Jet Fuel 8	ND	mg/L	0.055	1
<u>TRAVEL BLANK::AIEA GULCH WELLS PUMP 1 (331-201-TP071) (202202170723)</u>						Sampled on 02/15/2022 0915			
SW 8015B - (SUB)Gas Fraction Hydrocarbons									
02/18/22	02/18/22 19:59			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1
<u>AIEA GULCH WELLS PUMP 2 (331-202-TP072) (202202170724)</u>						Sampled on 02/15/2022 0915			
SW 8015B - (SUB)Gas Fraction Hydrocarbons									
02/18/22	02/18/22 20:35			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1
SW 8015B - TPH 8015 Diesel and Motor Oil									
02/21/22	02/23/22 02:28			(SW 8015B)	TPH Diesel	ND	mg/L	0.027	1
02/21/22	02/23/22 02:28			(SW 8015B)	TPH Motor Oil	ND	mg/L	0.055	1
EPA 8015 - Jet Fuel 5 C8-C18									
02/21/22	02/23/22 02:28			(EPA 8015)	Jet Fuel 5	ND	mg/L	0.055	1
EPA 8015 - Jet Fuel 8 C8-C18									
	02/23/22 02:28			(EPA 8015)	Jet Fuel 8	ND	mg/L	0.055	1
<u>TRAVEL BLANK::AIEA GULCH WELLS PUMP 2 -331-202-TP072 (202202170725)</u>						Sampled on 02/15/2022 0915			
SW 8015B - (SUB)Gas Fraction Hydrocarbons									
02/18/22	02/18/22 21:12			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1

Rounding on totals after summation.
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.



3051 Fujita Street
Torrance, CA 90505
Tel: (310)-618-8889

Date: 03-11-2022
EMAX Batch No.: 22B192

Attn: Jackie Contreras

Eurofins Eaton Analytical
750 Royal Oaks Dr., Suite 100
Monrovia, CA 91016-3629

Subject: Laboratory Report
Project: 988199


Enclosed is the Laboratory report for samples received on 02/18/22.
The data reported relate only to samples listed below :

Sample ID	Control #	Col Date	Matrix	Analysis
202202170722	B192-01	02/15/22	WATER	TPH GASOLINE TPH
202202170723	B192-02	02/15/22	WATER	TPH GASOLINE
202202170724	B192-03	02/15/22	WATER	TPH GASOLINE TPH
202202170725	B192-04	02/15/22	WATER	TPH GASOLINE

The results are summarized on the following pages.

Please feel free to call if you have any questions concerning these results.


Sincerely yours,


Gaspar J. Pang
Laboratory Director

This report is confidential and intended solely for the use of the individual or entity to whom it is addressed. This report shall not be reproduced except in full or without the written approval of EMAX.

EMAX certifies that results included in this report meets all TNI & DOD requirements unless noted in the Case Narrative.

NELAP Accredited Certificate Number CA002912021-19
ANAB Accredited DoD ELAP and ISO/IEC 17025 Certificate Number L2278 Testing
California ELAP Accredited Certificate Number 2672



Eaton Analytical

Ship To:
EMAX Laboratories, Inc.
3051 Fujita St.
Torrance, CA 90505

Phone: 310-618-8889 Fax: 310-618-0818

Folder #: 988199 Report Due: 02/24/2022

Sample ID: 202202170722 ¹ Client Sample ID for reference on: AIEA GULCH WELLS PUMP 1 (331-201-TP071)

Sample type: SW 8015B EPA 5030C
SW 8015B EPA 3550B
EPA 8015 EPA 8015
EPA 8015 Jet Fuel 5 C8-C18
Jet Fuel 8 C8-C18

Method: SW 8015B EPA 5030C
SW 8015B EPA 3550B
EPA 8015 EPA 8015
EPA 8015 Jet Fuel 5 C8-C18
Jet Fuel 8 C8-C18

Analysis Requested: (SUB)Gas Fraction Hydrocarbons
TPH 8015 Diesel and Motor Oil
Jet Fuel 5 C8-C18
Jet Fuel 8 C8-C18

Sample ID: 202202170723 ² Client Sample ID for reference on: TRAVEL BLANK: AIEA GULCH WELLS PUMP 1 (331-201-TP071)

Sample type: SW 8015B EPA 5030C

Method: SW 8015B EPA 5030C

Analysis Requested: (SUB)Gas Fraction Hydrocarbons

Relinquished by: [Signature]

Received by: [Signature]

Relinquished by: [Signature]

Received by: [Signature]

Date: 2/18/22 Time: 10:23

Date: 2/18/22 Time: 10:23

Date: _____ Time: _____

Date: _____ Time: _____

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS

An Acknowledgement of Receipt is requested to attain Jackie Contreras

Temp. 4.0/3.5, 3.2/2.7

Date: 2/18/2022

22B192

Submittal Form

*REPORTING REQUIREMENTS: Do Not Combine Reports with any other samples submitted under different Folder Numbers!
Report & Invoice must have the Folder# 988199 Job # 1000014

Report all quality control data according to Method. Include dates analyzed, Date extracted (if extracted) and Method reference on the report.
Results must have Complete data & QC with Approval Signature

Reports: Jackie Contreras Sub-Contracting Administrator
EMAIL TO: Eaton-MonroviaSubContract@eurofins.com
Eurofins Eaton Analytical, LLC 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016
Phone (626) 386-1165 Fax (626) 386-1122
Invoices to: Eurofins Eaton Analytical, LLC
Accounts Payable 2425 New Holland Pike, Lancaster, PA 17605

Provide in each Report the
Specified State Certification # and
Exp Date for requested tests + matrix
Samples from: HAWAII

Sample Date & Time Matrix Clip Code PWSID
02/15/22 0915 DW JLS

Sample Point ID: Static ID:

Sample Date & Time Matrix Clip Code PWSID
02/15/22 0915 DW JLS

Sample Point ID: Static ID:

Sample ID 202202170724 (3) **Client Sample ID for reference on!** 22B92 **Sample Date & Time** 02/15/22 09:15 DW **Matrix** DW **Clip Code** **PWSID** **Static ID:** JLS
Sample type: AIEA GULCH WELLS PUMP 2 (331-202-TP072) **Sample Point ID:** **Facility ID:**

Method	Prep Method	Analysis Requested
SW 8015B	EPA 5030C	(SUB)Gas Fraction Hydrocarbons
SW 8015B	EPA 3550B	TPH 8015 Diesel and Motor Oil
EPA 8015	EPA 8015	Jet Fuel 5 C8-C18
EPA 8015		Jet Fuel 8 C8-C18

Sample ID 202202170725 (4) **Client Sample ID for reference on!** **Sample Date & Time** 02/15/22 09:15 DW **Matrix** DW **Clip Code** **PWSID** **Static ID:** JLS
Sample type: TRAVEL BLANK-AIEA GULCH WELLS PUMP 2 -331-202-TP072 **Sample Point ID:** **Facility ID:**

Method	Prep Method	Analysis Requested
SW 8015B	EPA 5030C	(SUB)Gas Fraction Hydrocarbons

Relinquished by: VP **Sample Control** **Date:** 2/18/22 **Time:** 10:23
Received by: [Signature] **Date:** 2/18/22 **Time:** 10:23
Relinquished by: _____ **Sample Control** **Date:** _____ **Time:** _____
Received by: _____ **Date:** _____ **Time:** _____

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS
 An Acknowledgment of Receipt is requested to attn: Jackie Contreras

Type of Delivery	Airbill / Tracking Number	ECN 22B192
<input type="checkbox"/> Fedex <input type="checkbox"/> UPS <input type="checkbox"/> GSO <input type="checkbox"/> Others		Recipient Alan Ramos
<input type="checkbox"/> EMAX Courier <input checked="" type="checkbox"/> Client Delivery		Date 02/18/22 Time 10:23

COC INSPECTION

<input checked="" type="checkbox"/> Client Name	<input checked="" type="checkbox"/> Client PM/FC	<input type="checkbox"/> Sampler Name	<input checked="" type="checkbox"/> Sampling Date/Time	<input checked="" type="checkbox"/> Sample ID	<input checked="" type="checkbox"/> Matrix
<input checked="" type="checkbox"/> Address	<input checked="" type="checkbox"/> Tel # / Fax #	<input type="checkbox"/> Courier Signature	<input checked="" type="checkbox"/> Analysis Required	<input type="checkbox"/> Preservative (if any)	<input type="checkbox"/> TAT
Safety Issues (if any)	<input type="checkbox"/> High concentrations expected	<input type="checkbox"/> From Superfund Site	<input type="checkbox"/> Rad screening required		

Note: _____

PACKAGING INSPECTION

Container <i>correction</i>	<input checked="" type="checkbox"/> Cooler	<input type="checkbox"/> Box	<input type="checkbox"/> Other
Condition <i>Factor:</i>	<input type="checkbox"/> Custody Seal	<input type="checkbox"/> Intact	<input type="checkbox"/> Damaged
Packaging <i>-0.5</i>	<input checked="" type="checkbox"/> Bubble Pack	<input type="checkbox"/> Styrofoam	<input type="checkbox"/> Popcorn
Temperatures (Cool, ≤6 °C but not frozen)	<input checked="" type="checkbox"/> Cooler 1 <u>4.0/35</u> °C	<input checked="" type="checkbox"/> Cooler 2 <u>3.2/2.7</u> °C	<input type="checkbox"/> Cooler 3 _____ °C
Thermometer:	<input type="checkbox"/> Cooler 6 _____ °C	<input type="checkbox"/> Cooler 7 _____ °C	<input type="checkbox"/> Cooler 4 _____ °C
	<input type="checkbox"/> Cooler 8 _____ °C	<input type="checkbox"/> Cooler 9 _____ °C	<input type="checkbox"/> Cooler 5 _____ °C
	A - S/N <u>210191066</u> <i>14/14</i>	B - S/N <u>210271396</u>	<input type="checkbox"/> Cooler 10 _____ °C
		C - S/N <u>210271399</u>	D - S/N _____

Comments: Temperature is out of range. PM was informed IMMEDIATELY.

Note: _____

DISCREPANCIES

LabSampleID	LabSampleContainerID	Code	ClientSample Label ID / Information	Corrective Action
1,3	4-12, 19-24	D22		R8
2	13-14	D23		R1

pH holding time requirement for water samples is 15 mins. Water samples for pH analysis are received beyond 15 minutes from sampling time.

NOTES/OBSERVATIONS: Extra volume for sample 1 Ambers.

LEGEND:

<p>Code Description- Sample Management</p> <p>D1 Analysis is not indicated in _____</p> <p>D2 Analysis mismatch COC vs label</p> <p>D3 Sample ID mismatch COC vs label</p> <p>D4 Sample ID is not indicated in _____</p> <p>D5 Container -[improper] [leaking] [broken]</p> <p>D6 Date/Time is not indicated in _____</p> <p>D7 Date/Time mismatch COC vs label</p> <p>D8 Sample listed in COC is not received</p> <p>D9 Sample received is not listed in COC</p> <p>D10 No initial/date on corrections in COC/label</p> <p>D11 Container count mismatch COC vs received</p> <p>D12 Container size mismatch COC vs received</p>	<p>Code Description-Sample Management</p> <p>D13 Out of Holding Time</p> <p>D14 Bubble is >6mm</p> <p>D15 No trip blank in cooler</p> <p>D16 Preservation not indicated in _____</p> <p>D17 Preservation mismatch COC vs label</p> <p>D18 Insufficient chemical preservative</p> <p>D19 Insufficient Sample</p> <p>D20 No filtration info for dissolved analysis</p> <p>D21 No sample for moisture determination</p> <p>D22 <u>Jet Fuel 8, Analysis not indicated</u></p> <p>D23 <u>Two dates 7/14/15 on label</u></p> <p>D24 _____</p>	<p><input type="checkbox"/> Continue to next page.</p> <p>Code Description-Sample Management</p> <p>R1 Proceed as indicated in COC <input type="checkbox"/> Label</p> <p>R2 Refer to attached instruction</p> <p>R3 Cancel the analysis</p> <p>R4 Use vial with smallest bubble first</p> <p>R5 Log-in with latest sampling date and time+1 min</p> <p>R6 Adjust pH as necessary</p> <p>R7 Filter and preserved as necessary</p> <p>R8 <u>Informed client</u></p> <p>R9 _____</p> <p>R10 _____</p> <p>R11 _____</p> <p>R12 _____</p>
--	---	---

REVIEWS:

Sample Labeling <u>Marla Rivera</u>	SRF <u>Debra</u>	PM <u>RB</u>
Date <u>02/18/22</u>	Date <u>2/18/22</u>	Date <u>2/18/22</u>

REPORTING CONVENTIONS

DATA QUALIFIERS:

Lab Qualifier	AFCEE Qualifier	Description
J	F	Indicates that the analyte is positively identified and the result is less than RL but greater than MDL.
N		Indicates presumptive evidence of a compound.
B	B	Indicates that the analyte is found in the associated method blank as well as in the sample at above QC level.
E	J	Indicates that the result is above the maximum calibration range or estimated value.
*	*	Out of QC limit.

Note: The above qualifiers are used to flag the results unless the project requires a different set of qualification criteria.

ACRONYMS AND ABBREVIATIONS:

CRDL	Contract Required Detection Limit
RL	Reporting Limit
MRL	Method Reporting Limit
PQL	Practical Quantitation Limit
MDL	Method Detection Limit
DO	Diluted out

DATES

The date and time information for leaching and preparation reflect the beginning date and time of the procedure unless the method, protocol, or project specifically requires otherwise.

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

988199

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

SDG#: 22B192

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 988199

SDG : 22B192

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

A total of four(4) water samples were received on 02/18/22 to be analyzed for Total Petroleum Hydrocarbons by Purge and Trap in accordance with Method 5030B/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. VG39B11B - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. VG39B11L/VG39B11C were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Gasoline was within MS QC limits in B192-01M/B192-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogate was added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

Client : EUROFINS EATON ANALYTICAL
 Project : 988199
 SDG NO. : 22B192
 Instrument ID : GCT039

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis Date/Time	Extraction Date/Time	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
202202170722	VG39B11B	1	NA	02/18/2216:21	02/18/2216:21	EB18005A	EB18003A	22VG39B11	Method Blank
202202170722MS	VG39B11L	1	NA	02/18/2216:57	02/18/2216:57	EB18006A	EB18003A	22VG39B11	Lab Control Sample (LCS)
202202170723	VG39B11C	1	NA	02/18/2217:34	02/18/2217:34	EB18007A	EB18003A	22VG39B11	LCS Duplicate
202202170724	B192-01	1	NA	02/18/2218:10	02/18/2218:10	EB18008A	EB18003A	22VG39B11	Field Sample
202202170725	B192-01M	1	NA	02/18/2218:46	02/18/2218:46	EB18009A	EB18003A	22VG39B11	Matrix Spike Sample (MS)
	B192-01S	1	NA	02/18/2219:23	02/18/2219:23	EB18010A	EB18003A	22VG39B11	MS Duplicate (MSD)
	B192-02	1	NA	02/18/2219:59	02/18/2219:59	EB18011A	EB18003A	22VG39B11	Field Sample
	B192-03	1	NA	02/18/2220:35	02/18/2220:35	EB18012A	EB18003A	22VG39B11	Field Sample
	B192-04	1	NA	02/18/2221:12	02/18/2221:12	EB18013A	EB18003A	22VG39B11	Field Sample

FN - Filename
 % Moist - Percent Moisture

SAMPLE RESULTS

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client       : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project      : 988199                      Date Received: 02/18/22
Batch No.    : 22B192                      Date Extracted: 02/18/22 18:10
Sample ID    : 202202170722                Date Analyzed: 02/18/22 18:10
Lab Samp ID  : B192-01                     Dilution Factor: 1
Lab File ID  : EB18008A                    Matrix: WATER
Ext Btch ID  : 22VG39B11                   % Moisture: NA
Calib. Ref. : EB18003A                     Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)		
GASOLINE	ND	0.020	0.010		
SURROGATE PARAMETERS					
	RESULT	SPK_AMT	%RECOVERY	QC LIMIT	
Bromofluorobenzene	0.0346	0.0400	86	60-140	

Notes:

Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project     : 988199                     Date Received: 02/18/22
Batch No.   : 22B192                     Date Extracted: 02/18/22 19:59
Sample ID   : 202202170723              Date Analyzed: 02/18/22 19:59
Lab Samp ID : B192-02                   Dilution Factor: 1
Lab File ID : EB18011A                  Matrix: WATER
Ext Btch ID : 22VG39B11                 % Moisture: NA
Calib. Ref. : EB18003A                  Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)		
GASOLINE	ND	0.020	0.010		
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT	
Bromofluorobenzene	0.0329	0.0400	82	60-140	

Notes:

Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

QC SUMMARIES

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/18/22 16:21
Project     : 988199                     Date Received: 02/18/22
Batch No.   : 22B192                     Date Extracted: 02/18/22 16:21
Sample ID   : MBLK1W                    Date Analyzed: 02/18/22 16:21
Lab Samp ID : VG39B11B                  Dilution Factor: 1
Lab File ID : EB18005A                  Matrix: WATER
Ext Btch ID: 22VG39B11                  % Moisture: NA
Calib. Ref.: EB18003A                   Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
GASOLINE	ND	0.020	0.010	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0338	0.0400	85	60-140

Notes:

Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 988199
BATCH NO. : 22B192
METHOD : 5030B/8015B

MATRIX : WATER		% MOISTURE:NA
DILUTION FACTOR: 1	1	1
SAMPLE ID : MBLK1W	LCS1W	LCD1W
LAB SAMPLE ID : VG39B11B	VG39B11L	VG39B11C
LAB FILE ID : EB18005A	EB18006A	EB18007A
DATE PREPARED : 02/18/22 16:21	02/18/22 16:57	02/18/22 17:34
DATE ANALYZED : 02/18/22 16:21	02/18/22 16:57	02/18/22 17:34
PREP BATCH : 22VG39B11	22VG39B11	22VG39B11
CALIBRATION REF: EB18003A	EB18003A	EB18003A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.531	106	0.500	0.514	103	3	60-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0458	115	0.0400	0.0447	112	70-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 988199
BATCH NO. : 22B192
METHOD : 5030B/8015B

MATRIX : WATER		% MOISTURE:NA
DILUTION FACTOR: 1	1	1
SAMPLE ID : 202202170722	202202170722MS	202202170722MSD
LAB SAMPLE ID : B192-01	B192-01M	B192-01S
LAB FILE ID : EB18008A	EB18009A	EB18010A
DATE PREPARED : 02/18/22 18:10	02/18/22 18:46	02/18/22 19:23
DATE ANALYZED : 02/18/22 18:10	02/18/22 18:46	02/18/22 19:23
PREP BATCH : 22VG39B11	22VG39B11	22VG39B11
CALIBRATION REF: EB18003A	EB18003A	EB18003A

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.529	106	0.500	0.502	100	5	50-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0456	114	0.0400	0.0427	107	60-140

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

988199

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG#: 22B192

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 988199

SDG : 22B192

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

A total of two(2) water samples were received on 02/18/22 to be analyzed for Total Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSB027WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for Diesel was within LCS QC limits in DSB027WL. Refer to LCS summary form for details.

Matrix QC Sample

No matrix QC sample was provided on this SDG. One(1) set of MS/MSD was analyzed. Diesel was within MS QC limits in 22B177-01M/22B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 988199

SDG : 22B192

METHOD 3520C/8015B
PETROLEUM HYDROCARBONS BY EXTRACTION

A total of two (2) water samples were received on 02/18/22 to be analyzed for Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one (1) method blank was analyzed. DSB027WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one (1) LCS was analyzed. Percent recovery for JP5 was within LCS QC limits in J5B027WL. Refer to LCS summary form for details.

Matrix QC Sample

No matrix QC sample was provided on this SDG. One (1) set of MS/MSD was analyzed. JP5 was within MS QC limits in 22B177-01M/22B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 988199

SDG : 22B192

METHOD 3520C/8015B
PETROLEUM HYDROCARBONS BY EXTRACTION

A total of two(2) water samples were received on 02/18/22 to be analyzed for Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSB027WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for JP8 was within LCS QC limits in J8B027WL. Refer to LCS summary form for details.

Matrix QC Sample

No matrix QC sample was provided on this SDG. One(1) set of MS/MSD was analyzed. JP8 was within MS QC limits in 22B177-01M/22B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG NO. : 22B192
Instrument ID : D5

Client : EUROFINS EATON ANALYTICAL
Project : 988199

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
MBLK1W	DSB027WB	1	NA	02/22/2221:14	02/21/2210:30	LB22012A	LB22006A	22DSB027W	Method Blank
LCS1W	DSB027WL	1	NA	02/22/2221:32	02/21/2210:30	LB22013A	LB22006A	22DSB027W	Lab Control Sample (LCS)
202202170722	B192-01	1	NA	02/23/2202:10	02/21/2210:30	LB22028A	LB22006A	22DSB027W	Field Sample
202202170724	B192-03	1	NA	02/23/2202:28	02/21/2210:30	LB22029A	LB22006A	22DSB027W	Field Sample

FN - Filename
% Moist - Percent Moisture

LAB CHRONICLE
PETROLEUM HYDROCARBONS BY EXTRACTION

SDG NO. : 22B192
Instrument ID : D5

Client : EUROFINS EATON ANALYTICAL
Project : 988199

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Notes
MBLK1W	DSB027WB	1	NA	02/22/2221:14	02/21/2210:30	LB22012A	LB22007A	22DSB027W Method Blank
LCS1W	J5B027WL	1	NA	02/22/2221:51	02/21/2210:30	LB22014A	LB22007A	22DSB027W Lab Control Sample (LCS)
202202170722	B192-01	1	NA	02/23/2202:10	02/21/2210:30	LB22028A	LB22007A	22DSB027W Field Sample
202202170724	B192-03	1	NA	02/23/2202:28	02/21/2210:30	LB22029A	LB22007A	22DSB027W Field Sample

FN - Filename
% Moist - Percent Moisture

LAB CHRONICLE
 PETROLEUM HYDROCARBONS BY EXTRACTION

Client : EUROFINS EATON ANALYTICAL
 Project : 988199

SGJ NO. : 22B192
 Instrument ID : D5

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Notes
MBLK1W	DS8027WB	1	NA	02/22/2221:14	02/21/2210:30	LB22012A	LB22008A	22DSB027W Method Blank
LCS1W	J88027WL	1	NA	02/22/2222:09	02/21/2210:30	LB22015A	LB22008A	22DSB027W Lab Control Sample (LCS)
202202170722	B192-01	1	NA	02/23/2202:10	02/21/2210:30	LB22028A	LB22008A	22DSB027W Field Sample
202202170724	B192-03	1	NA	02/23/2202:28	02/21/2210:30	LB22029A	LB22008A	22DSB027W Field Sample

FN - Filename
 % Moist - Percent Moisture

SAMPLE RESULTS

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

=====
Client : EUROFINS EATON ANALYTICAL Date Collected: 02/15/22 09:15
Project : 988199 Date Received: 02/18/22
Batch No. : 22B192 Date Extracted: 02/21/22 10:30
Sample ID : 202202170722 Date Analyzed: 02/23/22 02:10
Lab Samp ID: 22B192-01 Dilution Factor: 1
Lab File ID: LB22028A Matrix: WATER
Ext Btch ID: 22DSB027W % Moisture: NA
Calib. Ref.: LB22006A Instrument ID: D5
=====

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.028	0.014	
Motor Oil	ND	0.055	0.028	

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.504	0.550	92	60-130
Hexacosane	0.127	0.158	92	60-130

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 910ml Final Volume : 5ml
Prepared by : P0reto Analyzed by : SDeeso

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project     : 988199                     Date Received: 02/18/22
Batch No.   : 22B192                     Date Extracted: 02/21/22 10:30
Sample ID   : 202202170722              Date Analyzed: 02/23/22 02:10
Lab Samp ID: 22B192-01                  Dilution Factor: 1
Lab File ID: LB22028A                   Matrix: WATER
Ext Btch ID: 22DSB027W                  % Moisture: NA
Calib. Ref.: LB22007A                   Instrument ID: D5
=====
    
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
JP5	ND	0.055	0.028

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.504	0.550	92	60-130
Hexacosane	0.127	0.138	92	60-130

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP5 C8-C18
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
 Sample Amount : 910ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project    : 988199                      Date Received: 02/18/22
Batch No.  : 22B192                      Date Extracted: 02/21/22 10:30
Sample ID  : 202202170722               Date Analyzed: 02/23/22 02:10
Lab Samp ID: 22B192-01                  Dilution Factor: 1
Lab File ID: LB22028A                   Matrix: WATER
Ext Btch ID: 22DSB027W                  % Moisture: NA
Calib. Ref.: LB22008A                   Instrument ID: D5
=====
  
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
JP8	ND	0.055	0.028

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.504	0.550	92	60-130
Hexacosane	0.127	0.138	92	60-130

Notes:

RL : Reporting Limit

Parameter H-C Range

JP8 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 910ml

Final Volume : 5ml

Prepared by : POrto

Analyzed by : SDeeso

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project    : 988199                      Date Received: 02/18/22
Batch No.  : 22B192                      Date Extracted: 02/21/22 10:30
Sample ID  : 202202170724               Date Analyzed: 02/23/22 02:28
Lab Samp ID: 22B192-03                  Dilution Factor: 1
Lab File ID: LB22029A                   Matrix: WATER
Ext Btch ID: 22DSB027W                  % Moisture: NA
Calib. Ref.: LB22006A                   Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.027	0.014	
Motor Oil	ND	0.055	0.027	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.530	0.545	97	60-130
Hexacosane	0.128	0.136	94	60-130

Notes:
Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 920ml Final Volume : 5ml
Prepared by : P0reto Analyzed by : SDeeso

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project    : 988199                      Date Received: 02/18/22
Batch No.  : 22B192                      Date Extracted: 02/21/22 10:30
Sample ID  : 202202170724                Date Analyzed: 02/23/22 02:28
Lab Samp ID: 22B192-03                   Dilution Factor: 1
Lab File ID: LB22029A                    Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22007A                    Instrument ID: D5
=====
    
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
JP5	ND	0.055	0.027	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.530	0.545	97	60-130
Hexacosane	0.128	0.136	94	60-130

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP5 C8-C18
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
 Sample Amount : 920ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/15/22 09:15
Project     : 988199                     Date Received: 02/18/22
Batch No.   : 22B192                     Date Extracted: 02/21/22 10:30
Sample ID   : 202202170724              Date Analyzed: 02/23/22 02:28
Lab Samp ID: 22B192-03                   Dilution Factor: 1
Lab File ID: LB22029A                    Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22008A                    Instrument ID: D5
=====
    
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
JP8	ND	0.055	0.027

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.530	0.545	97	60 130
Hexacosane	0.128	0.136	94	60-130

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP8 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 920ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

QC SUMMARIES

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```
=====
Client      : EUROFINS EATON ANALYTICAL    Date Collected: 02/21/22 10:30
Project     : 988199                      Date Received: 02/21/22
Batch No.   : 22B192                      Date Extracted: 02/21/22 10:30
Sample ID   : MBLK1W                      Date Analyzed: 02/22/22 21:14
Lab Samp ID: DSB027WB                    Dilution Factor: 1
Lab File ID: LB22012A                   Matrix: WATER
Ext Btch ID: 22DSB027W                  % Moisture: NA
Calib. Ref.: LB22006A                   Instrument ID: D5
=====
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.025	0.012	
Motor Oil	ND	0.050	0.025	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.384	0.500	77	60-130
Hexacosane	0.110	0.125	88	60-130

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml Final Volume : 5ml
Prepared by : POrto Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 988199
BATCH NO. : 22B192
METHOD : 3520C/8015B

=====

MATRIX	: WATER	% MOISTURE:NA
DILUTION FACTOR:	1	1
SAMPLE ID	: MBLK1W	LCS1W
LAB SAMPLE ID	: DSB027WB	DSB027WL
LAB FILE ID	: LB22012A	LB22013A
DATE PREPARED	: 02/21/22 10:30	02/21/22 10:30
DATE ANALYZED	: 02/22/22 21:14	02/22/22 21:32
PREP BATCH	: 22DSB027W	22DSB027W
CALIBRATION REF:	LB22006A	LB22006A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----	-----
Diesel	ND	2.50	2.72	109	50-130

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----
Bromobenzene	0.500	0.567	113	60-130
Hexacosane	0.125	0.130	104	60-130

=====

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202202160931                       202202160931MSD
LAB SAMPLE ID : 22B177-01                         22B177-01S
LAB FILE ID  : LB22017A                           LB22019A
DATE PREPARED : 02/21/22 10:30                    02/21/22 10:30
DATE ANALYZED : 02/22/22 22:46                    02/22/22 23:23
PREP BATCH   : 22DSB027W                           22DSB027W
CALIBRATION REF: LB22006A                           LB22006A
=====

```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Diesel	ND	2.50	2.89	116	2.55	3.16	124	9	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.500	0.478	96	0.510	0.512	100	60-130
Hexacosane	0.125	0.122	98	0.127	0.132	104	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/21/22 10:30
Project    : 988199                       Date Received: 02/21/22
Batch No.  : 22B192                       Date Extracted: 02/21/22 10:30
Sample ID  : MBLK1W                       Date Analyzed: 02/22/22 21:14
Lab Samp ID: DSB027WB                     Dilution Factor: 1
Lab File ID: LB22012A                     Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22007A                    Instrument ID: D5
=====
  
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
JP5	ND	0.050	0.025	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.384	0.500	77	60-130
Hexacosane	0.110	0.125	88	60-130

Notes:

RL : Reporting Limit

Parameter H-C Range

JP5 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml

Final Volume : 5ml

Prepared by : POrreto

Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 988199
BATCH NO. : 22B192
METHOD : 3520C/8015B

MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : MBLK1W LCS1W
LAB SAMPLE ID : DSB027WB J5B027WL
LAB FILE ID : LB22012A LB22014A
DATE PREPARED : 02/21/22 10:30 02/21/22 10:30
DATE ANALYZED : 02/22/22 21:14 02/22/22 21:51
PREP BATCH : 22DSB027W 22DSB027W
CALIBRATION REF: LB22007A LB22007A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
JP5	ND	2.50	2.25	90	30-160

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
Bromobenzene	0.500	0.450	90	60-130
Hexacosane	0.125	0.119	95	60-130

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                                % MOISTURE:NA
DILUTION FACTOR: 1                                1
SAMPLE ID   : 202202160931                        202202160931MSD
LAB SAMPLE ID : 22B177-01                          22B177-01S
LAB FILE ID  : LB22017A                            LB22020A
DATE PREPARED : 02/21/22 10:30                    02/21/22 10:30
DATE ANALYZED : 02/22/22 22:46                    02/23/22 00:00
PREP BATCH   : 22DSB027W                          22DSB027W
CALIBRATION REF: LB22007A                          LB22007A
=====

```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
JP5	ND	2.53	2.69	107	2.53	2.59	103	4	30-160	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.505	0.495	98	0.505	0.501	99	60-130
Hexacosane	0.126	0.121	96	0.126	0.120	95	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/21/22 10:30
Project     : 988199                     Date Received: 02/21/22
Batch No.   : 22B192                     Date Extracted: 02/21/22 10:30
Sample ID   : MBLK1W                     Date Analyzed: 02/22/22 21:14
Lab Samp ID: DSB027WB                    Dilution Factor: 1
Lab File ID: LB22012A                    Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22008A                    Instrument ID: D5
=====
  
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
JP8	ND	0.050	0.025

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.384	0.500	77	60-130
Hexacosane	0.110	0.125	88	60-130

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP8 C8-C18
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
 Sample Amount : 1000ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 988199
BATCH NO. : 22B192
METHOD : 3520C/8015B

MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : MBLK1W LCS1W
LAB SAMPLE ID : DSB027WB J8B027WL
LAB FILE ID : LB22012A LB22015A
DATE PREPARED : 02/21/22 10:30 02/21/22 10:30
DATE ANALYZED : 02/22/22 21:14 02/22/22 22:09
PREP BATCH : 22DSB027W 22DSB027W
CALIBRATION REF: LB22008A LB22008A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
JP8	ND	2.50	2.21	88	30-160

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
Bromobenzene	0.500	0.528	106	60-130
Hexacosane	0.125	0.122	98	60-130

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202202160931                       202202160931MSD
LAB SAMPLE ID : 22B177-01                         22B177-01S
LAB FILE ID  : LB22017A                           LB22022A
DATE PREPARED : 02/21/22 10:30                   02/21/22 10:30
DATE ANALYZED : 02/22/22 22:46                   02/23/22 00:38
PREP BATCH   : 22DSB027W                         22DSB027W
CALIBRATION REF: LB22008A                        LB22008A
=====

```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
JP8	ND	2.62	2.58	98	2.62	2.26	86	13	30-160	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.525	0.563	107	0.525	0.491	94	60-130
Hexacosane	0.131	0.128	98	0.131	0.123	94	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate