


750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Report

for

Honolulu Board of Water Supply
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843
Attention: Erwin Kawata
Fax: 808-550-5018



Utah ELCP CA00006

DEB: Debbie L Frank
Project Manager

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022) - EMAX

* Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.

* Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis.

* As applicable, this report consists of the cover page, State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms.

* Test results relate only to the sample(s) tested.

* Test results apply to the sample(s) as received, unless otherwise noted in the comments report (ISO/IEC 17025:2017).

* This report shall not be reproduced except in full, without the written approval of the laboratory.

* This report includes ISO/IEC 17025 and non-ISO 17025 accredited methods.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Montana	Cert 0035
Arizona	AZ0778	Nebraska	NE-OS-21-13
Arkansas	CA00006	Nevada	CA00006
California	2813	New Hampshire *	2959
Colorado	CA00006	New Jersey *	CA 008
Connecticut	PH-0107	New Mexico	CA00006
Delaware	CA 006	New York *	11320
Florida *	E871024	North Carolina	06701
Georgia	947	North Dakota	R-009
Guam	21-008R	Ohio - 537.1	87786
Hawaii	CA00006	Oregon *	4034
Idaho	CA00006	Pennsylvania *	68-00565
Illinois	200033	Puerto Rico	CA00006
Indiana	C-CA-01	Rhode Island	LAO00326
Iowa – Asbestos	413	South Carolina	87016
Kansas *	E-10268	South Dakota	CA11320
Kentucky	90107	Tennessee	TN02839
Louisiana *	LA008	Texas *	T104704230-20-18
Maine	CA00006	Utah (Primary AB) *	CA00006
Maryland	224	Vermont	VT0114
Marianas Islands	MP0004	Virginia *	460260
Massachusetts	M-CA006	Washington	C838
Michigan	9906	EPA Region 5	CA00006
Mississippi	CA00006	Los Angeles County Sanitation Districts	10264

* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025:2017 Accredited Method List

The test listed below are accredited and met the requirements of ISO/IEC 17025 as verify by A2LA.

Refer to our certificates and scope of accreditations (no. 5890-1 and 5890-2) found at:

<https://www.eurofinsus.com/Eaton>

Test(s)	Method(s)	Potable Water *	Waste Water
Enterococci	Enterolert	x	x
Escherichia coli (Enumeration)	SM 9221 B.1 SM 9221 F	x	
Fecal Coliform (P/A and Enumeration)	SM 9221 C (MTF/EC), SM 9221 E (MTF/EC)	x	x
Fecal Streptococci and Enterococci	SM 9230 B	x	x
Heterotrophic Bacteria	SM 9215 B	x	
Legionella	Legiolert®	x	
Pseudomonas aeruginosa	Idexx Pseudalart	x	
Total Coliform (P/A and Enumeration)	SM 9221A, SM 9221B, SM 9221 C	x	x
Total Coliform, Total Coliform with Chlorine Present	SM 9221 B	x	x
Total Coliform/E. coli (P/A and Enumeration, Idexx Colilert, Idexx Colilert 18, Colisure)	SM 9223	x	
Total Microcystins and Nodularins	EPA 546	X	
Yeast and Mold	SM 9610	x	
1,2,3-Trichloropropane (TCP) at 5 PPT	CA SRL 524M-TCP	x	
1,4-Dioxane	EPA 522	x	
2,3,7,8-TCDD	Modified EPA 1613 B	x	
Acrylamide	+ LCMS 2440)	x	
Algal Toxins/Microcys in	+ LCMS 3570	x	
Alkalinity	SM 2320B	x	x
Ammonia	EPA 350.1, SM 4500-NH3 H		x
Asbestos	EPA 100.2	x	x
Bicarbonate Alkalinity as HCO3	SM 2330 B	x	x
BOD/CBOD	SM 5210 B		x
Bromate	+ LCMS- 2447	x	
Carbonate as CO3	SM 2330 B	x	x
Carbonyls	EPA 556	x	x
Chemical Oxygen Demand	EPA 410.4, SM 5220D		x
Chlorinated Acids	EPA 515.4	x	
Chlorine Dioxide	Palin Test Chlordio X Plus, SM 4500-CLO2 D	x	
Chlorine, Free, Combined, Total Residual, Chloramines	SM 4500-Cl G	x	
Color	SM2120B	x	
Conductivity	EPA 120.1, SM 2510B	x	x
Corrosivity (Langelier Index), Carbonate as CO3, Hydroxide as OH Calculated	SM 2330 B	x	
Cyanide (Amenable)	SM 4500-CN G	x	x
Cyanide (Free)	SM 4500CN F	x	x
Cyanide (Total)	EPA 335.4	x	x
Cyanogen Chloride (Screen)	+ 335 Mod (WC-24467)	x	
Diquat and Paraquat	EPA 549.2	x	
DBP and HAA	SM 6251 B	x	
Dissolved Organic Carbon	SM 5310 C	x	
Dissolved Oxygen	SM 4500-O G		x
EDB/DCBP/TCP	EPA 504.1	x	
EDB/DBCP and Disinfection Byproducts	EPA 551.1	x	
EDTA and NTA	+ WC-2454	x	
Endothall	EPA 548.1, +(LCMS-2445)	x	
Fluoride	SM 4500F C	x	x
Glyphosate	EPA 547	x	
Glyphosate and AMPA	+ LCMS-3618	x	
Gross Alpha and Gross Beta	EPA 900.0	x	x

Test(s)	Method(s)	Potable Water *	Waste Water
Gross Alpha coprecipitation	SM 7110 C	x	x
Hardness	SM 2340 B	x	x
Hexavalent Chromium	EPA 218.6,	x	x
Hexavalent Chromium	EPA 218.7,	x	
Hexavalent Chromium	SM 3500-Cr B		x
Inorganic Anions and DBPs	EPA 300.0	x	x
Norganic Anions and DBPs	EPA 300.1	x	
Kjeldahl Nitrogen	EPA 351.2		x
Metals	EPA 200.7, EPA200.8	x	x
Nitrosamines	EEA-Agilent 521.1 (GCMS-24250)	x	
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x
Odor	SM2150B	x	
Organohalide Pesticides and PCB	EPA 505	x	
Ortho Phosphate	SM 4500P E	x	
Oxyhalides Disinfect ion Byproducts	EPA 317.0	x	
Perchlorate	EPA 331.0	x	
Perchlorate (Low and High Levels)	EPA 314.0	x	
Perfluorinated Alkyl Acids	EPA 533, EPA 537, EPA 537.1	x	
PPCP and EDC	+ LCMS-2443	x	
pH	EPA 150.1 SM 4500-H+ B	x	x
Phenolics – Low Level	+WC 2493 (EPA 420.2 and EPA 420.4 MOD)	x	x
Phenylurea Pesticides/Herbicides	+ LCMS-2448	x	
Radium-226, Radium-228	GA Tech (Rad-2374)	x	
Radon-222	SM 7500RN	x	
Residue (Filterable)	SM 2540C	x	x
Residue (Non-Filterable)	SM 2540D		x
Residue (Total)	SM 2540B		x
Residue (Volatile)	EPA 160.4		x
Semi-Volatile Compounds	EPA 525.2	x	
Silica	SM 4500-SiO2 C	x	x
Sulfide	SM 4500-S D		x
Sulfite	SM 4500-SO3 B	x	x
Surfactants	SM 5540C	x	x
Taste and Odor	SM 6040 E	x	
Total Organic Carbon	SM 5310 C	x	x
Total Phenols	EPA 420.1		x
Total Phenols	EPA 420.4	x	x
Triazine Pesticides and their Degradates	+ LCMS-3617	x	
Turbidity	EPA 180.1	x	x
Uranium by ICP/MS	EPA 200.8	x	
UV 254 Organic Constituents	SM 5910B	x	
VOCs	EPA 524.2	x	
VOCs	+(GCMS 2412) by EPA 524.2 modified	x	

(*) includes: Bottled Water, Drinking Water and Water as Component of Food & Beverage.

(+) In-House Method

Acknowledgement of Samples Received

Addr: **Honolulu Board of Water Supply**
 630 South Beretania Street
 Public Service Bldg." Room 308
 Honolulu, HI 96843

Attn: Erwin Kawata
 Phone: 808-748-5091

Client ID: HONOLULU
 Folder #: 1007133
 Project: RED-HILL
 Sample Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX
 Project Manager: Debbie L Frank
 Phone: (626) 386-1149
 PO #: C20525101 exp 05312023

The following samples were received from you on **June 02, 2022 at 1012**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
<u>202206030089</u>	MOANALUA WELLS (331-223-TP202)	05/31/2022 1029
	SDWIS PWSID: HI0000331 SDWIS FACILITY ID: TP202 SDWIS SAMPLE POINT ID: 223	
	@525PLUS C PLUS TICS (SUB)Gas Fraction Hydrocarbons TPH 8015 Diesel and Motor Oil	
<u>202206030090</u>	TB:MOANALUA WELLS (331-223-TP202)	05/31/2022 1029
	(SUB)Gas Fraction Hydrocarbons	

Test Description

@525PLUS C PLUS TICS -- Semivolatiles by GCMS



Kit Order for BOARD OF WATER SUPPLY, CITY AND COUNTY OF

Debbie L Frank is your Eurofins Eaton Analytical, LLC Service Manager

Created Date & Time: 4/27/2022 3:15:49PM

750 Royal Oaks Drive, Suite 100
 Monrovia, California 91016-3629
 (626) 386-1100 FAX (866) 988-3757

Note: Sampler Please return this paper with your samples

Client ID: HONOLULU
 Project Code: RED-HILL Bottle Orders
 Group Name: Weekly TPH-8015_RED-HILL (2022) - EMAX
 PO#/JOB#: C20525101 exp 05312023
 Description: Every 1 week on Wed

Kit #: 318557
 Created By: - [AutoGenerated]
 Deliver By: 05/04/2022
 STG: Bottle Orders
 Ice Type: W

Ship Sample Kits to
 Honolulu Board of Water Supply
 630 South Beretania Street
 Chemistry Lab
 Honolulu, HI 96843
 Attn: Ron Fenstermacher
 Phone: 808-748-5841
 Fax: 808-550-5572

Send Report to
 Honolulu Board of Water Supply
 630 South Beretania Street
 Public Service Bldg. Room 308
 Honolulu, HI 96843
 Attn: Erwin Kawata
 Phone: 808-748-5091
 Fax: 808-550-5018

Billing Address
 Honolulu Board of Water Supply
 630 South Beretania Street
 Public Service Bldg. Room 308
 Honolulu, HI 96843
 Attn: Erwin Kawata
 Phone: 808-748-5091
 Fax: 808-550-5018

# of	Sample Tests	Bottle Qty - Type [preservative information]	Total	UN DOT #
6	TPH 8015 Diesel and Motor Oil_C	✓ 3 - 1L amber glass [1 ml Thio 8%]	18	
5	@525PLUS C PLUS TICS	2 - 1L amber glass [45mg Sulfite xls+1 vial 2 ml 6N HCl]	10	UN1789
1	@525PLUS C PLUS TICS	✓ 3 - 1L amber glass [45mg Sulfite xls+1 vial 2 ml 6N HCl]	3	UN1789
6	8015 Gas_C	✓ 4 - 40ml amber glass vial [1 drop Thio (8%)]	24	
6	8015 Gas_C TB	✓ 2 - 40ml amber glass vial [1 drop Thio (8%) + H2O]	12	
Sum Tests: 24			Sum Bottles: 67	

Comments

WEEKLY SAMPLING

5/31/2022

Moanalua Wells



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EBA Folder Number: 102137

SAMPLE TEMP RECEIVED:
Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = 649A (Observation = 4.6 °C) (Corr.Factor = 0.3 °C) (Final = 4.3 °C)

TYPE OF ICE: Real Synthetic No ice CONDITION OF ICE: Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In FedEx UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrant

1 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

4) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

6) pH Check, Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____

6) Chlorine check, Manufacturer: Sansafe, Lot No.: _____ Expiration Date: _____ Results: _____

7) VOA and Radon Headspace: No Samples with Headspace: Samples with Headspace (see below): _____

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)
Exempt from headspace concerns: Methods 816.4, HAA(8281,892), 805, SPME, @CH, 832LCMS, 888, 838, Anatoxin, LCMS methods using 40 ml vials, International clients:

Sample ID	Bottle #	None/≤8	>8mm	Test	Sample ID	Bottle #	None/≤8	>8mm	Test	Sample ID	Bottle #	None/≤8	>8mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): _____

RECEIVED BY: [Signature] SIGNATURE PRINT NAME: G. REITER COMPANY/TITLE: Eurofins Eaton Analytical DATE: 06/02/22 TIME: 10:12

SAMPLES CHECKED AGAINST COC BY: _____ SIGNATURE PRINT NAME: _____ COMPANY/TITLE: Eurofins Eaton Analytical DATE: _____ TIME: _____

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: 165173

SAMPLE TEMP RECEIVED:

Notes: If samples are out of temperature range, let the ASMA know. ASMA will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = 649A (Observation = 4.3 °C) (Corr. Factor = -0.3 °C) (Final = 4.0 °C)

TYPE OF ICE: Real Synthetic No Ice CONDITION OF ICE: Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-in / FedEX UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤ 8°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥ 10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation °C) (Corr. Factor °C) (Final °C)	2 = (Observation °C) (Corr. Factor °C) (Final °C)
3 = (Observation °C) (Corr. Factor °C) (Final °C)	4 = (Observation °C) (Corr. Factor °C) (Final °C)

4) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check, Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____

6) Chlorine check, Manufacturer: Samsafe. Lot No.: _____ Expiration Date: _____ Results: _____

7) VOA and Radon Headspace: No Samples with Headspace: Samples with Headspace (see below):

Headspace Documentation (use additional VOC and Radon Internal COC for additional bottles)

Exmpl from headspace concern: Methods #15.4, HAA(0251,022), 005, SPME, @CH, 032LCMS, 006, 008, Anstoxin, LCMS methods using 40 ml vials, international client

Sampl ID Bottle #	None/<8	>8mm	Test	Sampl ID Bottle #	None/<8	>8mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors):

RECEIVED BY: <u>[Signature]</u>	PRINT NAME: <u>G. REITNER</u>	COMPANY/TITLE: <u>Eurofins Eaton Analytical</u>	DATE: <u>06/02/22</u>	TIME: <u>10:12</u>
SAMPLES CHECKED AGAINST GOOD BY: <u>[Signature]</u>	PRINT NAME: <u>[Signature]</u>	COMPANY/TITLE: <u>Eurofins Eaton Analytical</u>	DATE: _____	TIME: _____

INTERNAL CHAIN OF CUSTODY RECORD

Eaton Analytical

EEA Folder Number 108877

SAMPLE TEMP RECEIVED:

Notes: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = 649A (Observation = 4.2 °C) (Corr. Factor -0.3 °C) (Final = 3.9 °C)

TYPE OF ICE: Real Synthetic No Ice CONDITION OF ICE: Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / **FedEx** / UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: <10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: <10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = °C) (Corr. Factor = °C) (Final = °C)	2 = (Observation = °C) (Corr. Factor = °C) (Final = °C)
3 = (Observation = °C) (Corr. Factor = °C) (Final = °C)	4 = (Observation = °C) (Corr. Factor = °C) (Final = °C)

4) Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check, Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____

6) Chlorine check, Manufacturer: Samsafe, Lot No.: _____ Expiration Date: _____ Results: _____

7) VOA and Radon Headspace: No Samples with Headspace; Samples with Headspace (see below): _____

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Exempt from headspace concerns: Methods 815.4, HAA(825,852), 805, SPME, @CH, 832LCMS, 888, 898, Anatoxin, LCMS methods using 40 ml vials, International clients:

Samp. ID	Bottle #	None/<8	>8mm	Test	Samp ID	Bottle #	None/<8	>8mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors):


RECEIVED BY: <u>[Signature]</u>	PRINT NAME: <u>G. RETNER</u>	COMPANY/TITLE: <u>Eurofins Eaton Analytical</u>	DATE: <u>06/02/22</u>	TIME: <u>10:12</u>
SAMPLES CHECKED AGAINST GOODBY!	PRINT NAME: _____	COMPANY/TITLE: <u>Eurofins Eaton Analytical</u>	DATE: _____	TIME: _____

ORIGIN ID: HIKA (808) 748-5840
 BWS CHEM LAB
 HONOLULU BOARD OF WATER SUPPLY
 630 S. BERETANIA ST.
 CHEMICAL LABORATORY
 HONOLULU, HI 96843
 UNITED STATES US

SHIP DATE: 01 JUN 22
 ACT WGT: 62.00 LB
 CAD: 100205419/INET 4490


TO BROOKS
 EUROFINS EATON ANALYTICAL, INC
 750 ROYAL OAKS DR
 SUITE 100
 MONROVIA CA 91016
 (626) 386-1178
 REF: INV. PO. DEPT.

577J2274FF4A



1 of 3
 TRK# 7770 1457 6277
 ## MASTER ##
WZ WHPA
 91016
 CA-US BUR

THU - 02 JUN 10:30A
 PRIORITY OVERNIGHT



After printing this label:
 1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
 2. Fold the printed page along the horizontal line.
 3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.
 Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.


ORIGIN ID:HIKA (808) 748-5840
BWS CHEMILAB
HONOLULU BOARD OF WATER SUPPLY
630 S. BERETANIA ST.
CHEMICAL LABORATORY
HONOLULU, HI 96843
UNITED STATES US

SHIP DATE: 01 JUN 22
ACTWGT: 62.00 LB
CAD: 100205419/NET14490

BILL RECIPIENT

TO **BROOKS**
EUROFINS EATON ANALYTICAL, INC
750 ROYAL OAKS DR
SUITE 100
MONROVIA CA 91016
(626) 386-1178 REF:
INV. PO. DEPT:

577J2274FFEA4




THU - 02 JUN 10:30A
PRIORITY OVERNIGHT

2 of 3
MPS# 7770 1457 5650
Mstr# 7770 1457 6277

0201

91016
BUR
CA-US

WZ WHPA





After printing this label:
1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.
Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.
Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

ORIGIN ID:HIKA (808) 748-5840
 BWS CHEM LAB SHIP DATE: 01 JUN 22
 HONOLULU BOARD OF WATER SUPPLY ACTWGT: 62.00 LB
 630 S. BERETANIA ST. CAD: 100205419/NET4490
 CHEMICAL LABORATORY
 HONOLULU, HI 96843
 UNITED STATES US
 BILL RECIPIENT

TO **BROOKS**
EUROFINS EATON ANALYTICAL, INC
750 ROYAL OAKS DR
SUITE 100
MONROVIA CA 91016
 REF: (626) 386-1178


577J2274FF44

PO. INVT. DEPT.



J22202041201W

3 of 3
 MPS# 7770 1457 3955
 Mstr# 7770 1457 6277
WZ WHPA
 CA-US
91016
BUR
 THU - 02 JUN 10:30A
 PRIORITY OVERNIGHT



After printing this label:
 1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
 2. Fold the printed page along the horizontal line.
 3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.
 Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022)
- EMAX

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Folder Comments

Analytical results for TPH Gas, Diesel, and Mortor Oil are submitted by EMAX Laboratories, Torrance, CA

ND reporting (subcontract lab reports)

MDL is listed due to report format restrictions; it is not used in reporting. Analytical results reported as ND, are ND at the RL.

Tentatively Identified compounds (TIC).

The analyte has been "tentatively identified" as present and the associated numerical value is the estimated concentration in the sample. The analytes are not positively identified or quantified. Presentation of results in this report does not indicate actual presence of the compound identified in the TIC summary. Information is for study purposes only.

@525.2 (SVOC by GCMS)

Sample 202206030089: @525.2 TIC's: ND

(525.2-SVOC)L4

Caffeine is low bias, Caffeine is not a specific target for this testing. Not Reported.

Flags Legend:

L4 - The associated blank spike recovery was below method acceptance limits.

R7 - LFB/LFBD RPD exceeded the laboratory acceptance limit. Recovery met acceptance criteria.



Eaton Analytical

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Hits

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022)
- EMAX

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Samples Received on:
06/02/2022 1012

Analyzed	Analyte	Sample ID	Result	HI Limit	Units	MRL
----------	---------	-----------	--------	----------	-------	-----

SUMMARY OF POSITIVE DATA ONLY

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply
 Erwin Kawata
 630 South Beretania Street
 Public Service Bldg.™ Room 308
 Honolulu, HI 96843

Samples Received on:
 06/02/2022 1012

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution	
MOANALUA WELLS (331-223-TP202) (202206030089)						Sampled on 05/31/2022 1029				
Facility ID: TP202										
Sample Point ID: 223										
PWSID: HI0000331										
EPA 525.2 - Semivolatiles by GCMS										
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	1-Methylnaphthalene	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	2,4-DDD	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	2,4-DDE	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	2,4-DDT	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	2,4-Dinitrotoluene	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	2,6-Dinitrotoluene	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	2-methylnaphthalene	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	4,4-DDD	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	4,4-DDE	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	4,4-DDT	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Acenaphthene	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Acenaphthylene	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Acetochlor	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Alachlor	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Alpha-BHC	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	alpha-Chlordane	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Anthracene	ND	ug/L	0.020	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Atrazine	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Benz(a)Anthracene	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Benzo(a)pyrene	ND	ug/L	0.020	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Benzo(b)Fluoranthene	ND	ug/L	0.020	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Benzo(g,h,i)Perylene	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Benzo(k)Fluoranthene	ND	ug/L	0.020	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Beta-BHC	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Bromacil	ND (R7)	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Butachlor	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Butylbenzylphthalate	ND	ug/L	0.50	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Caffeine by method 525mod	NR (L4,R7)	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Chlorobenzilate	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Chloroneb	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Chlorothalonil(Draconil,Bravo)	ND	ug/L	0.10	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Chlorpyrifos (Dursban)	ND	ug/L	0.050	1
06/06/22	06/08/22	15:20	1411092	1411458	(EPA 525.2)	Chrysene	ND	ug/L	0.020	1

Rounding on totals after summation.
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply
 Erwin Kawata
 630 South Beretania Street
 Public Service Bldg.” Room 308
 Honolulu, HI 96843

Samples Received on:
 06/02/2022 1012

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Delta-BHC	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Di-(2-Ethylhexyl)adipate	ND	ug/L	0.60	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Di(2-Ethylhexyl)phthalate	ND	ug/L	0.60	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Diazinon (Qualitative)	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Dibenz(a,h)Anthracene	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Dichlorvos (DDVP)	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Dieldrin	ND	ug/L	0.20	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Diethylphthalate	ND	ug/L	0.50	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Dimethoate	ND (R7)	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Dimethylphthalate	ND	ug/L	0.50	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Di-n-Butylphthalate	ND	ug/L	1.0	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Di-n-octylphthalate	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Endosulfan I (Alpha)	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Endosulfan II (Beta)	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Endosulfan Sulfate	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Endrin	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Endrin Aldehyde	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	EPTC	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Fluoranthene	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Fluorene	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	gamma-Chlordane	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Heptachlor	ND	ug/L	0.040	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Heptachlor Epoxide (isomer B)	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Hexachlorobenzene	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Hexachlorocyclopentadiene	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Indeno(1,2,3,c,d)Pyrene	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Isophorone	ND	ug/L	0.50	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Lindane	ND	ug/L	0.040	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Malathion	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Methoxychlor	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Metolachlor	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Metribuzin	ND (R7)	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Molinatate	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Naphthalene	ND	ug/L	0.30	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Parathion	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Pendimethalin	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Permethrin (mixed isomers)	ND	ug/L	0.20	1

Rounding on totals after summation.
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply
 Erwin Kawata
 630 South Beretania Street
 Public Service Bldg." Room 308
 Honolulu, HI 96843

Samples Received on:
 06/02/2022 1012

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Phenanthrene	ND	ug/L	0.040	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Propachlor	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Pyrene	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Simazine	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Terbacil	ND (R7)	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Terbutylazine	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Thiobencarb (ELAP)	ND	ug/L	0.20	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	trans-Nonachlor	ND	ug/L	0.050	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Trifluralin	ND	ug/L	0.10	1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	1,3-Dimethyl-2-nitrobenzene	93	%		1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Acenaphthene-d10	78	%		1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Chrysene-d12	88	%		1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Perylene-d12	96	%		1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Phenanthrene-d10	84	%		1
06/06/22	06/08/22 15:20	1411092	1411458	(EPA 525.2)	Triphenylphosphate	112	%		1
SW 8015B - (SUB)Gas Fraction Hydrocarbons									
06/03/22	06/03/22 18:07			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1
SW 8015B - TPH 8015 Diesel and Motor Oil									
06/06/22	06/07/22 20:22			(SW 8015B)	TPH Diesel	ND	mg/L	0.024	1
06/06/22	06/07/22 20:22			(SW 8015B)	TPH Motor Oil	ND	mg/L	0.049	1
TB:MOANALUA WELLS (331-223-TP202) (202206030090)							Sampled on 05/31/2022 1029		
SW 8015B - (SUB)Gas Fraction Hydrocarbons									
06/03/22	06/03/22 18:43			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1

Rounding on totals after summation.
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.



Eaton Analytical

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory QC Summary

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022)
- EMAX

Honolulu Board of Water Supply

Semivolatiles by GCMS

Prep Batch: 1411092 Analytical Batch: 1411458

202206030089

MOANALUA WELLS (331-223-TP202)

Analysis Date: 06/08/2022

Analyzed by: JWC

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
Semivolatiles by GCMS by EPA 525.2									
Prep Batch: 1411092 Analytical Batch: 1411458					Analysis Date: 06/08/2022				
DUP_202205260805	1,3-Dimethyl-2-nitrobenzene (S)			94.0	%	94	(70-130)		
LCS1	1,3-Dimethyl-2-nitrobenzene (S)		5	93.6	%	94	(70-130)		
LC 2	1,3 Dimethyl 2 nitrobenzene ()		5	90.6	%	91	(70-130)		
MBLK	1,3-Dimethyl-2-nitrobenzene (S)			92.2	%	92	(70-130)		
MRL_CHK	1,3-Dimethyl-2-nitrobenzene (S)		5	94.2	%	94	(70-130)		
MS_202205260804	1,3-Dimethyl-2-nitrobenzene (S)		5	92.6	%	93	(70-130)		
DUP_202205260805	1-Methylnaphthalene			ND	ug/L		(0-20)		
LCS1	1-Methylnaphthalene		2	1.78	ug/L	89	(70-130)		
LCS2	1-Methylnaphthalene		2	1.77	ug/L	88	(70-130)	20	0.56
MBLK	1-Methylnaphthalene			<0.1	ug/L				
MRL_CHK	1-Methylnaphthalene		0.1	0.0960	ug/L	96	(50-150)		
MS_202205260804	1-Methylnaphthalene		2	1.81	ug/L	90	(70-130)		
DUP_202205260805	2,4-DDD			ND	ug/L		(0-20)		
LCS1	2,4-DDD		2	1.97	ug/L	99	(70-130)		
LCS2	2,4-DDD		2	2.08	ug/L	104	(70-130)	20	5.4
MBLK	2,4-DDD			<0.1	ug/L				
MRL_CHK	2,4-DDD		0.1	0.111	ug/L	111	(50-150)		
MS_202205260804	2,4-DDD		2	2.09	ug/L	105	(70-130)		
DUP_202205260805	2,4-DDE			ND	ug/L		(0-20)		
LCS1	2,4-DDE		2	1.98	ug/L	99	(70-130)		
LCS2	2,4-DDE		2	2.08	ug/L	104	(70-130)	20	4.9
MBLK	2,4-DDE			<0.1	ug/L				
MRL_CHK	2,4-DDE		0.1	0.104	ug/L	104	(50-150)		
MS_202205260804	2,4-DDE		2	2.03	ug/L	102	(70-130)		
DUP_202205260805	2,4-DDT			ND	ug/L		(0-20)		
LCS1	2,4-DDT		2	2.02	ug/L	101	(70-130)		
LCS2	2,4-DDT		2	2.14	ug/L	107	(70-130)	20	5.8
MBLK	2,4-DDT			<0.1	ug/L				
MRL_CHK	2,4-DDT		0.1	0.108	ug/L	108	(50-150)		
MS_202205260804	2,4-DDT		2	2.12	ug/L	106	(70-130)		
DUP_202205260805	2,4-Dinitrotoluene			ND	ug/L		(0-20)		
LCS1	2,4-Dinitrotoluene		2	1.64	ug/L	82	(70-130)		
LCS2	2,4-Dinitrotoluene		2	1.93	ug/L	97	(70-130)	20	16
MBLK	2,4-Dinitrotoluene			<0.1	ug/L				
MRL_CHK	2,4-Dinitrotoluene		0.1	0.102	ug/L	102	(50-150)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MS_202205260804	2,4-Dinitrotoluene		2	2.06	ug/L	103	(70-130)		
DUP_202205260805	2,6-Dinitrotoluene			ND	ug/L		(0-20)		
LCS1	2,6-Dinitrotoluene		2	1.77	ug/L	89	(70-130)		
LCS2	2,6-Dinitrotoluene		2	2.10	ug/L	105	(70-130)	20	17
MBLK	2,6-Dinitrotoluene			<0.1	ug/L				
MRL_CHK	2,6-Dinitrotoluene		0.1	0.0930	ug/L	93	(50-150)		
MS_202205260804	2,6-Dinitrotoluene		2	2.14	ug/L	107	(70-130)		
DUP_202205260805	2-methylnaphthalene			ND	ug/L		(0-20)		
LCS1	2-methylnaphthalene		2	1.90	ug/L	95	(70-130)		
LCS2	2-methylnaphthalene		2	1.87	ug/L	94	(70-130)	20	1.6
MBLK	2-methylnaphthalene			<0.1	ug/L				
MRL_CHK	2-methylnaphthalene		0.1	0.0990	ug/L	99	(50-150)		
MS_202205260804	2-methylnaphthalene		2	1.89	ug/L	95	(70-130)		
DUP_202205260805	4,4-DDD			ND	ug/L		(0-20)		
LCS1	4,4-DDD		2	2.06	ug/L	103	(70-130)		
LCS2	4,4-DDD		2	2.15	ug/L	108	(70-130)	20	4.3
MBLK	4,4-DDD			<0.1	ug/L				
MRL_CHK	4,4-DDD		0.1	0.103	ug/L	103	(50-150)		
MS_202205260804	4,4-DDD		2	2.16	ug/L	108	(70-130)		
DUP_202205260805	4,4-DDE			ND	ug/L		(0-20)		
LCS1	4,4-DDE		2	2.12	ug/L	106	(70-130)		
LCS2	4,4-DDE		2	2.26	ug/L	113	(70-130)	20	6.4
MBLK	4,4-DDE			<0.1	ug/L				
MRL_CHK	4,4-DDE		0.1	0.105	ug/L	105	(50-150)		
MS_202205260804	4,4-DDE		2	2.25	ug/L	112	(70-130)		
DUP_202205260805	4,4-DDT			ND	ug/L		(0-20)		
LCS1	4,4-DDT		2	2.16	ug/L	108	(70-130)		
LCS2	4,4-DDT		2	2.31	ug/L	115	(70-130)	20	6.7
MBLK	4,4-DDT			<0.1	ug/L				
MRL_CHK	4,4-DDT		0.1	0.0990	ug/L	99	(50-150)		
MS_202205260804	4,4-DDT		2	2.32	ug/L	116	(70-130)		
DUP_202205260805	Acenaphthene			ND	ug/L		(0-20)		
LCS1	Acenaphthene		2	1.69	ug/L	85	(70-130)		
LCS2	Acenaphthene		2	1.71	ug/L	86	(70-130)	20	1.2
MBLK	Acenaphthene			<0.1	ug/L				
MRL_CHK	Acenaphthene		0.1	0.0870	ug/L	87	(50-150)		
MS_202205260804	Acenaphthene		2	1.74	ug/L	87	(70-130)		
DUP_202205260805	Acenaphthene-d10 (I)			76.4	%	76	(50-150)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
LCS1	Acenaphthene-d10 (I)		5	98.0	%	98	(50-150)		
LCS2	Acenaphthene-d10 (I)		5	88.0	%	88	(50-150)		
MBLK	Acenaphthene-d10 (I)			83.3	%	83	(50-150)		
MRL_CHK	Acenaphthene-d10 (I)		5	93.4	%	93	(50-150)		
MS_202205260804	Acenaphthene-d10 (I)		5	80.0	%	80	(50-150)		
DUP_202205260805	Acenaphthylene			ND	ug/L		(0-20)		
LCS1	Acenaphthylene		2	1.74	ug/L	87	(70-130)		
LCS2	Acenaphthylene		2	1.75	ug/L	88	(70-130)	20	0.57
MBLK	Acenaphthylene			<0.1	ug/L				
MRL_CHK	Acenaphthylene		0.1	0.0750	ug/L	75	(50-150)		
MS_202205260804	Acenaphthylene		2	1.78	ug/L	89	(70-130)		
DUP_202205260805	Acetochlor			ND	ug/L		(0-20)		
LCS1	Acetochlor		2	1.89	ug/L	95	(70-130)		
LCS2	Acetochlor		2	1.94	ug/L	97	(70-130)	20	2.6
MBLK	Acetochlor			<0.1	ug/L				
MRL_CHK	Acetochlor		0.05	0.0400	ug/L	80	(50-150)		
MS_202205260804	Acetochlor		2	1.98	ug/L	99	(70-130)		
DUP_202205260805	Alachlor			ND	ug/L		(0-20)		
LCS1	Alachlor		2	1.89	ug/L	94	(70-130)		
LCS2	Alachlor		2	1.94	ug/L	97	(70-130)	20	2.6
MBLK	Alachlor			<0.05	ug/L				
MRL_CHK	Alachlor		0.05	0.0480	ug/L	96	(50-150)		
MS_202205260804	Alachlor		2	1.95	ug/L	98	(70-130)		
DUP_202205260805	Alpha-BHC			ND	ug/L		(0-20)		
LCS1	Alpha-BHC		2	1.93	ug/L	96	(70-130)		
LCS2	Alpha-BHC		2	1.96	ug/L	98	(70-130)	20	1.5
MBLK	Alpha-BHC			<0.1	ug/L				
MRL_CHK	Alpha-BHC		0.1	0.111	ug/L	111	(50-150)		
MS_202205260804	Alpha-BHC		2	2.03	ug/L	101	(70-130)		
DUP_202205260805	alpha-Chlordane			ND	ug/L		(0-20)		
LCS1	alpha-Chlordane		2	2.12	ug/L	106	(70-130)		
LCS2	alpha-Chlordane		2	2.19	ug/L	109	(70-130)	20	2.8
MBLK	alpha-Chlordane			<0.05	ug/L				
MRL_CHK	alpha-Chlordane		0.05	0.0540	ug/L	108	(50-150)		
MS_202205260804	alpha-Chlordane		2	2.20	ug/L	110	(70-130)		
DUP_202205260805	Anthracene			ND	ug/L		(0-20)		
LCS1	Anthracene		2	1.84	ug/L	92	(70-130)		
LCS2	Anthracene		2	1.86	ug/L	93	(70-130)	20	1.1

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MBLK	Anthracene			<0.02	ug/L				
MRL_CHK	Anthracene		0.02	0.0180	ug/L	90	(50-150)		
MS_202205260804	Anthracene		2	1.87	ug/L	93	(70-130)		
DUP_202205260805	Atrazine	ND		ND	ug/L		(0-20)		
LCS1	Atrazine		2	2.14	ug/L	107	(70-130)		
LCS2	Atrazine		2	2.28	ug/L	114	(70-130)	20	6.3
MBLK	Atrazine			<0.05	ug/L				
MRL_CHK	Atrazine		0.05	0.0550	ug/L	110	(50-150)		
MS_202205260804	Atrazine	ND	2	2.40	ug/L	120	(70-130)		
DUP_202205260805	Benz(a)Anthracene			ND	ug/L		(0-20)		
LCS1	Benz(a)Anthracene		2	2.02	ug/L	101	(70-130)		
LCS2	Benz(a)Anthracene		2	2.17	ug/L	109	(70-130)	20	7.2
MBLK	Benz(a)Anthracene			<0.05	ug/L				
MRL_CHK	Benz(a)Anthracene		0.05	0.0510	ug/L	102	(50-150)		
MS_202205260804	Benz(a)Anthracene		2	2.16	ug/L	108	(70-130)		
DUP_202205260805	Benzo(a)pyrene	ND		ND	ug/L		(0-20)		
LCS1	Benzo(a)pyrene		2	1.91	ug/L	96	(70-130)		
LCS2	Benzo(a)pyrene		2	2.00	ug/L	100	(70-130)	20	4.6
MBLK	Benzo(a)pyrene			<0.02	ug/L				
MRL_CHK	Benzo(a)pyrene		0.02	0.0150	ug/L	75	(50-150)		
MS_202205260804	Benzo(a)pyrene	ND	2	1.97	ug/L	99	(70-130)		
DUP_202205260805	Benzo(b)Fluoranthene			ND	ug/L		(0-20)		
LCS1	Benzo(b)Fluoranthene		2	1.92	ug/L	96	(70-130)		
LCS2	Benzo(b)Fluoranthene		2	2.08	ug/L	104	(70-130)	20	8.0
MBLK	Benzo(b)Fluoranthene			<0.02	ug/L				
MRL_CHK	Benzo(b)Fluoranthene		0.02	0.0190	ug/L	95	(50-150)		
MS_202205260804	Benzo(b)Fluoranthene		2	1.99	ug/L	100	(70-130)		
DUP_202205260805	Benzo(g,h,i)Perylene			ND	ug/L		(0-20)		
LCS1	Benzo(g,h,i)Perylene		2	2.21	ug/L	110	(70-130)		
LCS2	Benzo(g,h,i)Perylene		2	2.28	ug/L	114	(70-130)	20	3.1
MBLK	Benzo(g,h,i)Perylene			<0.05	ug/L				
MRL_CHK	Benzo(g,h,i)Perylene		0.05	0.0460	ug/L	92	(50-150)		
MS_202205260804	Benzo(g,h,i)Perylene		2	2.19	ug/L	110	(70-130)		
DUP_202205260805	Benzo(k)Fluoranthene			ND	ug/L		(0-20)		
LCS1	Benzo(k)Fluoranthene		2	2.15	ug/L	108	(70-130)		
LCS2	Benzo(k)Fluoranthene		2	2.21	ug/L	111	(70-130)	20	2.8
MBLK	Benzo(k)Fluoranthene			<0.02	ug/L				
MRL_CHK	Benzo(k)Fluoranthene		0.02	0.0170	ug/L	85	(50-150)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MS_202205260804	Benzo(k)Fluoranthene		2	2.14	ug/L	107	(70-130)		
DUP_202205260805	Beta-BHC			ND	ug/L		(0-20)		
LCS1	Beta-BHC		2	1.97	ug/L	99	(70-130)		
LCS2	Beta-BHC		2	2.02	ug/L	101	(70-130)	20	2.5
MBLK	Beta-BHC			<0.1	ug/L				
MRL_CHK	Beta-BHC		0.1	0.0960	ug/L	96	(50-150)		
MS_202205260804	Beta-BHC		2	2.14	ug/L	107	(70-130)		
DUP_202205260805	Bromacil			ND	ug/L		(0-20)		
LCS1	Bromacil		2	1.62	ug/L	81	(70-130)		
LCS2	Bromacil		2	2.38	ug/L	119	(70-130)	20	<u>37</u>
MBLK	Bromacil			<0.2	ug/L				
MRL_CHK	Bromacil		0.1	0.113	ug/L	113	(50-150)		
MS_202205260804	Bromacil		2	2.48	ug/L	124	(70-130)		
DUP_202205260805	Butachlor			ND	ug/L		(0-20)		
LCS1	Butachlor		2	1.93	ug/L	96	(70-130)		
LCS2	Butachlor		2	2.02	ug/L	101	(70-130)	20	4.6
MBLK	Butachlor			<0.05	ug/L				
MRL_CHK	Butachlor		0.05	0.0500	ug/L	100	(50-150)		
MS_202205260804	Butachlor		2	2.04	ug/L	102	(70-130)		
DUP_202205260805	Butylbenzylphthalate			ND	ug/L		(0-20)		
LCS1	Butylbenzylphthalate		2	1.92	ug/L	96	(70-130)		
LCS2	Butylbenzylphthalate		2	2.05	ug/L	102	(70-130)	20	6.5
MBLK	Butylbenzylphthalate			<0.5	ug/L				
MRL_CHK	Butylbenzylphthalate		0.15	0.163	ug/L	109	(50-150)		
MS_202205260804	Butylbenzylphthalate		2	2.10	ug/L	105	(70-130)		
DUP_202205260805	Caffeine by method 525mod			ND	ug/L		(0-20)		
LCS1	Caffeine by method 525mod		2	0.742	ug/L	<u>37</u>	(45-137)		
LCS2	Caffeine by method 525mod		2	1.80	ug/L	90	(45-137)	20	<u>83</u>
MBLK	Caffeine by method 525mod			<0.05	ug/L				
MRL_CHK	Caffeine by method 525mod		0.05	0.0410	ug/L	82	(50-150)		
MS_202205260804	Caffeine by method 525mod		2	2.15	ug/L	108	(46-144)		
DUP_202205260805	Chlorobenzilate			ND	ug/L		(0-20)		
LCS1	Chlorobenzilate		2	1.96	ug/L	98	(70-130)		
LC 2	Chlorobenzilate		2	2.06	ug/L	103	(70 130)	20	5.0
MBLK	Chlorobenzilate			<0.1	ug/L				
MRL_CHK	Chlorobenzilate		0.1	0.0950	ug/L	95	(50-150)		
MS_202205260804	Chlorobenzilate		2	2.14	ug/L	107	(70-130)		
DUP_202205260805	Chloroneb			ND	ug/L		(0-20)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
LCS1	Chloroneb		2	1.92	ug/L	96	(70-130)		
LCS2	Chloroneb		2	1.93	ug/L	97	(70-130)	20	0.52
MBLK	Chloroneb			<0.1	ug/L				
MRL_CHK	Chloroneb		0.1	0.101	ug/L	101	(50-150)		
MS_202205260804	Chloroneb		2	1.96	ug/L	98	(70-130)		
DUP_202205260805	Chlorothalonil(Draconil,Bravo)			ND	ug/L		(0-20)		
LCS1	Chlorothalonil(Draconil,Bravo)		2	2.27	ug/L	114	(70-130)		
LCS2	Chlorothalonil(Draconil,Bravo)		2	2.33	ug/L	117	(70-130)	20	2.6
MBLK	Chlorothalonil(Draconil,Bravo)			<0.1	ug/L				
MRL_CHK	Chlorothalonil(Draconil,Bravo)		0.1	0.105	ug/L	105	(50-150)		
MS_202205260804	Chlorothalonil(Draconil,Bravo)		2	2.37	ug/L	118	(70-130)		
DUP_202205260805	Chlorpyrifos (Dursban)			ND	ug/L		(0-20)		
LCS1	Chlorpyrifos (Dursban)		2	1.95	ug/L	98	(70-130)		
LCS2	Chlorpyrifos (Dursban)		2	1.97	ug/L	99	(70-130)	20	1.0
MBLK	Chlorpyrifos (Dursban)			<0.05	ug/L				
MRL_CHK	Chlorpyrifos (Dursban)		0.05	0.0430	ug/L	86	(50-150)		
MS_202205260804	Chlorpyrifos (Dursban)		2	1.99	ug/L	100	(70-130)		
DUP_202205260805	Chrysene			ND	ug/L		(0-20)		
LCS1	Chrysene		2	1.96	ug/L	98	(70-130)		
LCS2	Chrysene		2	2.01	ug/L	101	(70-130)	20	2.5
MBLK	Chrysene			<0.02	ug/L				
MRL_CHK	Chrysene		0.02	0.0190	ug/L	95	(50-150)		
MS_202205260804	Chrysene		2	2.01	ug/L	101	(70-130)		
DUP_202205260805	Chrysene-d12 (I)			92.3	%	92	(50-150)		
LCS1	Chrysene-d12 (I)		5	107	%	107	(50-150)		
LCS2	Chrysene-d12 (I)		5	104	%	104	(50-150)		
MBLK	Chrysene-d12 (I)			97.4	%	97	(50-150)		
MRL_CHK	Chrysene-d12 (I)		5	99.3	%	99	(50-150)		
MS_202205260804	Chrysene-d12 (I)		5	97.0	%	97	(50-150)		
DUP_202205260805	Delta-BHC			ND	ug/L		(0-20)		
LCS1	Delta-BHC		2	1.84	ug/L	92	(70-130)		
LCS2	Delta-BHC		2	1.86	ug/L	93	(70-130)	20	1.1
MBLK	Delta-BHC			<0.1	ug/L				
MRL_CHK	Delta-BHC		0.1	0.108	ug/L	108	(50-150)		
MS_202205260804	Delta-BHC		2	1.92	ug/L	96	(70-130)		
DUP_202205260805	Di-(2-Ethylhexyl)adipate	ND		ND	ug/L		(0-20)		
LCS1	Di-(2-Ethylhexyl)adipate		2	1.90	ug/L	95	(70-130)		
LCS2	Di-(2-Ethylhexyl)adipate		2	2.14	ug/L	107	(70-130)	20	12

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MBLK	Di-(2-Ethylhexyl)adipate			<0.6	ug/L				
MRL_CHK	Di-(2-Ethylhexyl)adipate		0.3	0.331	ug/L	110	(50-150)		
MS_202205260804	Di-(2-Ethylhexyl)adipate	ND	2	2.03	ug/L	102	(70-130)		
DUP_202205260805	Di(2-E hylhexyl)phthalate	ND		ND	ug/L		(0-20)		
LCS1	Di(2-E hylhexyl)phthalate		2	1.71	ug/L	85	(70-130)		
LCS2	Di(2-E hylhexyl)phthalate		2	1.84	ug/L	92	(70-130)	20	7.3
MBLK	Di(2-E hylhexyl)phthalate			<0.6	ug/L				
MRL_CHK	Di(2-E hylhexyl)phthalate		0.6	0.588	ug/L	98	(50-150)		
MS_202205260804	Di(2-E hylhexyl)phthalate	ND	2	1.74	ug/L	87	(70-130)		
DUP_202205260805	Diazinon (Qualitative)			ND	ug/L		(0-20)		
LCS1	Diazinon (Qualitative)		2	1.60	ug/L	80	(15-132)		
LCS2	Diazinon (Qualitative)		2	1.72	ug/L	86	(15-132)	20	7.2
MBLK	Diazinon (Qualitative)			<0.10	ug/L				
MRL_CHK	Diazinon (Qualitative)		0.1	0.0790	ug/L	79	(15-132)		
MS_202205260804	Diazinon (Qualitative)		2	1.92	ug/L	96	(15-132)		
DUP_202205260805	Dibenz(a,h)Anthracene			ND	ug/L		(0-20)		
LCS1	Dibenz(a,h)Anthracene		2	2.13	ug/L	107	(70-130)		
LCS2	Dibenz(a,h)Anthracene		2	2.26	ug/L	113	(70-130)	20	5.9
MBLK	Dibenz(a,h)Anthracene			<0.05	ug/L				
MRL_CHK	Dibenz(a,h)Anthracene		0.05	0.0420	ug/L	84	(50-150)		
MS_202205260804	Dibenz(a,h)Anthracene		2	2.12	ug/L	106	(70-130)		
DUP_202205260805	Dichlorvos (DDVP)			ND	ug/L		(0-20)		
LCS1	Dichlorvos (DDVP)		2	2.07	ug/L	103	(70-130)		
LCS2	Dichlorvos (DDVP)		2	2.21	ug/L	111	(70-130)	20	6.5
MBLK	Dichlorvos (DDVP)			<0.05	ug/L				
MRL_CHK	Dichlorvos (DDVP)		0.05	0.0540	ug/L	108	(50-150)		
MS_202205260804	Dichlorvos (DDVP)		2	2.28	ug/L	114	(70-130)		
DUP_202205260805	Dieldrin			ND	ug/L		(0-20)		
LCS1	Dieldrin		2	1.84	ug/L	92	(70-130)		
LCS2	Dieldrin		2	1.92	ug/L	96	(70-130)	20	4.3
MBLK	Dieldrin			<0.2	ug/L				
MRL_CHK	Dieldrin		0.1	0.115	ug/L	115	(50-150)		
MS_202205260804	Dieldrin		2	1.99	ug/L	99	(70-130)		
DUP_202205260805	Diethylphthalate			ND	ug/L		(0-20)		
LCS1	Diethylphthalate		2	1.85	ug/L	93	(70-130)		
LCS2	Diethylphthalate		2	1.89	ug/L	95	(70-130)	20	2.1
MBLK	Diethylphthalate			<0.5	ug/L				
MRL_CHK	Diethylphthalate		0.15	0.166	ug/L	111	(50-150)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MS_202205260804	Diethylphthalate		2	1.95	ug/L	98	(70-130)		
DUP_202205260805	Dimethoate			ND	ug/L		(0-20)		
LCS1	Dimethoate		2	0.744	ug/L	37	(35-100)		
LCS2	Dimethoate		2	1.73	ug/L	87	(35-100)	20	<u>80</u>
MBLK	Dimethoate			<0.1	ug/L				
MRL_CHK	Dimethoate		0.1	0.0950	ug/L	95	(35-100)		
MS_202205260804	Dimethoate		2	2.02	ug/L	101	(34-111)		
DUP_202205260805	Dimethylphthalate			ND	ug/L		(0-20)		
LCS1	Dimethylphthalate		2	1.97	ug/L	99	(70-130)		
LCS2	Dimethylphthalate		2	2.04	ug/L	102	(70-130)	20	3.5
MBLK	Dimethylphthalate			<0.5	ug/L				
MRL_CHK	Dimethylphthalate		0.3	0.295	ug/L	98	(50-150)		
MS_202205260804	Dimethylphthalate		2	2.09	ug/L	105	(70-130)		
DUP_202205260805	Di-n-Butylphthalate			ND	ug/L		(0-20)		
LCS1	Di-n-Butylphthalate		4	3.39	ug/L	85	(70-130)		
LCS2	Di-n-Butylphthalate		4	3.50	ug/L	88	(70-130)	20	3.2
MBLK	Di-n-Butylphthalate			<1	ug/L				
MRL_CHK	Di-n-Butylphthalate		0.3	0.316	ug/L	105	(50-150)		
MS_202205260804	Di-n-Butylphthalate		4	3.59	ug/L	90	(70-130)		
DUP_202205260805	Di-N-octylphthalate			ND	ug/L		(0-20)		
LCS1	Di-N-octylphthalate		2	1.42	ug/L	71	(70-130)		
LCS2	Di-N-octylphthalate		2	1.55	ug/L	78	(70-130)	20	8.8
MBLK	Di-N-octylphthalate			<0.1	ug/L				
MRL_CHK	Di-N-octylphthalate		0.1	0.0760	ug/L	76	(50-150)		
MS_202205260804	Di-N-octylphthalate		2	1.45	ug/L	72	(70-130)		
DUP_202205260805	Endosulfan I (Alpha)			ND	ug/L		(0-20)		
LCS1	Endosulfan I (Alpha)		2	1.84	ug/L	92	(70-130)		
LCS2	Endosulfan I (Alpha)		2	1.95	ug/L	97	(70-130)	20	5.8
MBLK	Endosulfan I (Alpha)			<0.1	ug/L				
MRL_CHK	Endosulfan I (Alpha)		0.1	0.106	ug/L	106	(50-150)		
MS_202205260804	Endosulfan I (Alpha)		2	1.99	ug/L	99	(70-130)		
DUP_202205260805	Endosulfan II (Beta)			ND	ug/L		(0-20)		
LCS1	Endosulfan II (Beta)		2	1.95	ug/L	98	(70-130)		
LCS2	Endosulfan II (Beta)		2	2.05	ug/L	103	(70-130)	20	5.0
MBLK	Endosulfan II (Beta)			<0.1	ug/L				
MRL_CHK	Endosulfan II (Beta)		0.1	0.112	ug/L	112	(50-150)		
MS_202205260804	Endosulfan II (Beta)		2	2.11	ug/L	106	(70-130)		
DUP_202205260805	Endosulfan Sulfate			ND	ug/L		(0-20)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
LCS1	Endosulfan Sulfate		2	2.23	ug/L	111	(70-130)		
LCS2	Endosulfan Sulfate		2	2.37	ug/L	119	(70-130)	20	6.1
MBLK	Endosulfan Sulfate			<0.1	ug/L				
MRL_CHK	Endosulfan Sulfate		0.1	0.108	ug/L	108	(50-150)		
MS_202205260804	Endosulfan Sulfate		2	2.47	ug/L	124	(70-130)		
DUP_202205260805	Endrin			ND	ug/L		(0-20)		
LCS1	Endrin		2	1.85	ug/L	93	(70-130)		
LCS2	Endrin		2	1.98	ug/L	99	(70-130)	20	6.8
MBLK	Endrin			<0.1	ug/L				
MRL_CHK	Endrin		0.1	0.0990	ug/L	99	(50-150)		
MS_202205260804	Endrin		2	1.95	ug/L	98	(70-130)		
DUP_202205260805	Endrin Aldehyde			ND	ug/L		(0-20)		
LCS1	Endrin Aldehyde		2	2.03	ug/L	101	(70-130)		
LCS2	Endrin Aldehyde		2	2.16	ug/L	108	(70-130)	20	6.2
MBLK	Endrin Aldehyde			<0.1	ug/L				
MRL_CHK	Endrin Aldehyde		0.1	0.0860	ug/L	86	(50-150)		
MS_202205260804	Endrin Aldehyde		2	2.19	ug/L	110	(70-130)		
DUP_202205260805	EPTC			ND	ug/L		(0-20)		
LCS1	EPTC		2	1.84	ug/L	92	(70-130)		
LCS2	EPTC		2	1.84	ug/L	92	(70-130)	20	0.0
MBLK	EPTC			<0.1	ug/L				
MRL_CHK	EPTC		0.1	0.0960	ug/L	96	(50-150)		
MS_202205260804	EPTC		2	1.86	ug/L	93	(70-130)		
DUP_202205260805	Fluoranthene			ND	ug/L		(0-20)		
LCS1	Fluoranthene		2	1.93	ug/L	97	(70-130)		
LCS2	Fluoranthene		2	2.02	ug/L	101	(70-130)	20	4.6
MBLK	Fluoranthene			<0.1	ug/L				
MRL_CHK	Fluoranthene		0.05	0.0490	ug/L	98	(50-150)		
MS_202205260804	Fluoranthene		2	2.04	ug/L	102	(70-130)		
DUP_202205260805	Fluorene			ND	ug/L		(0-20)		
LCS1	Fluorene		2	1.88	ug/L	94	(70-130)		
LCS2	Fluorene		2	1.91	ug/L	96	(70-130)	20	1.6
MBLK	Fluorene			<0.05	ug/L				
MRL_CHK	Fluorene		0.05	0.0480	ug/L	96	(50-150)		
MS_202205260804	Fluorene		2	1.97	ug/L	99	(70-130)		
DUP_202205260805	gamma-Chlordane			ND	ug/L		(0-20)		
LCS1	gamma-Chlordane		2	2.13	ug/L	106	(70-130)		
LCS2	gamma-Chlordane		2	2.21	ug/L	110	(70-130)	20	3.7

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MBLK	gamma-Chlordane			<0.05	ug/L				
MRL_CHK	gamma-Chlordane		0.05	0.0560	ug/L	112	(50-150)		
MS_202205260804	gamma-Chlordane		2	2.17	ug/L	108	(70-130)		
DUP_202205260805	Heptachlor			ND	ug/L		(0-20)		
LCS1	Heptachlor		2	1.73	ug/L	87	(70-130)		
LCS2	Heptachlor		2	1.76	ug/L	88	(70-130)	20	1.7
MBLK	Heptachlor			<0.04	ug/L				
MRL_CHK	Heptachlor		0.04	0.0430	ug/L	108	(50-150)		
MS_202205260804	Heptachlor		2	1.78	ug/L	89	(70-130)		
DUP_202205260805	Heptachlor Epoxide (isomer B)			ND	ug/L		(0-20)		
LCS1	Heptachlor Epoxide (isomer B)		2	2.12	ug/L	106	(70-130)		
LCS2	Heptachlor Epoxide (isomer B)		2	2.19	ug/L	110	(70-130)	20	3.3
MBLK	Heptachlor Epoxide (isomer B)			<0.05	ug/L				
MRL_CHK	Heptachlor Epoxide (isomer B)		0.05	0.0500	ug/L	100	(50-150)		
MS_202205260804	Heptachlor Epoxide (isomer B)		2	2.19	ug/L	110	(70-130)		
DUP_202205260805	Hexachlorobenzene	ND		ND	ug/L		(0-20)		
LCS1	Hexachlorobenzene		2	2.02	ug/L	101	(70-130)		
LCS2	Hexachlorobenzene		2	2.07	ug/L	104	(70-130)	20	2.4
MBLK	Hexachlorobenzene			<0.05	ug/L				
MRL_CHK	Hexachlorobenzene		0.05	0.0600	ug/L	120	(50-150)		
MS_202205260804	Hexachlorobenzene	ND	2	2.10	ug/L	105	(70-130)		
DUP_202205260805	Hexachlorocyclopentadiene	ND		ND	ug/L		(0-20)		
LCS1	Hexachlorocyclopentadiene		2	1.94	ug/L	97	(70-130)		
LCS2	Hexachlorocyclopentadiene		2	1.94	ug/L	97	(70-130)	20	0.0
MBLK	Hexachlorocyclopentadiene			<0.05	ug/L				
MRL_CHK	Hexachlorocyclopentadiene		0.05	0.0440	ug/L	88	(50-150)		
MS_202205260804	Hexachlorocyclopentadiene	ND	2	1.94	ug/L	97	(70-130)		
DUP_202205260805	Indeno(1,2,3,c,d)Pyrene			ND	ug/L		(0-20)		
LCS1	Indeno(1,2,3,c,d)Pyrene		2	2.17	ug/L	108	(70-130)		
LCS2	Indeno(1,2,3,c,d)Pyrene		2	2.22	ug/L	111	(70-130)	20	2.7
MBLK	Indeno(1,2,3,c,d)Pyrene			<0.05	ug/L				
MRL_CHK	Indeno(1,2,3,c,d)Pyrene		0.05	0.0420	ug/L	84	(50-150)		
MS_202205260804	Indeno(1,2,3,c,d)Pyrene		2	2.12	ug/L	106	(70-130)		
DUP_202205260805	Isophorone			ND	ug/L		(0-20)		
LCS1	Isophorone		2	1.88	ug/L	94	(70-130)		
LCS2	Isophorone		2	1.93	ug/L	97	(70-130)	20	2.6
MBLK	Isophorone			<0.5	ug/L				
MRL_CHK	Isophorone		0.1	0.0920	ug/L	92	(50-150)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MS_202205260804	Isophorone		2	1.96	ug/L	98	(70-130)		
DUP_202205260805	Lindane			ND	ug/L		(0-20)		
LCS1	Lindane		2	1.81	ug/L	90	(70-130)		
LCS2	Lindane		2	1.84	ug/L	92	(70-130)	20	1.6
MBLK	Lindane			<0.04	ug/L				
MRL_CHK	Lindane		0.04	0.0440	ug/L	110	(50-150)		
MS_202205260804	Lindane		2	1.95	ug/L	97	(70-130)		
DUP_202205260805	Malathion			ND	ug/L		(0-20)		
LCS1	Malathion		2	2.03	ug/L	101	(70-130)		
LCS2	Malathion		2	2.10	ug/L	105	(70-130)	20	3.4
MBLK	Malathion			<0.1	ug/L				
MRL_CHK	Malathion		0.1	0.0900	ug/L	90	(50-150)		
MS_202205260804	Malathion		2	2.12	ug/L	106	(70-130)		
DUP_202205260805	Methoxychlor			ND	ug/L		(0-20)		
LCS1	Methoxychlor		2	2.16	ug/L	108	(70-130)		
LCS2	Methoxychlor		2	2.18	ug/L	109	(70-130)	20	0.92
MBLK	Methoxychlor			<0.1	ug/L				
MRL_CHK	Methoxychlor		0.1	0.0960	ug/L	96	(50-150)		
MS_202205260804	Methoxychlor		2	2.21	ug/L	111	(70-130)		
DUP_202205260805	Metolachlor			ND	ug/L		(0-20)		
LCS1	Metolachlor		2	1.88	ug/L	94	(70-130)		
LCS2	Metolachlor		2	1.92	ug/L	96	(70-130)	20	2.1
MBLK	Metolachlor			<0.05	ug/L				
MRL_CHK	Metolachlor		0.05	0.0460	ug/L	92	(50-150)		
MS_202205260804	Metolachlor		2	1.94	ug/L	97	(70-130)		
DUP_202205260805	Metribuzin			ND	ug/L		(0-20)		
LCS1	Metribuzin		2	1.57	ug/L	79	(70-130)		
LCS2	Metribuzin		2	2.02	ug/L	101	(70-130)	20	<u>25</u>
MBLK	Metribuzin			<0.05	ug/L				
MRL_CHK	Metribuzin		0.05	0.0450	ug/L	90	(50-150)		
MS_202205260804	Metribuzin		2	2.14	ug/L	107	(70-130)		
DUP_202205260805	Molinate	ND		ND	ug/L		(0-20)		
LCS1	Molinate		2	1.95	ug/L	97	(70-130)		
LC 2	Molinate		2	1.98	ug/L	99	(70 130)	20	1.5
MBLK	Molinate			<0.1	ug/L				
MRL_CHK	Molinate		0.1	0.103	ug/L	103	(50-150)		
MS_202205260804	Molinate	ND	2	2.02	ug/L	101	(70-130)		
DUP_202205260805	Naphthalene			ND	ug/L		(0-20)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
LCS1	Naphthalene		2	1.77	ug/L	89	(70-130)		
LCS2	Naphthalene		2	1.75	ug/L	88	(70-130)	20	1.1
MBLK	Naphthalene			<0.3	ug/L				
MRL_CHK	Naphthalene		0.1	0.0930	ug/L	93	(50-150)		
MS_202205260804	Naphthalene		2	1.78	ug/L	89	(70-130)		
DUP_202205260805	Parathion			ND	ug/L		(0-20)		
LCS1	Parathion		2	1.96	ug/L	98	(70-130)		
LCS2	Parathion		2	2.02	ug/L	101	(70-130)	20	3.0
MBLK	Parathion			<0.1	ug/L				
MRL_CHK	Parathion		0.1	0.0880	ug/L	88	(50-150)		
MS_202205260804	Parathion		2	2.05	ug/L	103	(70-130)		
DUP_202205260805	Pendimethalin			ND	ug/L		(0-20)		
LCS1	Pendimethalin		2	1.98	ug/L	99	(70-130)		
LCS2	Pendimethalin		2	2.05	ug/L	103	(70-130)	20	3.5
MBLK	Pendimethalin			<0.1	ug/L				
MRL_CHK	Pendimethalin		0.1	0.0980	ug/L	98	(50-150)		
MS_202205260804	Pendimethalin		2	2.09	ug/L	104	(70-130)		
DUP_202205260805	Permethrin (mixed isomers)			ND	ug/L		(0-20)		
LCS1	Permethrin (mixed isomers)		4	3.34	ug/L	83	(70-130)		
LCS2	Permethrin (mixed isomers)		4	3.54	ug/L	89	(70-130)	20	6.1
MBLK	Permethrin (mixed isomers)			<0.2	ug/L				
MRL_CHK	Permethrin (mixed isomers)		0.2	0.178	ug/L	89	(50-150)		
MS_202205260804	Permethrin (mixed isomers)		4	3.49	ug/L	87	(70-130)		
DUP_202205260805	Perylene-d12 (S)			93.8	%	94	(70-130)		
LCS1	Perylene-d12 (S)		5	96.4	%	96	(70-130)		
LCS2	Perylene-d12 (S)		5	99.4	%	99	(70-130)		
MBLK	Perylene-d12 (S)			93.8	%	94	(70-130)		
MRL_CHK	Perylene-d12 (S)		5	90.4	%	90	(70-130)		
MS_202205260804	Perylene-d12 (S)		5	95.8	%	96	(70-130)		
DUP_202205260805	Phenanthrene			ND	ug/L		(0-20)		
LCS1	Phenanthrene		2	1.76	ug/L	88	(70-130)		
LCS2	Phenanthrene		2	1.78	ug/L	89	(70-130)	20	1.1
MBLK	Phenanthrene			<0.04	ug/L				
MRL_CHK	Phenanthrene		0.02	0.0200	ug/L	100	(50-150)		
MS_202205260804	Phenanthrene		2	1.80	ug/L	90	(70-130)		
DUP_202205260805	Phenanthrene-d10 (I)			82.9	%	83	(50-150)		
LCS1	Phenanthrene-d10 (I)		5	102	%	102	(50-150)		
LCS2	Phenanthrene-d10 (I)		5	93.6	%	94	(50-150)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 1007133
 Project: RED-HILL
 Group: Weekly TPH-8015_RED-HILL (2022)
 - EMAX

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MBLK	Phenanthrene-d10 (I)			88.2	%	88	(50-150)		
MRL_CHK	Phenanthrene-d10 (I)		5	96.8	%	97	(50-150)		
MS_202205260804	Phenanthrene-d10 (I)		5	88.2	%	88	(50-150)		
DUP_202205260805	Propachlor			ND	ug/L		(0-20)		
LCS1	Propachlor		2	1.88	ug/L	94	(70-130)		
LCS2	Propachlor		2	1.97	ug/L	98	(70-130)	20	4.7
MBLK	Propachlor			<0.05	ug/L				
MRL_CHK	Propachlor		0.05	0.0490	ug/L	98	(50-150)		
MS_202205260804	Propachlor		2	2.10	ug/L	105	(70-130)		
DUP_202205260805	Pyrene			ND	ug/L		(0-20)		
LCS1	Pyrene		2	1.99	ug/L	99	(70-130)		
LCS2	Pyrene		2	2.04	ug/L	102	(70-130)	20	2.5
MBLK	Pyrene			<0.05	ug/L				
MRL_CHK	Pyrene		0.05	0.0480	ug/L	96	(50-150)		
MS_202205260804	Pyrene		2	2.10	ug/L	105	(70-130)		
DUP_202205260805	Simazine	ND		ND	ug/L		(0-20)		
LCS1	Simazine		2	1.96	ug/L	98	(70-130)		
LCS2	Simazine		2	2.26	ug/L	113	(70-130)	20	14
MBLK	Simazine			<0.05	ug/L				
MRL_CHK	Simazine		0.05	0.0560	ug/L	112	(50-150)		
MS_202205260804	Simazine	ND	2	2.42	ug/L	121	(70-130)		
DUP_202205260805	Terbacil			ND	ug/L		(0-20)		
LCS1	Terbacil		2	1.72	ug/L	86	(70-130)		
LCS2	Terbacil		2	2.21	ug/L	111	(70-130)	20	25
MBLK	Terbacil			<0.1	ug/L				
MRL_CHK	Terbacil		0.1	0.126	ug/L	126	(50-150)		
MS_202205260804	Terbacil		2	2.25	ug/L	113	(70-130)		
DUP_202205260805	Terbutylazine			ND	ug/L		(0-20)		
LCS1	Terbutylazine		2	2.09	ug/L	105	(70-130)		
LCS2	Terbutylazine		2	2.19	ug/L	110	(70-130)	20	4.7
MBLK	Terbutylazine			<0.1	ug/L				
MRL_CHK	Terbutylazine		0.1	0.101	ug/L	101	(50-150)		
MS_202205260804	Terbutylazine		2	2.32	ug/L	116	(70-130)		
DUP_202205260805	Thiobencarb	ND		ND	ug/L		(0 20)		
LCS1	Thiobencarb		2	1.85	ug/L	93	(70-130)		
LCS2	Thiobencarb		2	1.89	ug/L	94	(70-130)	20	2.1
MBLK	Thiobencarb			<0.2	ug/L				
MRL_CHK	Thiobencarb		0.1	0.0960	ug/L	96	(50-150)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Honolulu Board of Water Supply

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield(%)	Limits (%)	RPD Limit(%)	RPD%
MS_202205260804	Thiobencarb	ND	2	1.94	ug/L	97	(70-130)		
DUP_202205260805	trans-Nonachlor			ND	ug/L		(0-20)		
LCS1	trans-Nonachlor		2	2.26	ug/L	113	(70-130)		
LCS2	trans-Nonachlor		2	2.31	ug/L	115	(70-130)	20	2.2
MBLK	trans-Nonachlor			<0.05	ug/L				
MRL_CHK	trans-Nonachlor		0.05	0.0540	ug/L	108	(50-150)		
MS_202205260804	trans-Nonachlor		2	2.34	ug/L	117	(70-130)		
DUP_202205260805	Trifluralin			ND	ug/L		(0-20)		
LCS1	Trifluralin		2	1.90	ug/L	95	(70-130)		
LCS2	Trifluralin		2	1.92	ug/L	96	(70-130)	20	1.1
MBLK	Trifluralin			<0.1	ug/L				
MRL_CHK	Trifluralin		0.1	0.0930	ug/L	93	(50-150)		
MS_202205260804	Trifluralin		2	1.98	ug/L	99	(70-130)		
DUP_202205260805	Triphenylphosphate (S)			121	%	121	(70-130)		
LCS1	Triphenylphosphate (S)		5	111	%	111	(70-130)		
LCS2	Triphenylphosphate (S)		5	117	%	117	(70-130)		
MBLK	Triphenylphosphate (S)			117	%	117	(70-130)		
MRL_CHK	Triphenylphosphate (S)		5	107	%	107	(70-130)		
MS_202205260804	Triphenylphosphate (S)		5	120	%	120	(70-130)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.



Eaton Analytical

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Hits

Report: 1007133
Project: RED-HILL
Group: Weekly TPH-8015_RED-HILL (2022)
- EMAX

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Samples Received on:
06/02/2022 1012

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
----------	---------	-----------	--------	-------------	-------	-----

SUMMARY OF POSITIVE DATA ONLY



3051 Fujita Street
Torrance, CA 90505
Tel: (310)-618-8889

Date: 06-16-2022
EMAX Batch No.: 22F042

Attn: Jackie Contreras

Eurofins Eaton Analytical
750 Royal Oaks Dr., Suite 100
Monrovia, CA 91016-3629

Subject: Laboratory Report
Project: 1007133


Enclosed is the Laboratory report for samples received on 06/03/22.
The data reported relate only to samples listed below :

Sample ID	Control #	Col Date	Matrix	Analysis
202206030089	F042-01	05/31/22	WATER	TPH GASOLINE TPH DIESEL & MOTOR OIL
202206030090	F042-02	05/31/22	WATER	TPH GASOLINE

The results are summarized on the following pages.

Please feel free to call if you have any questions concerning these results.

Sincerely yours,



Caspar J. Pang
Laboratory Director

This report is confidential and intended solely for the use of the individual or entity to whom it is addressed. This report shall not be reproduced except in full or without the written approval of EMAX.

EMAX certifies that results included in this report meets all TNI & DOD requirements unless noted in the Case Narrative.

NELAP Accredited Certificate Number CA002912022-22
ANAB Accredited DoD ELAP and ISO/IEC 17025 Certificate Number L2278 Testing
California ELAP Accredited Certificate Number 2672



Eaton Analytical

Ship To:
EMAX Laboratories, Inc.
3051 Fujita St.
Torrance, CA 90505

Phone: 310-618-8889 Fax: 310-618-0818

Folder #: 1007133
Report Due: 06/09/2022

Submittal Form

Date: 6/3/2022

*REPORTING REQUIREMENTS: Do Not Combine Reports with any other samples submitted under different Folder Numbers!
Report & Invoice must have the Folder# 1007133 Job # 1000014

Report all quality control data according to Method. Include dates analyzed. Date extracted (if extracted) and Method reference on the report.
Results must have Complete data & QC with Approval Signature.

Reports: Jackie Contreras Sub-Contracting Administrator
EMAIL TO: Eaton-MonroviaSubContract@eurofins.com
Eurofins Eaton Analytical, LLC 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016
Phone (626) 386-1165 Fax (626) 386-1122
Invoices to: Eurofins Eaton Analytical, LLC
Accounts Payable 2425 New Holland Pike, Lancaster, PA 17605

Provide in each Report the
Specified State Certification # and
Exp Date for requested tests + matrix.
Samples from: HAWAII

Sample ID: 202206030089
Client Sample ID for reference on!
MOANALUA WELLS (331-223-TP202)

Sample type: Sample Event:

Method: SW 8015B
Prep Method: EPA 5030C
Analysis Requested: (SUB)Gas Fraction Hydrocarbons
SW 8015B
Prep Method: EPA 3550B
Analysis Requested: TPH 8015 Diesel and Motor Oil

Sample ID: 202206030090
Client Sample ID for reference on!
TB:MOANALUA WELLS (331-223-TP202)

Sample type: Sample Event:

Method: SW 8015B
Prep Method: EPA 5030C
Analysis Requested: (SUB)Gas Fraction Hydrocarbons

Sample Date & Time Matrix
05/31/22 1029 DW

Clip Code

PWSID

JLS

Sample Point ID:

Static ID:

Sample Date & Time Matrix
05/31/22 1029 DW

Clip Code

PWSID

JLS

Sample Point ID:

Static ID:

Relinquished by: [Signature] Sample Control

Received by: [Signature] Sample Control

Relinquished by: [Signature] Sample Control

Received by: [Signature] Sample Control

Date: 6/3/22 Time: 11:53

Date: Time

Date: 6/3/22 Time: 11:53

Date: Time

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS

An Acknowledgement of Receipt is requested to attn: Jackie Contreras

Temp 049/s.1 @ 3.3/3.5



Type of Delivery <input type="checkbox"/> Fedex <input type="checkbox"/> UPS <input type="checkbox"/> GSO <input type="checkbox"/> Others <input type="checkbox"/> EMAX Courier <input checked="" type="checkbox"/> Client Delivery	Airbill / Tracking Number	ECN <u>22F042</u> Recipient <u>Cecilia Chavez</u> Date <u>6/3/22</u> Time <u>1113</u>
---	---------------------------	---

COC INSPECTION

<input checked="" type="checkbox"/> Client Name	<input checked="" type="checkbox"/> Client PM/FC	<input type="checkbox"/> Sampler Name	<input checked="" type="checkbox"/> Sampling Date/Time	<input checked="" type="checkbox"/> Sample ID	<input checked="" type="checkbox"/> Matrix
<input type="checkbox"/> Address	<input checked="" type="checkbox"/> Tel # / Fax #	<input checked="" type="checkbox"/> Courier Signature	<input type="checkbox"/> Analysis Required	<input type="checkbox"/> Preservative (if any)	<input type="checkbox"/> TAT
Safety Issues (if any) Note: _____					
<input type="checkbox"/> High concentrations expected					
<input type="checkbox"/> From Superfund Site					
<input type="checkbox"/> Rad screening required					

PACKAGING INSPECTION

Container <u>Correction factor +0.2</u>	<input checked="" type="checkbox"/> Cooler	<input type="checkbox"/> Box	<input type="checkbox"/> Other
Condition	<input type="checkbox"/> Custody Seal	<input type="checkbox"/> Intact	<input type="checkbox"/> Damaged
Packaging	<input checked="" type="checkbox"/> Bubble Pack	<input type="checkbox"/> Styrofoam	<input type="checkbox"/> Popcorn
Temperatures (Cool, ≤6 °C but not frozen)	<input checked="" type="checkbox"/> Cooler 1 <u>4.9/5.1</u> °C	<input checked="" type="checkbox"/> Cooler 2 <u>2.3/2.5</u> °C	<input type="checkbox"/> Cooler 3 _____ °C
Thermometer: <u>A - S/N 210583479</u>	<input type="checkbox"/> Cooler 6 _____ °C	<input type="checkbox"/> Cooler 7 _____ °C	<input type="checkbox"/> Cooler 4 _____ °C
	<input type="checkbox"/> Cooler 8 _____ °C	<input type="checkbox"/> Cooler 9 _____ °C	<input type="checkbox"/> Cooler 5 _____ °C
	<input type="checkbox"/> Cooler 10 _____ °C	<input type="checkbox"/> Cooler 11 _____ °C	<input type="checkbox"/> Cooler 12 _____ °C

Comments: Temperature is out of range. PM was informed IMMEDIATELY.
 Note: _____

DISCREPANCIES

LabSampleID	LabSampleContainerID	Code	ClientSample Label ID / Information	Corrective Action
<u>2</u>	<u>9</u>	<u>D22</u>	<u>label reads 5/14/22 & 5/31/22</u>	<u>R1</u>

pH holding time requirement for water samples is 15 mins. Water samples for pH analysis are received beyond 15 minutes from sampling time. MB 6/5/22

NOTES/OBSERVATIONS:
 SAMPLE MATRIX IS DRINKING WATER? YES NO
Analyses is covered in sample 1 container 2 & 3.

LEGEND:

<p><input type="checkbox"/> Continue to next page.</p> <p>Code Description- Sample Management</p> <p>D1 Analysis is not indicated in _____</p> <p>D2 Analysis mismatch COC vs label</p> <p>D3 Sample ID mismatch COC vs label</p> <p>D4 Sample ID is not indicated in _____</p> <p>D5 Container -[improper] [leaking] [broken]</p> <p>D6 Date/Time is not indicated in _____</p> <p>D7 Date/Time mismatch COC vs label</p> <p>D8 Sample listed in COC is not received</p> <p>D9 Sample received is not listed in COC</p> <p>D10 No initial/date on corrections in COC/label</p> <p>D11 Container count mismatch COC vs received</p> <p>D12 Container size mismatch COC vs received</p>	<p>Code Description-Sample Management</p> <p>D13 Out of Holding Time</p> <p>D14 Bubble is >6mm</p> <p>D15 No trip blank in cooler</p> <p>D16 Preservation not indicated in _____</p> <p>D17 Preservation mismatch COC vs label</p> <p>D18 Insufficient chemical preservative</p> <p>D19 Insufficient Sample</p> <p>D20 No filtration info for dissolved analysis</p> <p>D21 No sample for moisture determination</p> <p><u>D22 2 different dates on label</u></p> <p>D23 _____</p> <p>D24 _____</p>	<p>Code Description-Sample Management</p> <p>R1 Proceed as indicated in <input checked="" type="checkbox"/> COC <input type="checkbox"/> Label</p> <p>R2 Refer to attached instruction</p> <p>R3 Cancel the analysis</p> <p>R4 Use vial with smallest bubble first</p> <p>R5 Log-in with latest sampling date and time+1 min</p> <p>R6 Adjust pH as necessary</p> <p>R7 Filter and preserved as necessary</p> <p>R8 _____</p> <p>R9 _____</p> <p>R10 _____</p> <p>R11 _____</p> <p>R12 _____</p>
--	--	--

REVIEWS:

Sample Labeling <u>SHOWIN Zamora</u>	SRF <u>Cecilia</u>	PM <u>MB</u>
Date <u>6/3/22</u>	Date <u>6/3/22</u>	Date <u>6/5/22</u>

REPORTING CONVENTIONS

DATA QUALIFIERS:

Lab Qualifier	AFCEE Qualifier	Description
J	F	Indicates that the analyte is positively identified and the result is less than RL but greater than MDL.
N		Indicates presumptive evidence of a compound.
B	B	Indicates that the analyte is found in the associated method blank as well as in the sample at above QC level.
E	J	Indicates that the result is above the maximum calibration range or estimated value.
*	*	Out of QC limit.

Note: The above qualifiers are used to flag the results unless the project requires a different set of qualification criteria.

ACRONYMS AND ABBREVIATIONS:

CRDL	Contract Required Detection Limit
RL	Reporting Limit
MRL	Method Reporting Limit
PQL	Practical Quantitation Limit
MDL	Method Detection Limit
DO	Diluted out

DATES

The date and time information for leaching and preparation reflect the beginning date and time of the procedure unless the method, protocol, or project specifically requires otherwise.

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

1007133

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

SDG#: 22F042

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 1007133

SDG : 22F042

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

A total of two(2) water samples were received on 06/03/22 to be analyzed for Total Petroleum Hydrocarbons by Purge and Trap in accordance with Method 5030B/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. VG39F02B - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. VG39F02L/VG39F02C were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Gasoline was within MS QC limits in F041-01M/F041-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogate was added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

SAMPLE RESULTS

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 05/31/22 10:29
Project     : 1007133                     Date Received: 06/03/22
Batch No.   : 22F042                       Date Extracted: 06/03/22 18:43
Sample ID   : 202206030090                Date Analyzed: 06/03/22 18:43
Lab Samp ID: F042-02                       Dilution Factor: 1
Lab File ID: EF03012A                       Matrix: WATER
Ext Btch ID: 22VG39F02                     % Moisture: NA
Calib. Ref.: EF03004A                       Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
GASOLINE	ND	0.020	0.010

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0356	0.0400	89	60-140

Notes:

Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

QC SUMMARIES

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 1007133
BATCH NO. : 22F042
METHOD : 5030B/8015B

```

=====
MATRIX      : WATER                                % MOISTURE:NA
DILUTION FACTOR: 1                                1
SAMPLE ID   : MBLK1W                               LCS1W         LCD1W
LAB SAMPLE ID : VG39F02B                           VG39F02L      VG39F02C
LAB FILE ID  : EF03005A                             EF03006A      EF03007A
DATE PREPARED : 06/03/22 14:31                       06/03/22 15:07 06/03/22 15:43
DATE ANALYZED : 06/03/22 14:31                       06/03/22 15:07 06/03/22 15:43
PREP BATCH   : 22VG39F02                             22VG39F02     22VG39F02
CALIBRATION REF: EF03004A                             EF03004A      EF03004A
=====

```

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.473	95	0.500	0.487	97	3	60-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0453	113	0.0400	0.0464	116	70-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 1007135
BATCH NO. : 22F041
METHOD : 5030B/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202206030095                       202206030095MS  202206030095MSD
LAB SAMPLE ID : F041-01                           F041-01M       F041-01S
LAB FILE ID  : EF03016A                           EF03017A       EF03018A
DATE PREPARED : 06/03/22 21:06                   06/03/22 21:42  06/03/22 22:18
DATE ANALYZED : 06/03/22 21:06                   06/03/22 21:42  06/03/22 22:18
PREP BATCH   : 22VG39F02                           22VG39F02      22VG39F02
CALIBRATION REF: EF03015A                           EF03015A       EF03015A
  
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.535	107	0.500	0.557	111	4	50-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0473	118	0.0400	0.0462	116	60-140

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

1007133

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG#: 22F042

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 1007133

SDG : 22F042

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 06/03/22 to be analyzed for Total Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSF005WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. DSF005WL/DSF005WC were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Diesel was within MS QC limits in 22F041-01M/22F041-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

SAMPLE RESULTS

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL      Date Collected: 05/31/22 10:29
Project     : 1007133                        Date Received: 06/03/22
Batch No.   : 22F042                          Date Extracted: 06/06/22 10:15
Sample ID   : 202206030089                   Date Analyzed: 06/07/22 20:22
Lab Samp ID: 22F042-01                        Dilution Factor: 1
Lab File ID: LF07031A                          Matrix: WATER
Ext Btch ID: 22DSF005W                          % Moisture: NA
Calib. Ref.: LF07018A                          Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.024	0.012	
Motor Oil	ND	0.049	0.024	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.349	0.485	72	60-130
Hexacosane	0.107	0.121	89	60-130

Notes:
Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 1030ml Final Volume : 5ml
Prepared by : JMuert Analyzed by : CMpang

QC SUMMARIES

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 06/06/22 10:15
Project     : 1007133                     Date Received: 06/06/22
Batch No.   : 22F042                       Date Extracted: 06/06/22 10:15
Sample ID   : MBLK1W                       Date Analyzed: 06/07/22 17:35
Lab Samp ID: DSF005WB                      Dilution Factor: 1
Lab File ID: LF07022A                      Matrix: WATER
Ext Btch ID: 22DSF005W                    % Moisture: NA
Calib. Ref.: LF07018A                     Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.025	0.012	
Motor Oil	ND	0.050	0.025	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.416	0.500	83	60-130
Hexacosane	0.122	0.125	98	60-130

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml Final Volume : 5ml
Prepared by : JMuert Analyzed by : CMpang

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 1007133
BATCH NO. : 22F042
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : MBLK1W                             LCS1W           LCD1W
LAB SAMPLE ID : DSF005WB                         DSF005WL       DSF005WC
LAB FILE ID  : LF07022A                         LF07023A       LF07024A
DATE PREPARED : 06/06/22 10:15                 06/06/22 10:15 06/06/22 10:15
DATE ANALYZED : 06/07/22 17:35                 06/07/22 17:54 06/07/22 18:12
PREP BATCH   : 22DSF005W                       22DSF005W      22DSF005W
CALIBRATION REF: LF07018A                      LF07018A       LF07018A
  
```

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Diesel	ND	2.50	2.38	95	2.50	2.43	97	2	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromobenzene	0.500	0.512	102	0.500	0.551	110	60-130
Hexacosane	0.125	0.126	101	0.125	0.135	108	60-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 1007135
BATCH NO. : 22F041
METHOD : 3520C/8015B

```

=====
MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : 202206030095 202206030095MSD 202206030095MSD
LAB SAMPLE ID : 22F041-01 22F041-01M 22F041-01S
LAB FILE ID : LF07025A LF07026A LF07027A
DATE PREPARED : 06/06/22 10:15 06/06/22 10:15 06/06/22 10:15
DATE ANALYZED : 06/07/22 18:31 06/07/22 18:49 06/07/22 19:08
PREP BATCH : 22DSF005W 22DSF005W 22DSF005W
CALIBRATION REF: LF07018A LF07018A LF07018A
  
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Diesel	ND	2.58	2.48	96	2.62	2.38	91	4	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.515	0.500	97	0.525	0.507	97	60-130
Hexacosane	0.129	0.137	106	0.131	0.128	98	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate